


Data Path : Z:\HPCHEM1\BNA G\Data\BG102717\
 Data File : BG030506.D
 Acq On : 27 Oct 2017 20:43
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG101617.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Oct 16 17:32:15 2017



AutoFind: Scans 2492, 2493, 2494; Background Corrected with Scan 2485

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.6	20968	PASS
68	69	0.00	2	1.1	220	PASS
69	198	0.00	100	32.9	19946	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	40.4	24488	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	60608	PASS
199	198	5	9	6.8	4134	PASS
275	198	10	60	27.9	16896	PASS
365	198	1	100	3.6	2170	PASS
441	442	0.01	24	16.1	8300	PASS
442	198	50	100	85.2	51626	PASS
443	442	15	24	20.4	10532	PASS