

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102722\
 Data File : BG055318.D
 Acq On : 27 Oct 2022 10:53
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 27 15:19:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 24 16:54:05 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.552	0.481	12.9	84	0.00
3	Pyridine	1.661	1.583	4.7	86	0.00
4	n-Nitrosodimethylamine	0.738	0.630	14.6	83	0.00
5 S	2-Fluorophenol	1.142	1.124	1.6	96	0.00
6	Aniline	2.166	2.098	3.1	90	0.00
7 S	Phenol-d6	1.655	1.627	1.7	92	0.00
8	2-Chlorophenol	1.328	1.343	-1.1	96	0.00
9	Benzaldehyde	1.130	0.982	13.1	88	0.00
10 C	Phenol	1.854	1.843	0.6	94	0.00
11	bis(2-Chloroethyl)ether	1.399	1.362	2.6	93	0.00
12	1,3-Dichlorobenzene	1.449	1.490	-2.8	99	0.00
13 C	1,4-Dichlorobenzene	1.480	1.522	-2.8	100	0.00
14	1,2-Dichlorobenzene	1.425	1.445	-1.4	97	0.00
15	Benzyl Alcohol	1.385	1.374	0.8	91	0.00
16	2,2'-oxybis(1-Chloropropane	2.290	2.081	9.1	86	-0.01
17	2-Methylphenol	1.311	1.309	0.2	93	0.00
18	Hexachloroethane	0.527	0.536	-1.7	98	0.00
19 P	n-Nitroso-di-n-propylamine	1.374	1.337	2.7	90	0.00
20	3+4-Methylphenols	1.878	1.884	-0.3	92	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
22	Acetophenone	0.511	0.511	0.0	92	0.00
23 S	Nitrobenzene-d5	0.392	0.386	1.5	90	0.00
24	Nitrobenzene	0.408	0.403	1.2	90	0.00
25	Isophorone	0.791	0.772	2.4	88	0.00
26 C	2-Nitrophenol	0.179	0.195	-8.9	98	0.00
27	2,4-Dimethylphenol	0.279	0.295	-5.7	98	0.00
28	bis(2-Chloroethoxy)methane	0.399	0.409	-2.5	93	0.00
29 C	2,4-Dichlorophenol	0.299	0.331	-10.7	100	0.00
30	1,2,4-Trichlorobenzene	0.305	0.342	-12.1	103	0.00
31	Naphthalene	1.067	1.093	-2.4	94	0.00
32	Benzoic acid	0.194	0.184	5.2	80	0.01
33	4-Chloroaniline	0.456	0.472	-3.5	92	0.00
34 C	Hexachlorobutadiene	0.174	0.204	-17.2	110	0.00
35	Caprolactam	0.139	0.131	5.8	82	0.00
36 C	4-Chloro-3-methylphenol	0.376	0.392	-4.3	91	0.00
37	2-Methylnaphthalene	0.769	0.802	-4.3	94	0.00
38	1-Methylnaphthalene	0.759	0.781	-2.9	93	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	0.482	0.529	-9.8	101	0.00
41 P	Hexachlorocyclopentadiene	0.198	0.212	-7.1	91	0.00
42 S	2,4,6-Tribromophenol	0.217	0.245	-12.9	97	0.00
43 C	2,4,6-Trichlorophenol	0.356	0.396	-11.2	100	0.00
44	2,4,5-Trichlorophenol	0.394	0.444	-12.7	98	0.00
45 S	2-Fluorobiphenyl	1.222	1.255	-2.7	96	0.00
46	1,1'-Biphenyl	1.372	1.414	-3.1	94	0.00
47	2-Chloronaphthalene	1.082	1.133	-4.7	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102722\
 Data File : BG055318.D
 Acq On : 27 Oct 2022 10:53
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 27 15:19:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 24 16:54:05 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.423	0.409	3.3	85	0.00
49	Acenaphthylene	1.867	1.882	-0.8	91	0.00
50	Dimethylphthalate	1.570	1.613	-2.7	93	0.00
51	2,6-Dinitrotoluene	0.320	0.338	-5.6	93	0.00
52 C	Acenaphthene	1.190	1.221	-2.6	92	0.00
53	3-Nitroaniline	0.402	0.400	0.5	87	0.00
54 P	2,4-Dinitrophenol	0.153	0.176	-15.0	94	0.00
55	Dibenzofuran	1.786	1.823	-2.1	92	0.00
56 P	4-Nitrophenol	0.287	0.281	2.1	82	0.01
57	2,4-Dinitrotoluene	0.464	0.486	-4.7	92	0.00
58	Fluorene	1.486	1.490	-0.3	91	0.00
59	2,3,4,6-Tetrachlorophenol	0.351	0.375	-6.8	93	0.00
60	Diethylphthalate	1.670	1.700	-1.8	91	0.00
61	4-Chlorophenyl-phenylether	0.683	0.739	-8.2	97	0.00
62	4-Nitroaniline	0.422	0.421	0.2	89	0.00
63	Azobenzene	1.627	1.556	4.4	86	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
65	4,6-Dinitro-2-methylphenol	0.110	0.126	-14.5	96	0.00
66 c	n-Nitrosodiphenylamine	0.557	0.582	-4.5	92	0.00
67	4-Bromophenyl-phenylether	0.194	0.219	-12.9	97	0.00
68	Hexachlorobenzene	0.220	0.245	-11.4	99	0.00
69	Atrazine	0.189	0.208	-10.1	96	0.00
70 C	Pentachlorophenol	0.108	0.119	-10.2	89	0.00
71	Phenanthrene	1.067	1.089	-2.1	90	0.00
72	Anthracene	1.043	1.066	-2.2	89	0.00
73	Carbazole	1.003	1.005	-0.2	89	0.00
74	Di-n-butylphthalate	1.254	1.281	-2.2	91	0.00
75 C	Fluoranthene	1.243	1.213	2.4	87	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
77	Benzidine	0.484	0.464	4.1	78	0.00
78	Pyrene	1.371	1.378	-0.5	87	0.00
79 S	Terphenyl-d14	0.963	0.988	-2.6	90	0.00
80	Butylbenzylphthalate	0.648	0.604	6.8	84	0.00
81	Benzo(a)anthracene	1.343	1.302	3.1	88	0.00
82	3,3'-Dichlorobenzidine	0.437	0.452	-3.4	91	0.00
83	Chrysene	1.276	1.238	3.0	87	0.00
84	Bis(2-ethylhexyl)phthalate	0.925	0.867	6.3	85	0.00
85 c	Di-n-octyl phthalate	1.542	1.468	4.8	85	0.00
86 I	Perylene-d12	1.000	1.000	0.0	93	0.02
87	Indeno(1,2,3-cd)pyrene	1.477	1.521	-3.0	93	0.02
88	Benzo(b)fluoranthene	1.203	1.210	-0.6	91	0.00
89	Benzo(k)fluoranthene	1.215	1.207	0.7	89	0.01
90 C	Benzo(a)pyrene	1.032	1.035	-0.3	90	0.02
91	Dibenzo(a,h)anthracene	1.212	1.264	-4.3	94	0.02
92	Benzo(g,h,i)perylene	1.202	1.234	-2.7	93	0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102722\
Data File : BG055318.D
Acq On : 27 Oct 2022 10:53
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 27 15:19:16 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 24 16:54:05 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0