

Data Path : Z:\HPCHEM1\BNA G\DATA\BG102815\
 Data File : BG019394.D
 Acq On : 28 Oct 2015 23:34
 Operator : UM/NP
 Sample : G4120-08
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 E5LQ6

Quant Time: Oct 29 00:54:38 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 29 01:04:20 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.20	152	31545	20.00	ng/ul	0.00
18) Naphthalene-d8	11.04	136	129550	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.83	164	105497	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.57	188	289917	20.00	ng/ul	0.00
78) Chrysene-d12	21.86	240	376274	20.00	ng/ul	0.00
86) Perylene-d12	25.24	264	350542	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.52	96	994	1.51	ng/uL	0.00
5) Phenol-d5	7.35	99	21642	7.47	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.52	67	66056	35.98	ng/ul	0.00
9) 2-Chlorophenol-d4	7.73	132	49950	27.63	ng/ul	0.00
13) 4-Methylphenol-d8	8.90	113	41811	18.13	ng/ul	0.00
19) Nitrobenzene-d5	9.38	128	33478	35.16	ng/ul	0.00
22) 2-Nitrophenol-d4	10.11	143	40544	35.87	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.65	165	80671	32.90	ng/ul	0.00
29) 4-Chloroaniline-d4	0.00	131	0d	0.00	ng/ul	0.00
44) Dimethylphthalate-d6	14.23	166	309390	35.57	ng/ul	0.00
47) Acenaphthylene-d8	14.53	160	336279	34.91	ng/ul	0.00
52) 4-Nitrophenol-d4	15.01	143	9505	6.88	ng/ul	0.00
58) Fluorene-d10	15.82	176	273927	34.04	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.92	200	63648	33.92	ng/ul	0.00
71) Anthracene-d10	17.67	188	459714	34.78	ng/ul	0.00
79) Pyrene-d10	19.95	212	573081	34.80	ng/ul	0.00
90) Benzo(a)pyrene-d12	25.01	264	612509	34.82	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG102815\
Data File : BG019394.D
Acq On : 28 Oct 2015 23:34
Operator : UM/NP
Sample : G4120-08
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E5LQ6

Quant Time: Oct 29 00:54:38 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 29 01:04:20 2015
Response via : Initial Calibration

