

Data Path : Z:\HPCHEM1\BNA_G\Data\BG102816\
 Data File : BG024539.D
 Acq On : 28 Oct 2016 14:57
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 28 23:26:42 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 26 11:14:13 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.499	0.521	-4.4	108	0.00
3	Pyridine	1.493	1.529	-2.4	105	0.00
4	n-Nitrosodimethylamine	0.773	0.745	3.6	97	0.00
5 S	2-Fluorophenol	1.220	1.209	0.9	103	0.00
6	Aniline	2.121	2.125	-0.2	101	0.00
7 S	Phenol-d6	1.674	1.720	-2.7	104	0.00
8	2-Chlorophenol	1.241	1.301	-4.8	110	0.00
9	Benzaldehyde	1.034	0.984	4.8	97	0.00
10 C	Phenol	1.725	1.781	-3.2	106	0.00
11	bis(2-Chloroethyl)ether	1.296	1.325	-2.2	107	0.00
12	1,3-Dichlorobenzene	1.536	1.541	-0.3	106	0.00
13 C	1,4-Dichlorobenzene	1.517	1.566	-3.2	107	0.00
14	1,2-Dichlorobenzene	1.459	1.497	-2.6	107	0.00
15	Benzyl Alcohol	1.557	1.607	-3.2	106	0.00
16	2,2'-oxybis(1-Chloropropane	2.405	2.421	-0.7	104	0.00
17	2-Methylphenol	1.143	1.180	-3.2	107	0.00
18	Hexachloroethane	0.583	0.614	-5.3	113	0.00
19 P	n-Nitroso-di-n-propylamine	1.234	1.241	-0.6	100	0.00
20	3+4-Methylphenols	1.535	1.636	-6.6	106	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
22	Acetophenone	0.514	0.517	-0.6	104	0.00
23 S	Nitrobenzene-d5	0.439	0.446	-1.6	106	0.00
24	Nitrobenzene	0.489	0.496	-1.4	104	0.00
25	Isophorone	0.817	0.789	3.4	98	0.00
26 C	2-Nitrophenol	0.181	0.174	3.9	99	0.00
27	2,4-Dimethylphenol	0.318	0.317	0.3	101	0.00
28	bis(2-Chloroethoxy)methane	0.423	0.428	-1.2	107	0.00
29 C	2,4-Dichlorophenol	0.333	0.333	0.0	101	0.00
30	1,2,4-Trichlorobenzene	0.395	0.397	-0.5	105	0.00
31	Naphthalene	0.998	1.003	-0.5	103	0.00
32	Benzoic acid	0.264	0.196	25.8#	77	0.00
33	4-Chloroaniline	0.433	0.431	0.5	98	0.00
34 C	Hexachlorobutadiene	0.309	0.315	-1.9	109	0.00
35	Caprolactam	0.122	0.106	13.1	88	0.00
36 C	4-Chloro-3-methylphenol	0.402	0.395	1.7	98	0.00
37	2-Methylnaphthalene	0.728	0.728	0.0	103	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
39	1,2,4,5-Tetrachlorobenzene	0.675	0.676	-0.1	102	0.00
40 P	Hexachlorocyclopentadiene	0.380	0.350	7.9	90	0.00
41 S	2,4,6-Tribromophenol	0.299	0.300	-0.3	93	0.00
42 C	2,4,6-Trichlorophenol	0.405	0.415	-2.5	102	0.00
43	2,4,5-Trichlorophenol	0.438	0.418	4.6	93	0.00
44 S	2-Fluorobiphenyl	1.203	1.218	-1.2	101	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG102816\
 Data File : BG024539.D
 Acq On : 28 Oct 2016 14:57
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 28 23:26:42 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 26 11:14:13 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.388	1.387	0.1	99	0.00
46	2-Chloronaphthalene	1.057	1.055	0.2	100	0.00
47	2-Nitroaniline	0.433	0.444	-2.5	93	0.00
48	Acenaphthylene	1.621	1.615	0.4	96	0.00
49	Dimethylphthalate	1.420	1.388	2.3	93	0.00
50	2,6-Dinitrotoluene	0.303	0.310	-2.3	94	0.00
51 C	Acenaphthene	1.208	1.079	10.7	87	0.00
52	3-Nitroaniline	0.314	0.308	1.9	92	0.00
53 P	2,4-Dinitrophenol	0.220	0.224	-1.8	92	0.00
54	Dibenzofuran	1.670	1.656	0.8	96	0.00
55 P	4-Nitrophenol	0.273	0.260	4.8	89	0.00
56	2,4-Dinitrotoluene	0.440	0.445	-1.1	90	0.00
57	Fluorene	1.385	1.376	0.6	95	0.00
58	2,3,4,6-Tetrachlorophenol	0.454	0.460	-1.3	94	0.00
59	Diethylphthalate	1.489	1.388	6.8	88	0.00
60	4-Chlorophenyl-phenylether	0.777	0.769	1.0	97	0.00
61	4-Nitroaniline	0.343	0.346	-0.9	95	0.00
62	Azobenzene	1.485	1.424	4.1	91	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
64	4,6-Dinitro-2-methylphenol	0.138	0.137	0.7	88	0.00
65 c	n-Nitrosodiphenylamine	0.547	0.540	1.3	91	0.00
66	4-Bromophenyl-phenylether	0.243	0.250	-2.9	95	0.00
67	Hexachlorobenzene	0.268	0.280	-4.5	94	0.00
68	Atrazine	0.235	0.231	1.7	88	0.00
69 C	Pentachlorophenol	0.177	0.164	7.3	81	0.00
70	Phenanthrene	1.019	1.016	0.3	92	0.00
71	Anthracene	1.028	1.025	0.3	91	0.00
72	Carbazole	0.938	0.926	1.3	91	0.00
73	Di-n-butylphthalate	1.081	1.079	0.2	91	0.00
74 C	Fluoranthene	1.289	1.315	-2.0	91	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
76	Benzidine	0.471	0.432	8.3	87	0.00
77	Pyrene	0.978	0.954	2.5	91	0.00
78 S	Terphenyl-d14	0.669	0.651	2.7	93	0.00
79	Butylbenzylphthalate	0.408	0.402	1.5	94	0.00
80	Benzo(a)anthracene	1.099	1.075	2.2	95	0.00
81	3,3'-Dichlorobenzidine	0.443	0.434	2.0	94	0.00
82	Chrysene	1.046	1.040	0.6	95	0.00
83	Bis(2-ethylhexyl)phthalate	0.583	0.572	1.9	94	0.00
84 c	Di-n-octyl phthalate	0.968	0.975	-0.7	96	0.00
85	Indeno(1,2,3-cd)pyrene	1.444	1.410	2.4	97	0.00
86 I	Perylene-d12	1.000	1.000	0.0	90	-0.01
87	Benzo(b)fluoranthene	1.081	1.070	1.0	96	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG102816\
Data File : BG024539.D
Acq On : 28 Oct 2016 14:57
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 23:26:42 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG102516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 26 11:14:13 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.044	1.042	0.2	96	0.00
89 C	Benzo(a)pyrene	1.056	1.058	-0.2	97	0.00
90	Dibenzo(a,h)anthracene	1.159	1.131	2.4	98	0.00
91	Benzo(g,h,i)perylene	1.158	1.113	3.9	98	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0