

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG102819\
 Data File : BG043350.D
 Acq On : 28 Oct 2019 15:05
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 28 15:42:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG102119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 21 16:37:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	43.970	-9.9	106	0.00
3	Pyridine	40.000	45.332	-13.3	98	0.00
4	n-Nitrosodimethylamine	40.000	44.349	-10.9	108	0.00
5 S	2-Fluorophenol	80.000	86.882	-8.6	105	0.00
6	Aniline	40.000	43.617	-9.0	104	0.00
7 S	Phenol-d6	80.000	87.237	-9.0	106	0.01
8	2-Chlorophenol	40.000	42.639	-6.6	104	0.00
9	Benzaldehyde	40.000	43.777	-9.4	110	0.00
10 C	Phenol	40.000	43.335	-8.3	104	0.01
11	bis(2-Chloroethyl)ether	40.000	41.876	-4.7	100	0.00
12	1,3-Dichlorobenzene	40.000	43.212	-8.0	106	0.00
13 C	1,4-Dichlorobenzene	40.000	43.376	-8.4	106	0.00
14	1,2-Dichlorobenzene	40.000	42.392	-6.0	103	0.00
15	Benzyl Alcohol	40.000	42.359	-5.9	100	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	42.222	-5.6	100	0.00
17	2-Methylphenol	40.000	42.447	-6.1	105	0.00
18	Hexachloroethane	40.000	43.259	-8.1	106	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.376	-3.4	101	0.00
20	3+4-Methylphenols	40.000	42.834	-7.1	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	40.777	-1.9	101	0.00
23 S	Nitrobenzene-d5	80.000	81.503	-1.9	103	0.00
24	Nitrobenzene	40.000	41.145	-2.9	103	0.00
25	Isophorone	40.000	40.586	-1.5	102	0.00
26 C	2-Nitrophenol	40.000	42.334	-5.8	109	0.00
27	2,4-Dimethylphenol	40.000	41.840	-4.6	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.810	-2.0	99	0.00
29 C	2,4-Dichlorophenol	40.000	42.214	-5.5	104	0.00
30	1,2,4-Trichlorobenzene	40.000	41.984	-5.0	107	0.00
31	Naphthalene	40.000	41.355	-3.4	105	0.00
32	Benzoic acid	40.000	36.068	9.8	101	0.01
33	4-Chloroaniline	40.000	42.305	-5.8	103	0.00
34 C	Hexachlorobutadiene	40.000	42.302	-5.8	108	0.00
35	Caprolactam	40.000	40.940	-2.3	100	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.880	-2.2	101	0.01
37	2-Methylnaphthalene	40.000	41.135	-2.8	103	0.00
38	1-Methylnaphthalene	40.000	41.293	-3.2	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.363	-5.9	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.897	-7.2	106	0.00
42 S	2,4,6-Tribromophenol	80.000	84.077	-5.1	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.645	-4.1	102	0.00
44	2,4,5-Trichlorophenol	40.000	42.791	-7.0	106	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG102819\
 Data File : BG043350.D
 Acq On : 28 Oct 2019 15:05
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 28 15:42:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG102119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 21 16:37:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	80.000	83.604	-4.5	104	0.00
46	1,1'-Biphenyl	40.000	41.502	-3.8	102	0.00
47	2-Chloronaphthalene	40.000	41.261	-3.2	104	0.00
48	2-Nitroaniline	40.000	42.148	-5.4	102	0.00
49	Acenaphthylene	40.000	42.052	-5.1	104	0.00
50	Dimethylphthalate	40.000	40.792	-2.0	102	0.00
51	2,6-Dinitrotoluene	40.000	40.604	-1.5	100	0.00
52 C	Acenaphthene	40.000	41.328	-3.3	103	0.00
53	3-Nitroaniline	40.000	41.999	-5.0	103	0.00
54 P	2,4-Dinitrophenol	40.000	36.422	8.9	94	0.00
55	Dibenzofuran	40.000	41.604	-4.0	104	0.00
56 P	4-Nitrophenol	40.000	41.993	-5.0	102	0.01
57	2,4-Dinitrotoluene	40.000	41.187	-3.0	103	0.00
58	Fluorene	40.000	41.767	-4.4	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.565	-3.9	98	0.00
60	Diethylphthalate	40.000	40.968	-2.4	102	0.00
61	4-Chlorophenyl-phenylether	40.000	41.536	-3.8	104	0.00
62	4-Nitroaniline	40.000	42.353	-5.9	103	0.00
63	Azobenzene	40.000	40.586	-1.5	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.596	-6.5	107	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.466	-3.7	102	0.00
67	4-Bromophenyl-phenylether	40.000	41.939	-4.8	104	0.00
68	Hexachlorobenzene	40.000	41.604	-4.0	104	0.00
69	Atrazine	40.000	37.575	6.1	95	0.00
70 C	Pentachlorophenol	40.000	42.912	-7.3	103	0.00
71	Phenanthrene	40.000	41.838	-4.6	103	0.00
72	Anthracene	40.000	41.836	-4.6	102	0.00
73	Carbazole	40.000	41.317	-3.3	101	0.00
74	Di-n-butylphthalate	40.000	41.962	-4.9	102	0.00
75 C	Fluoranthene	40.000	41.367	-3.4	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzidine	40.000	40.976	-2.4	95	0.00
78	Pyrene	40.000	41.527	-3.8	103	0.00
79 S	Terphenyl-d14	80.000	82.743	-3.4	102	0.00
80	Butylbenzylphthalate	40.000	41.491	-3.7	103	0.00
81	Benzo(a)anthracene	40.000	42.248	-5.6	107	0.00
82	3,3'-Dichlorobenzidine	40.000	42.640	-6.6	105	0.00
83	Chrysene	40.000	42.147	-5.4	105	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.429	-6.1	107	0.00
85 c	Di-n-octyl phthalate	40.000	42.474	-6.2	108	0.00
86	Indeno(1,2,3-cd)pyrene	40.000	43.340	-8.4	109	-0.01
87 I	Perylene-d12	20.000	20.000	0.0	109	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG102819\
Data File : BG043350.D
Acq On : 28 Oct 2019 15:05
Operator : JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 15:42:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG102119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 21 16:37:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	40.000	41.598	-4.0	108	0.00
89	Benzo(k)fluoranthene	40.000	41.240	-3.1	107	0.00
90 C	Benzo(a)pyrene	40.000	41.582	-4.0	108	0.00
91	Dibenzo(a,h)anthracene	40.000	41.824	-4.6	109	0.00
92	Benzo(q,h,i)perylene	40.000	41.782	-4.5	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0