


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG102819\
Data File : BG043349.D
Acq On : 28 Oct 2019 14:27
Operator : JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG102119.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Oct 21 16:37:43 2019



AutoFind: Scans 2508, 2509, 2510; Background Corrected with Scan 2501

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.7	10511	PASS
68	69	0.00	2	0.7	73	PASS
69	198	0.00	100	31.5	10811	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	40.1	13734	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	34272	PASS
199	198	5	9	6.7	2303	PASS
275	198	10	60	34.6	11848	PASS
365	198	1	100	4.6	1579	PASS
441	198	0.01	100	14.0	4781	PASS
442	198	50	100	94.4	32349	PASS
443	442	15	24	19.2	6216	PASS