


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG103015\
Data File : BG019431.D
Acq On : 30 Oct 2015 14:13
Operator : UM/NP
Sample : DFTPP12
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP12

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG102815.M
Title : SVOA CALIBRATION
Last Update : Sun Nov 01 05:57:20 2015



AutoFind: Scans 2575, 2576, 2577; Background Corrected with Scan 2570

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.1	11300	PASS
68	69	0.00	2	0.6	89	PASS
69	198	0.00	100	40.1	16097	PASS
70	69	0.00	2	1.6	259	PASS
127	198	10	80	39.8	16003	PASS
197	198	0.00	2	0.4	177	PASS
198	198	100	100	100.0	40165	PASS
199	198	5	9	6.2	2504	PASS
275	198	10	60	29.7	11939	PASS
365	198	1	100	4.0	1614	PASS
441	442	0.01	24	15.1	4125	PASS
442	198	50	100	67.9	27264	PASS
443	442	15	24	19.4	5301	PASS

m/z	Abundance
39.90	239.0
43.00	383.0
43.90	431.0
48.90	510.0
51.00	247.0
57.10	231.0
72.90	403.0
81.00	194.0
83.90	1204.0
85.80	681.0
87.90	299.0
123.00	204.0
129.10	257.0
152.00	164.0
186.00	181.0
207.00	552.0
221.20	208.0
226.70	167.0
248.70	159.0
280.80	283.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

m/z	Abundance
37.00	182.0
39.00	838.0
40.00	316.0
40.90	301.0
42.10	261.0
42.90	279.0
44.00	459.0
47.00	158.0
48.90	870.0
50.00	3253.0
51.00	9644.0
52.00	441.0
55.00	576.0
56.00	456.0
57.00	1118.0
59.90	168.0
61.00	508.0
61.90	383.0
63.00	557.0
64.00	158.0
65.00	257.0
66.90	248.0
67.90	268.0
68.90	14244.0
70.00	291.0
71.00	195.0
73.00	659.0
74.00	1134.0
74.90	2309.0
76.00	882.0
77.00	16147.0
78.00	1101.0
78.90	1081.0
79.90	776.0
80.90	1645.0
82.00	497.0
82.90	289.0
83.90	1005.0
84.90	533.0
85.80	1111.0
86.90	260.0
90.70	434.0
91.00	492.0
91.80	153.0
92.90	2350.0
97.10	177.0
97.90	1316.0
99.00	1315.0
100.90	1172.0
102.80	242.0
103.90	537.0
104.90	452.0
106.90	5184.0
107.90	768.0
109.90	7564.0
110.90	1294.0
111.90	176.0
115.00	162.0
116.00	378.0
116.90	4059.0
118.00	392.0
121.90	257.0
123.00	890.0
124.20	262.0
124.90	198.0
126.90	13803.0
128.00	1084.0
128.90	6057.0
129.80	411.0
133.10	174.0
134.00	258.0
134.90	499.0
135.70	200.0
136.20	221.0
136.90	320.0
140.90	1248.0
142.90	162.0
145.90	258.0
147.00	667.0
147.90	1456.0
149.00	242.0
151.10	270.0
152.00	266.0
153.70	244.0
154.90	287.0
156.00	766.0
158.00	188.0
159.10	180.0
159.70	176.0
161.00	837.0
163.70	206.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

164.90	287.0
165.90	217.0
166.90	2848.0
167.90	998.0
168.90	228.0
170.60	162.0
171.90	280.0
173.00	331.0
173.90	459.0
175.00	862.0
176.90	577.0
178.00	180.0
178.90	1372.0
180.00	768.0
180.80	601.0
181.90	194.0
184.00	267.0
184.70	521.0
186.00	4188.0
187.00	1725.0
187.80	254.0
188.80	337.0
191.90	546.0
193.00	740.0
195.90	755.0
197.90	34896.0
198.80	2153.0
200.10	174.0
201.70	223.0
202.70	417.0
204.00	1334.0
204.90	2264.0
205.90	10832.0
207.00	1806.0
208.00	530.0
209.90	150.0
210.60	514.0
210.80	510.0
211.60	171.0
211.90	164.0
215.80	210.0
216.90	4078.0
218.00	380.0
218.90	187.0
220.30	211.0
220.80	632.0
221.80	343.0
222.90	892.0
224.00	5949.0
224.90	1652.0
225.90	306.0
226.90	3107.0
227.90	810.0
228.90	747.0
230.90	241.0
233.80	342.0
234.90	258.0
236.90	299.0
241.10	243.0
242.00	259.0
242.80	243.0
244.00	4253.0
245.00	637.0
245.90	1176.0
250.00	182.0
254.90	23680.0
255.90	3733.0
256.80	287.0
258.00	1017.0
264.90	862.0
265.90	205.0
270.90	156.0
273.00	534.0
273.90	1590.0
274.90	8291.0
275.90	992.0
276.90	920.0
280.70	213.0
283.00	253.0
284.80	155.0
292.90	236.0
295.90	2778.0
296.90	514.0
303.00	525.0
314.30	179.0
314.80	215.0
316.10	192.0
323.10	764.0
324.00	212.0
326.80	236.0
334.00	701.0
345.70	196.0
352.10	335.0
352.90	207.0
365.00	761.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

372.00	331.0
372.90	173.0
377.20	155.0
403.00	253.0
423.00	725.0
423.80	353.0
441.00	2212.0
442.00	17512.0
443.00	2894.0
444.00	345.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

m/z	Abundance
38.20	291.0
38.90	716.0
41.00	335.0
41.90	222.0
42.90	303.0
44.00	306.0
46.90	393.0
48.90	917.0
50.00	3359.0
51.00	14353.0
52.10	643.0
55.00	273.0
56.00	484.0
57.00	1230.0
57.70	208.0
60.90	308.0
62.10	247.0
63.00	810.0
63.90	208.0
64.90	431.0
68.90	18176.0
69.90	287.0
71.10	288.0
73.00	415.0
73.90	2479.0
74.90	3758.0
76.00	1509.0
77.00	22120.0
78.00	1826.0
79.00	1623.0
79.90	1132.0
81.00	1520.0
81.90	431.0
83.00	775.0
83.90	1008.0
85.00	217.0
85.90	936.0
87.00	532.0
88.00	167.0
90.90	374.0
91.90	588.0
92.90	2670.0
98.00	2128.0
98.90	1455.0
100.00	181.0
100.90	780.0
102.90	431.0
103.90	858.0
104.90	459.0
106.00	183.0
106.90	7401.0
108.00	1120.0
108.80	409.0
109.90	9450.0
111.00	2018.0
111.90	265.0
116.00	453.0
116.90	5154.0
118.00	375.0
122.10	385.0
122.90	449.0
124.00	227.0
125.10	377.0
127.00	17952.0
128.00	1721.0
128.90	8776.0
130.00	754.0
131.00	247.0
133.80	351.0
134.90	929.0
135.80	534.0
136.90	459.0
138.90	318.0
141.00	1162.0
142.00	371.0
142.70	230.0
144.80	240.0
146.10	308.0
147.00	734.0
147.90	1747.0
148.90	429.0
150.80	248.0
152.90	351.0
155.00	369.0
155.90	1148.0
157.90	571.0
158.70	249.0
159.00	319.0
161.00	472.0
161.90	203.0
164.90	881.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

166.00	487.0
166.90	2191.0
167.90	1368.0
169.00	235.0
170.00	186.0
171.90	228.0
172.90	326.0
173.90	685.0
174.90	1264.0
175.90	277.0
176.90	732.0
178.90	2079.0
179.90	1207.0
181.00	850.0
184.00	225.0
184.90	1008.0
185.90	6344.0
187.00	2619.0
187.70	158.0
188.20	220.0
188.90	522.0
190.90	459.0
191.90	395.0
192.90	892.0
193.90	274.0
194.80	256.0
195.90	1178.0
197.90	45896.0
198.90	2536.0
199.80	273.0
201.20	165.0
202.90	391.0
203.90	2499.0
204.90	3481.0
206.00	14104.0
206.90	2142.0
207.90	809.0
208.90	305.0
209.90	263.0
210.80	464.0
216.00	313.0
216.90	4073.0
218.00	473.0
220.00	225.0
220.90	1396.0
222.00	681.0
222.90	1103.0
223.90	9630.0
224.90	2578.0
225.90	154.0
226.90	4199.0
227.90	669.0
228.90	619.0
229.70	186.0
230.80	154.0
233.80	234.0
234.90	390.0
236.00	388.0
236.80	260.0
239.10	253.0
240.00	157.0
240.80	392.0
241.80	408.0
242.90	305.0
244.00	5875.0
245.00	1143.0
245.90	1168.0
247.00	161.0
252.70	199.0
253.10	337.0
253.90	433.0
254.90	36096.0
255.90	5083.0
256.90	495.0
257.90	2885.0
258.80	405.0
264.80	1152.0
266.00	344.0
272.90	1297.0
273.90	2643.0
274.90	14725.0
276.00	1879.0
276.90	1065.0
277.90	336.0
280.90	253.0
282.80	185.0
284.00	243.0
285.00	250.0
292.90	397.0
294.00	162.0
295.00	240.0
295.90	5689.0
297.00	521.0
302.00	171.0
302.90	567.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

304.20	173.0
308.90	183.0
313.90	244.0
314.90	491.0
316.00	463.0
322.90	1608.0
323.80	260.0
326.80	321.0
334.00	1364.0
341.90	160.0
345.90	569.0
352.00	400.0
353.00	391.0
353.80	210.0
354.80	270.0
364.90	1916.0
365.90	573.0
369.80	157.0
372.00	849.0
372.80	214.0
402.90	191.0
420.90	518.0
421.70	276.0
423.00	1782.0
423.90	267.0
441.00	4633.0
442.00	30672.0
443.00	6078.0
444.00	313.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

m/z	Abundance
38.00	332.0
38.90	494.0
41.00	272.0
43.00	424.0
44.00	283.0
47.80	151.0
48.90	868.0
50.00	2879.0
51.00	10645.0
52.00	430.0
55.00	393.0
55.90	315.0
56.90	667.0
62.00	187.0
63.00	834.0
68.90	15873.0
70.00	201.0
71.00	245.0
73.00	445.0
73.90	1994.0
74.90	2832.0
76.00	793.0
77.00	17216.0
78.00	1301.0
79.00	1170.0
79.90	1164.0
81.00	1049.0
81.90	360.0
83.00	426.0
83.90	971.0
84.80	355.0
85.90	681.0
86.90	210.0
87.90	192.0
91.00	537.0
91.90	377.0
92.90	2709.0
97.90	2213.0
98.90	1223.0
100.90	761.0
101.70	171.0
102.90	238.0
103.90	543.0
104.90	828.0
107.00	6687.0
108.00	565.0
109.00	351.0
109.90	7626.0
110.90	1296.0
114.80	227.0
115.90	281.0
117.00	4797.0
117.90	348.0
118.80	196.0
122.00	410.0
122.90	712.0
123.90	683.0
125.00	382.0
125.70	208.0
126.90	16255.0
128.00	1667.0
128.90	7276.0
129.90	970.0
134.00	277.0
134.90	695.0
135.70	320.0
136.80	367.0
140.00	194.0
140.90	1087.0
141.80	230.0
142.10	224.0
145.90	195.0
147.00	454.0
147.90	1328.0
149.00	415.0
150.20	197.0
150.90	220.0
151.80	238.0
152.90	665.0
153.90	426.0
155.00	632.0
155.90	963.0
157.90	362.0
158.70	247.0
159.90	287.0
161.00	601.0
162.00	219.0
164.90	494.0
165.90	585.0
166.90	2333.0
167.90	1248.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

168.80	264.0
171.70	189.0
172.90	301.0
173.90	438.0
174.90	1466.0
175.80	580.0
176.90	565.0
178.10	275.0
178.90	2236.0
179.90	1518.0
180.90	795.0
181.90	269.0
183.40	171.0
184.00	284.0
184.90	923.0
185.90	6580.0
186.90	1959.0
188.00	229.0
188.90	672.0
189.80	180.0
191.00	463.0
192.00	707.0
192.90	499.0
193.70	196.0
195.90	1145.0
196.80	533.0
197.90	39704.0
198.90	2824.0
200.00	429.0
201.00	281.0
201.70	206.0
203.00	542.0
203.90	2154.0
204.90	3392.0
205.90	13740.0
207.00	2315.0
208.00	713.0
208.80	425.0
210.90	461.0
214.90	345.0
215.80	313.0
216.80	3912.0
217.90	565.0
220.20	406.0
220.90	867.0
221.40	342.0
221.90	318.0
223.00	1033.0
224.00	8652.0
224.90	1361.0
226.00	334.0
227.00	4084.0
227.90	506.0
229.00	712.0
230.00	162.0
231.00	269.0
233.80	291.0
234.90	376.0
236.00	194.0
237.00	276.0
238.80	195.0
239.60	150.0
241.00	384.0
241.90	354.0
243.10	446.0
243.90	5357.0
245.10	405.0
245.90	1474.0
246.90	360.0
248.80	363.0
252.00	213.0
253.10	183.0
253.80	429.0
254.90	33720.0
255.90	5186.0
256.80	298.0
257.90	2589.0
258.80	389.0
264.90	1001.0
272.90	1019.0
273.90	2121.0
274.90	12802.0
275.90	1925.0
276.90	1396.0
277.90	221.0
282.80	232.0
284.10	234.0
285.10	195.0
292.90	258.0
295.90	4362.0
297.00	808.0
302.90	652.0
303.80	278.0
304.10	253.0
314.10	215.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12

314.90	541.0
315.90	408.0
321.00	151.0
323.00	1329.0
323.80	310.0
326.70	282.0
327.80	236.0
333.80	914.0
334.90	213.0
340.90	326.0
352.00	519.0
353.80	388.0
364.90	2167.0
366.00	341.0
369.90	184.0
371.90	706.0
372.80	216.0
390.10	162.0
401.90	327.0
403.00	661.0
404.10	193.0
420.90	260.0
422.00	207.0
422.90	2271.0
424.00	443.0
441.00	5532.0
442.00	33608.0
443.00	6932.0
444.00	695.0

Instrument :
BNA_G
ClientSampleId :
DFTPP12