

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103124\
 Data File : BG063137.D
 Acq On : 31 Oct 2024 17:38
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG103124

Quant Time: Nov 01 05:14:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.507	0.482	4.9	94	0.00
3	Pyridine	1.223	1.271	-3.9	100	0.00
4	n-Nitrosodimethylamine	0.760	0.742	2.4	88	0.00
5 S	2-Fluorophenol	1.290	1.185	8.1	96	0.00
6	Aniline	1.520	1.494	1.7	98	0.00
7 S	Phenol-d6	1.690	1.682	0.5	99	0.00
8	2-Chlorophenol	1.198	1.182	1.3	98	0.00
9	Benzaldehyde	1.050	1.158	-10.3	109	0.00
10 C	Phenol	1.811	1.777	1.9	100	0.00
11	bis(2-Chloroethyl)ether	1.221	1.185	2.9	98	0.00
12	1,3-Dichlorobenzene	1.453	1.410	3.0	98	0.00
13 C	1,4-Dichlorobenzene	1.481	1.428	3.6	98	0.00
14	1,2-Dichlorobenzene	1.428	1.348	5.6	96	0.00
15	Benzyl Alcohol	1.358	1.413	-4.1	101	0.00
16	2,2'-oxybis(1-Chloropropane	2.215	2.159	2.5	99	0.00
17	2-Methylphenol	1.152	1.113	3.4	98	0.00
18	Hexachloroethane	0.530	0.525	0.9	97	0.00
19 P	n-Nitroso-di-n-propylamine	1.177	1.176	0.1	100	0.00
20	3+4-Methylphenols	1.536	1.520	1.0	98	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.536	0.527	1.7	99	0.00
23 S	Nitrobenzene-d5	0.460	0.459	0.2	103	0.00
24	Nitrobenzene	0.497	0.493	0.8	103	0.00
25	Isophorone	0.795	0.804	-1.1	102	0.00
26 C	2-Nitrophenol	0.183	0.182	0.5	103	0.00
27	2,4-Dimethylphenol	0.225	0.226	-0.4	104	0.00
28	bis(2-Chloroethoxy)methane	0.417	0.404	3.1	99	0.00
29 C	2,4-Dichlorophenol	0.340	0.331	2.6	99	0.00
30	1,2,4-Trichlorobenzene	0.418	0.405	3.1	98	0.00
31	Naphthalene	1.055	1.018	3.5	98	0.00
32	Benzoic acid	0.196	0.203	-3.6	102	0.00
33	4-Chloroaniline	0.356	0.365	-2.5	102	0.00
34 C	Hexachlorobutadiene	0.344	0.338	1.7	100	0.00
35	Caprolactam	0.108	0.103	4.6	95	0.00
36 C	4-Chloro-3-methylphenol	0.391	0.367	6.1	97	0.00
37	2-Methylnaphthalene	0.748	0.713	4.7	100	0.00
38	1-Methylnaphthalene	0.737	0.719	2.4	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	0.719	0.706	1.8	99	0.00
41 P	Hexachlorocyclopentadiene	0.234	0.241	-3.0	100	0.00
42 S	2,4,6-Tribromophenol	0.340	0.337	0.9	101	0.00
43 C	2,4,6-Trichlorophenol	0.430	0.432	-0.5	97	0.00
44	2,4,5-Trichlorophenol	0.473	0.475	-0.4	98	0.00
45 S	2-Fluorobiphenyl	1.350	1.352	-0.1	100	0.00
46	1,1'-Biphenyl	1.357	1.358	-0.1	101	0.00
47	2-Chloronaphthalene	1.137	1.118	1.7	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103124\
 Data File : BG063137.D
 Acq On : 31 Oct 2024 17:38
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG103124

Quant Time: Nov 01 05:14:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.399	0.397	0.5	96	0.00
49	Acenaphthylene	1.605	1.594	0.7	102	0.00
50	Dimethylphthalate	1.425	1.407	1.3	102	0.00
51	2,6-Dinitrotoluene	0.304	0.301	1.0	100	0.00
52 C	Acenaphthene	0.994	0.987	0.7	101	0.00
53	3-Nitroaniline	0.289	0.297	-2.8	101	0.00
54 P	2,4-Dinitrophenol	0.186	0.193	-3.8	101	0.00
55	Dibenzofuran	1.752	1.736	0.9	101	0.00
56 P	4-Nitrophenol	0.219	0.235	-7.3	106	0.00
57	2,4-Dinitrotoluene	0.435	0.440	-1.1	102	0.00
58	Fluorene	1.388	1.381	0.5	100	0.00
59	2,3,4,6-Tetrachlorophenol	0.454	0.469	-3.3	104	0.00
60	Diethylphthalate	1.464	1.462	0.1	100	0.00
61	4-Chlorophenyl-phenylether	0.863	0.849	1.6	99	0.00
62	4-Nitroaniline	0.319	0.309	3.1	99	0.00
63	Azobenzene	1.514	1.486	1.8	100	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	0.129	0.127	1.6	99	0.00
66 c	n-Nitrosodiphenylamine	0.499	0.487	2.4	100	0.00
67	4-Bromophenyl-phenylether	0.245	0.247	-0.8	102	0.00
68	Hexachlorobenzene	0.284	0.281	1.1	100	0.00
69	Atrazine	0.229	0.190	17.0	103	0.00
70 C	Pentachlorophenol	0.141	0.151	-7.1	100	0.00
71	Phenanthrene	0.970	0.940	3.1	100	0.00
72	Anthracene	0.952	0.927	2.6	99	0.00
73	Carbazole	0.896	0.851	5.0	99	0.00
74	Di-n-butylphthalate	1.017	0.963	5.3	99	0.00
75 C	Fluoranthene	1.372	1.261	8.1	98	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
77	Benzidine	0.308	0.135	56.2#	51	0.00
78	Pyrene	1.186	1.214	-2.4	98	0.00
79 S	Terphenyl-d14	0.972	0.989	-1.7	98	0.00
80	Butylbenzylphthalate	0.382	0.382	0.0	98	0.00
81	Benzo(a)anthracene	1.259	1.254	0.4	98	0.00
82	3,3'-Dichlorobenzidine	0.467	0.458	1.9	98	0.00
83	Chrysene	1.183	1.179	0.3	98	0.00
84	Bis(2-ethylhexyl)phthalate	0.554	0.556	-0.4	99	0.00
85 c	Di-n-octyl phthalate	0.927	0.941	-1.5	100	0.00
86 I	Perylene-d12	1.000	1.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	1.342	1.315	2.0	100	0.00
88	Benzo(b)fluoranthene	1.166	1.154	1.0	102	0.00
89	Benzo(k)fluoranthene	1.112	1.093	1.7	99	0.00
90 C	Benzo(a)pyrene	1.018	0.995	2.3	101	0.00
91	Dibenzo(a,h)anthracene	1.120	1.098	2.0	100	0.00
92	Benzo(g,h,i)perylene	1.108	1.076	2.9	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103124\
Data File : BG063137.D
Acq On : 31 Oct 2024 17:38
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG103124

Quant Time: Nov 01 05:14:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 01 05:13:40 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0