

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103124\
 Data File : BG063137.D
 Acq On : 31 Oct 2024 17:38
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG103124

Quant Time: Nov 01 05:14:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	38.054	4.9	94	0.00
3	Pyridine	40.000	41.563	-3.9	100	0.00
4	n-Nitrosodimethylamine	40.000	39.077	2.3	88	0.00
5 S	2-Fluorophenol	80.000	73.496	8.1	96	0.00
6	Aniline	40.000	39.323	1.7	98	0.00
7 S	Phenol-d6	80.000	79.603	0.5	99	0.00
8	2-Chlorophenol	40.000	39.465	1.3	98	0.00
9	Benzaldehyde	40.000	44.105	-10.3	109	0.00
10 C	Phenol	40.000	39.251	1.9	100	0.00
11	bis(2-Chloroethyl)ether	40.000	38.826	2.9	98	0.00
12	1,3-Dichlorobenzene	40.000	38.815	3.0	98	0.00
13 C	1,4-Dichlorobenzene	40.000	38.564	3.6	98	0.00
14	1,2-Dichlorobenzene	40.000	37.762	5.6	96	0.00
15	Benzyl Alcohol	40.000	41.608	-4.0	101	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.995	2.5	99	0.00
17	2-Methylphenol	40.000	38.637	3.4	98	0.00
18	Hexachloroethane	40.000	39.575	1.1	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.948	0.1	100	0.00
20	3+4-Methylphenols	40.000	39.574	1.1	98	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	39.326	1.7	99	0.00
23 S	Nitrobenzene-d5	80.000	79.833	0.2	103	0.00
24	Nitrobenzene	40.000	39.621	0.9	103	0.00
25	Isophorone	40.000	40.409	-1.0	102	0.00
26 C	2-Nitrophenol	40.000	39.815	0.5	103	0.00
27	2,4-Dimethylphenol	40.000	40.230	-0.6	104	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.700	3.2	99	0.00
29 C	2,4-Dichlorophenol	40.000	39.001	2.5	99	0.00
30	1,2,4-Trichlorobenzene	40.000	38.780	3.0	98	0.00
31	Naphthalene	40.000	38.584	3.5	98	0.00
32	Benzoic acid	40.000	41.435	-3.6	102	0.00
33	4-Chloroaniline	40.000	40.966	-2.4	102	0.00
34 C	Hexachlorobutadiene	40.000	39.385	1.5	100	0.00
35	Caprolactam	40.000	38.068	4.8	95	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.483	6.3	97	0.00
37	2-Methylnaphthalene	40.000	38.110	4.7	100	0.00
38	1-Methylnaphthalene	40.000	39.037	2.4	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.270	1.8	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.219	-3.0	100	0.00
42 S	2,4,6-Tribromophenol	80.000	79.441	0.7	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.187	-0.5	97	0.00
44	2,4,5-Trichlorophenol	40.000	40.162	-0.4	98	0.00
45 S	2-Fluorobiphenyl	80.000	80.126	-0.2	100	0.00
46	1,1'-Biphenyl	40.000	40.011	-0.0	101	0.00
47	2-Chloronaphthalene	40.000	39.334	1.7	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103124\
 Data File : BG063137.D
 Acq On : 31 Oct 2024 17:38
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG103124

Quant Time: Nov 01 05:14:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.836	0.4	96	0.00
49	Acenaphthylene	40.000	39.724	0.7	102	0.00
50	Dimethylphthalate	40.000	39.476	1.3	102	0.00
51	2,6-Dinitrotoluene	40.000	39.611	1.0	100	0.00
52 C	Acenaphthene	40.000	39.720	0.7	101	0.00
53	3-Nitroaniline	40.000	41.143	-2.9	101	0.00
54 P	2,4-Dinitrophenol	40.000	41.570	-3.9	101	0.00
55	Dibenzofuran	40.000	39.647	0.9	101	0.00
56 P	4-Nitrophenol	40.000	43.022	-7.6	106	0.00
57	2,4-Dinitrotoluene	40.000	40.427	-1.1	102	0.00
58	Fluorene	40.000	39.799	0.5	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.298	-3.2	104	0.00
60	Diethylphthalate	40.000	39.961	0.1	100	0.00
61	4-Chlorophenyl-phenylether	40.000	39.340	1.6	99	0.00
62	4-Nitroaniline	40.000	38.781	3.0	99	0.00
63	Azobenzene	40.000	39.269	1.8	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.662	0.8	99	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.006	2.5	100	0.00
67	4-Bromophenyl-phenylether	40.000	40.286	-0.7	102	0.00
68	Hexachlorobenzene	40.000	39.523	1.2	100	0.00
69	Atrazine	40.000	33.158	17.1	103	0.00
70 C	Pentachlorophenol	40.000	42.636	-6.6	100	0.00
71	Phenanthrene	40.000	38.779	3.1	100	0.00
72	Anthracene	40.000	38.972	2.6	99	0.00
73	Carbazole	40.000	37.979	5.1	99	0.00
74	Di-n-butylphthalate	40.000	37.886	5.3	99	0.00
75 C	Fluoranthene	40.000	36.740	8.1	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
77	Benzidine	40.000	17.476	56.3#	51	0.00
78	Pyrene	40.000	40.958	-2.4	98	0.00
79 S	Terphenyl-d14	80.000	81.429	-1.8	98	0.00
80	Butylbenzylphthalate	40.000	40.030	-0.1	98	0.00
81	Benzo(a)anthracene	40.000	39.817	0.5	98	0.00
82	3,3'-Dichlorobenzidine	40.000	39.227	1.9	98	0.00
83	Chrysene	40.000	39.845	0.4	98	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.124	-0.3	99	0.00
85 c	Di-n-octyl phthalate	40.000	40.606	-1.5	100	0.00
86 I	Perylene-d12	20.000	20.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.196	2.0	100	0.00
88	Benzo(b)fluoranthene	40.000	39.603	1.0	102	0.00
89	Benzo(k)fluoranthene	40.000	39.303	1.7	99	0.00
90 C	Benzo(a)pyrene	40.000	39.098	2.3	101	0.00
91	Dibenzo(a,h)anthracene	40.000	39.186	2.0	100	0.00
92	Benzo(g,h,i)perylene	40.000	38.871	2.8	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103124\
Data File : BG063137.D
Acq On : 31 Oct 2024 17:38
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG103124

Quant Time: Nov 01 05:14:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 01 05:13:40 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0