


Data Path : Z:\HPCHEM1\BNA G\DATA\BG110116\
 Data File : BG024590.D
 Acq On : 1 Nov 2016 13:24
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG102516.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Oct 26 11:14:13 2016



AutoFind: Scans 2583, 2584, 2585; Background Corrected with Scan 2576

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.9	15563	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	33.1	17819	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	36.2	19459	PASS
197	198	0.00	2	0.4	213	PASS
198	198	100	100	100.0	53808	PASS
199	198	5	9	5.3	2830	PASS
275	198	10	60	33.2	17879	PASS
365	198	1	100	5.5	2969	PASS
441	442	0.01	24	15.0	7941	PASS
442	198	50	100	98.2	52856	PASS
443	442	15	24	17.9	9485	PASS