


Data Path : Z:\HPCHEM1\BNA G\DATA\BG110116\
 Data File : BG024608.D
 Acq On : 2 Nov 2016 1:35
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG102516.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Oct 26 11:14:13 2016



AutoFind: Scans 2582, 2583, 2584; Background Corrected with Scan 2576

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.3	19606	PASS
68	69	0.00	2	0.8	175	PASS
69	198	0.00	100	38.2	21843	PASS
70	69	0.00	2	0.8	179	PASS
127	198	10	80	38.4	21997	PASS
197	198	0.00	2	0.4	257	PASS
198	198	100	100	100.0	57240	PASS
199	198	5	9	7.7	4379	PASS
275	198	10	60	33.0	18902	PASS
365	198	1	100	4.0	2308	PASS
441	442	0.01	24	16.6	7864	PASS
442	198	50	100	82.9	47480	PASS
443	442	15	24	20.0	9499	PASS