

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG110117\
 Data File : BG030643.D
 Acq On : 1 Nov 2017 20:46
 Operator : SJ/JU
 Sample : I6203-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 102717-TIDO-E-D-B

Quant Time: Nov 02 06:32:59 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 01 17:41:24 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.16	152	63791	20.00	ng	0.00
21) Naphthalene-d8	10.98	136	306082	20.00	ng	0.00
38) Acenaphthene-d10	14.77	164	208332	20.00	ng	0.00
63) Phenanthrene-d10	17.51	188	471121	20.00	ng	0.00
75) Chrysene-d12	21.79	240	535158	20.00	ng	0.00
86) Perylene-d12	25.09	264	531079	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.72	112	199868	52.34	ng	0.00
7) Phenol-d6	7.31	99	187890	35.05	ng	0.00
23) Nitrobenzene-d5	9.32	82	482710	100.22	ng	0.00
41) 2,4,6-Tribromophenol	16.26	330	263642	98.02	ng	0.00
44) 2-Fluorobiphenyl	13.40	172	1366348	96.19	ng	0.00
78) Terphenyl-d14	20.11	244	2152593	91.51	ng	0.00
Target Compounds						
51) Acenaphthene	14.84	154	26955	2.027	ng	96
74) Fluoranthene	19.55	202	57137	2.007	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG110117\
Data File : BG030643.D
Acq On : 1 Nov 2017 20:46
Operator : SJ/JU
Sample : I6203-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
102717-TIDO-E-D-B

Quant Time: Nov 02 06:32:59 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 01 17:41:24 2017
Response via : Initial Calibration

