

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG110117\
 Data File : BG030650.D
 Acq On : 2 Nov 2017 1:09
 Operator : SJ/JU
 Sample : I6204-04
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 102617-TIDO-E-U-S

Integration Parameters: rteint.p
 Integrator: RTE
 Smoothing : ON Filtering: 5
 Sampling : 1 Min Area: 3 % of largest Peak
 Start Thrs: 0.2 Max Peaks: 100
 Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
 Peak separation: 5

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG101617.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	5.238	327	332	339	rBV	69957	107903	1.68%	0.435%
2	5.714	407	413	424	rBV	406380	659060	10.28%	2.659%
3	7.312	675	685	701	rBV	290819	539713	8.42%	2.178%
4	7.688	741	749	760	rBV	747075	1317976	20.56%	5.318%
5	8.158	822	829	836	rBV	222783	376101	5.87%	1.518%
6	8.487	876	885	894	rBV	1058530	1856522	28.96%	7.491%
7	9.322	1018	1027	1037	rBV	841710	1528692	23.85%	6.168%
8	10.973	1298	1308	1319	rBV	331735	602437	9.40%	2.431%
9	13.399	1713	1721	1735	rBV	2782650	4286945	66.88%	17.298%
10	14.774	1947	1955	1962	rBV2	589395	987236	15.40%	3.984%
11	16.119	2178	2184	2191	rBV	46041	70652	1.10%	0.285%
12	16.255	2200	2207	2220	rBV	1379018	2128945	33.21%	8.591%
13	17.512	2414	2421	2434	rBV2	750468	1159087	18.08%	4.677%
14	20.103	2856	2862	2868	rBV	4682584	6410230	100.00%	25.866%
15	21.783	3141	3148	3158	rBV2	835842	1384800	21.60%	5.588%
16	25.097	3701	3712	3728	rBV2	450976	1366196	21.31%	5.513%


Sum of corrected areas: 24782495

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG110117\
Data File : BG030650.D
Acq On : 2 Nov 2017 1:09
Operator : SJ/JU
Sample : I6204-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
102617-TIDO-E-U-S

Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST11.L
TIC Integration Parameters: LSCINT.P



Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA G\DATA\BG110117\
Data File : BG030650.D
Acq On : 2 Nov 2017 1:09
Operator : SJ/JU
Sample : I6204-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
102617-TIDO-E-U-S










Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 1 2-Pentanone, 4-hydroxy-4-me... Concentration Rank 3

R.T.	EstConc	Area	Relative to ISTD	R.T.
5.24	5.74 ng	107903	1,4-Dichlorobenzene-d4	8.16

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			2-Pentanone, 4-hydroxy-4-methyl-	116	C6H12O2	000123-42-2	42
2			2,3-Butanedione, monooxime	101	C4H7NO2	000057-71-6	27
3			Acetic acid, 1,1-dimethylethyl e...	116	C6H12O2	000540-88-5	25
4			Propane, 1-ethoxy-2-methyl-	102	C6H14O	000627-02-1	23
5			Butane, 1-ethoxy-	102	C6H14O	000628-81-9	10



Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG110117\
Data File : BG030650.D
Acq On : 2 Nov 2017 1:09
Operator : SJ/JU
Sample : I6204-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
102617-TIDO-E-U-S










Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 2 unknown7.69 Concentration Rank 2

R.T.	EstConc	Area	Relative to ISTD	R.T.
7.69	70.09 ng	1317980	1,4-Dichlorobenzene-d4	8.16

Hit#	of	Tentative ID	MW	MolForm	CAS#	Qual
1	5	3-Chloro-6-fluoro-pyrazine	132	C4H2ClFN2	1000146-10-7	27
2		5-Fluoro-2-chloropyrimidine	132	C4H2ClFN2	062802-42-0	16
3		(E)-3-Chloro-2-methyl-2-pentenal	132	C6H9ClO	031357-76-3	16
4		5-Aminoindole	132	C8H8N2	005192-03-0	12
5		4-Chloro-2-fluoro-pyrimidine	132	C4H2ClFN2	051422-00-5	9



Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_G\DATA\BG110117\
Data File : BG030650.D
Acq On : 2 Nov 2017 1:09
Operator : SJ/JU
Sample : I6204-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
102617-TIDO-E-U-S

Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST11.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc
2-Pentanone, 4-hy...	5.24	5.7	ng	107903	1	8.16	376101	20.0
unknown7.69	7.69	70.1	ng	1317980	1	8.16	376101	20.0