

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110121\
 Data File : BG050794.D
 Acq On : 1 Nov 2021 11:41
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040

Quant Time: Nov 01 12:36:54 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 01 12:35:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.241	152	37252	20.000	ng	0.00
21) Naphthalene-d8	11.062	136	159093	20.000	ng	# 0.00
39) Acenaphthene-d10	14.863	164	104720	20.000	ng	0.00
64) Phenanthrene-d10	17.607	188	237777	20.000	ng	# 0.00
76) Chrysene-d12	21.908	240	198926	20.000	ng	# 0.00
86) Perylene-d12	25.310	264	199527	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.774	112	199512	80.097	ng	0.00
7) Phenol-d6	7.384	99	292268	81.042	ng	0.00
23) Nitrobenzene-d5	9.411	82	285038	74.167	ng	0.00
42) 2,4,6-Tribromophenol	16.350	330	82358	82.453	ng	0.00
45) 2-Fluorobiphenyl	13.482	172	541717	77.855	ng	0.00
79) Terphenyl-d14	20.192	244	860816	84.332	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.629	88	44107	34.190	ng	92
3) Pyridine	4.040	79	127564	36.196	ng	# 93
4) n-Nitrosodimethylamine	3.958	42	58224	35.058	ng	# 43
6) Aniline	7.554	93	164781	39.327	ng	94
8) 2-Chlorophenol	7.801	128	100746	42.008	ng	92
9) Benzaldehyde	7.372	77	91984	40.520	ng	88
10) Phenol	7.413	94	142368	40.602	ng	83
11) bis(2-Chloroethyl)ether	7.648	93	107369	39.232	ng	85
12) 1,3-Dichlorobenzene	8.124	146	108876	39.327	ng	97
13) 1,4-Dichlorobenzene	8.277	146	110489	39.587	ng	94
14) 1,2-Dichlorobenzene	8.594	146	105980	39.754	ng	96
15) Benzyl Alcohol	8.471	79	111663	39.181	ng	91
16) 2,2'-oxybis(1-Chloropr...	8.759	45	167907	37.619	ng	86
17) 2-Methylphenol	8.670	107	94667	41.031	ng	90
18) Hexachloroethane	9.328	117	41730	39.505	ng	92
19) n-Nitroso-di-n-propyla...	9.041	70	98315	37.403	ng	# 90
20) 3+4-Methylphenols	8.999	107	131618	40.529	ng	94
22) Acetophenone	9.064	105	168310	37.212	ng	# 96
24) Nitrobenzene	9.458	77	138788	36.319	ng	93
25) Isophorone	9.975	82	249753	36.638	ng	# 96
26) 2-Nitrophenol	10.169	139	59090	42.112	ng	# 89
27) 2,4-Dimethylphenol	10.216	122	93483	38.562	ng	89
28) bis(2-Chloroethoxy)met...	10.451	93	145598	37.355	ng	93
29) 2,4-Dichlorophenol	10.703	162	98075	39.692	ng	95
30) 1,2,4-Trichlorobenzene	10.921	180	108418	38.627	ng	95
31) Naphthalene	11.115	128	330439	38.803	ng	97
32) Benzoic acid	10.363	122	61715	34.301	ng	# 79
33) 4-Chloroaniline	11.220	127	142816	39.972	ng	95
34) Hexachlorobutadiene	11.385	225	65799	37.341	ng	96
35) Caprolactam	11.996	113	38373	40.578	ng	# 73
36) 4-Chloro-3-methylphenol	12.319	107	122118	39.541	ng	# 90
37) 2-Methylnaphthalene	12.701	142	226737	38.589	ng	93
38) 1-Methylnaphthalene	12.924	142	219952	38.604	ng	91
40) 1,2,4,5-Tetrachloroben...	13.065	216	118826	38.425	ng	97
41) Hexachlorocyclopentadiene	13.036	237	79584	44.529	ng	91
43) 2,4,6-Trichlorophenol	13.300	196	86978	39.914	ng	92

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110121\
 Data File : BG050794.D
 Acq On : 1 Nov 2021 11:41
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040


Quant Time: Nov 01 12:36:54 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 01 12:35:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.377	196	89610	39.332	ng	91
46) 1,1'-Biphenyl	13.694	154	297996	39.013	ng	94
47) 2-Chloronaphthalene	13.747	162	233167	39.755	ng	99
48) 2-Nitroaniline	13.947	65	94382	40.453	ng	94
49) Acenaphthylene	14.587	152	380030	39.545	ng	97
50) Dimethylphthalate	14.305	163	311464	39.349	ng	99
51) 2,6-Dinitrotoluene	14.434	165	65254	41.159	ng	85
52) Acenaphthene	14.928	154	223095	36.126	ng	93
53) 3-Nitroaniline	14.763	138	78611	41.918	ng	92
54) 2,4-Dinitrophenol	14.975	184	35145	35.733	ng	# 86
55) Dibenzofuran	15.257	168	374132	39.172	ng	98
56) 4-Nitrophenol	15.063	139	59095	39.174	ng	# 76
57) 2,4-Dinitrotoluene	15.216	165	93749	41.952	ng	94
58) Fluorene	15.903	166	290314	39.573	ng	97
59) 2,3,4,6-Tetrachlorophenol	15.398	232	73399	36.984	ng	94
60) Diethylphthalate	15.656	149	329366	39.603	ng	98
61) 4-Chlorophenyl-phenyle...	15.891	204	157727	39.294	ng	97
62) 4-Nitroaniline	15.927	138	82524	42.296	ng	# 63
63) Azobenzene	16.185	77	369439	38.375	ng	81
65) 4,6-Dinitro-2-methylph...	15.979	198	58018	40.624	ng	90
66) n-Nitrosodiphenylamine	16.103	169	263157	38.950	ng	97
67) 4-Bromophenyl-phenylether	16.784	248	95942	40.042	ng	96
68) Hexachlorobenzene	16.908	284	93094	39.092	ng	94
69) Atrazine	17.043	200	102519	38.520	ng	97
70) Pentachlorophenol	17.254	266	59979	36.998	ng	96
71) Phenanthrene	17.648	178	489773	39.088	ng	97
72) Anthracene	17.742	178	488917	39.267	ng	97
73) Carbazole	18.006	167	492580	39.694	ng	99
74) Di-n-butylphthalate	18.541	149	594901	40.850	ng	# 97
75) Fluoranthene	19.646	202	585206	37.994	ng	95
77) Benzidine	19.822	184	239145	37.092	ng	99
78) Pyrene	20.010	202	589405	41.761	ng	97
80) Butylbenzylphthalate	20.874	149	256346	42.940	ng	98
81) Benzo(a)anthracene	21.884	228	532131	40.427	ng	99
82) 3,3'-Dichlorobenzidine	21.790	252	181123	41.521	ng	# 97
83) Chrysene	21.955	228	504229	40.725	ng	96
84) Bis(2-ethylhexyl)phtha...	21.755	149	357903	42.193	ng	# 97
85) Di-n-octyl phthalate	23.030	149	605534	42.800	ng	98
87) Indeno(1,2,3-cd)pyrene	29.235	276	559939	40.292	ng	# 88
88) Benzo(b)fluoranthene	24.223	252	490717	40.158	ng	# 93
89) Benzo(k)fluoranthene	24.293	252	487640	40.564	ng	# 96
90) Benzo(a)pyrene	25.151	252	420694	40.492	ng	# 94
91) Dibenzo(a,h)anthracene	29.305	278	463783	40.348	ng	# 94
92) Benzo(g,h,i)perylene	30.468	276	452575	40.201	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 01 12:36:54 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 01 12:35:42 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.241 min Scan# 80
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 37252
Ion Ratio Lower Upper
152 100
150 158.7 119.4 179.0
115 65.6 60.2 90.4



#2
1,4-Dioxane
Concen: 34.190 ng
RT: 3.629 min Scan# 24
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion: 88 Resp: 44107
Ion Ratio Lower Upper
88 100
58 77.3 55.7 83.5
43 33.0 29.3 43.9









#3
Pyridine
Concen: 36.196 ng
RT: 4.040 min Scan# 94
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 127564
Ion Ratio Lower Upper
79 100
52 71.3 58.4 87.6
51 36.2 37.0 55.6#



#4
n-Nitrosodimethylamine
Concen: 35.058 ng
RT: 3.958 min Scan# 80
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion: 42 Resp: 58224
Ion Ratio Lower Upper
42 100
74 132.3 64.0 96.0#
44 7.0 2.7 4.1#









#5
2-Fluorophenol
Concen: 80.097 ng
RT: 5.774 min Scan# 31
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 199512
Ion Ratio Lower Upper
112 100
64 68.8 46.7 70.1
63 37.9 35.3 52.9



#6
Aniline
Concen: 39.327 ng
RT: 7.554 min Scan# 692
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion: 93 Resp: 164781
Ion Ratio Lower Upper
93 100
66 41.9 30.6 45.8
65 21.6 15.0 22.4









#7
Phenol-d6
Concen: 81.042 ng
RT: 7.384 min Scan# 60
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion: 99 Resp: 292268
Ion Ratio Lower Upper
99 100
42 19.5 20.5 30.7#
71 37.8 41.5 62.3#



#8
2-Chlorophenol
Concen: 42.008 ng
RT: 7.801 min Scan# 734
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion: 128 Resp: 100746
Ion Ratio Lower Upper
128 100
130 31.9 10.0 50.0
64 55.3 28.1 68.1









#9
Benzaldehyde
Concen: 40.520 ng
RT: 7.372 min Scan# 661
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 91984
Ion Ratio Lower Upper
77 100
105 83.2 68.8 108.8
106 84.3 47.9 87.9



#10
Phenol
Concen: 40.602 ng
RT: 7.413 min Scan# 668
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion: 94 Resp: 142368
Ion Ratio Lower Upper
94 100
65 32.0 12.9 52.9
66 40.2 39.4 79.4









#11
bis(2-Chloroethyl)ether
Concen: 39.232 ng
RT: 7.648 min Scan# 708
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 107369
Ion Ratio Lower Upper
93 100
63 85.3 50.4 90.4
95 30.5 15.3 55.3



#12
1,3-Dichlorobenzene
Concen: 39.327 ng
RT: 8.124 min Scan# 789
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:146 Resp: 108876
Ion Ratio Lower Upper
146 100
148 65.6 52.6 78.8
75 40.1 35.7 53.5









#13
1,4-Dichlorobenzene
Concen: 39.587 ng
RT: 8.277 min Scan# 81
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 110489
Ion Ratio Lower Upper
146 100
148 61.5 54.0 81.0
111 44.5 34.3 51.5



#14
1,2-Dichlorobenzene
Concen: 39.754 ng
RT: 8.594 min Scan# 869
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:146 Resp: 105980
Ion Ratio Lower Upper
146 100
148 63.0 47.5 71.3
111 47.1 39.1 58.7









#15
Benzyl Alcohol
Concen: 39.181 ng
RT: 8.471 min Scan# 84
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 111663




Ion	Ratio	Lower	Upper
79	100		
108	67.1	48.2	72.4
77	64.3	57.3	85.9




#16
2,2'-oxybis(1-Chloropropane)
Concen: 37.619 ng
RT: 8.759 min Scan# 897
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion: 45 Resp: 167907

Ion	Ratio	Lower	Upper
45	100		
77	14.2	2.2	42.2
79	10.5	0.0	34.3









#17
2-Methylphenol
Concen: 41.031 ng
RT: 8.670 min Scan# 885
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:107 Resp: 94667
Ion Ratio Lower Upper
107 100
108 108.4 84.6 127.0
77 51.0 50.7 76.1
79 49.8 49.4 74.0



#18
Hexachloroethane
Concen: 39.505 ng
RT: 9.328 min Scan# 994
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:117 Resp: 41730
Ion Ratio Lower Upper
117 100
119 94.8 74.4 111.6
201 96.9 88.6 133.0









#19
n-Nitroso-di-n-propylamine
Concen: 37.403 ng
RT: 9.041 min Scan# 949
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:	70	Resp:	98315
Ion Ratio	Lower	Upper	
70	100		
42	59.4	54.5	81.7
101	9.1	9.0	13.6
130	16.4	17.0	25.6



#20
3+4-Methylphenols
Concen: 40.529 ng
RT: 8.999 min Scan# 938
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:	107	Resp:	131618
Ion Ratio	Lower	Upper	
107	100		
108	86.0	69.3	109.3
77	37.8	25.2	65.2
79	32.5	15.1	55.1









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.062 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:	136	Resp:	159093
Ion Ratio	Lower	Upper	
136	100		
137	10.8	8.2	12.2
54	9.6	5.5	8.3#
68	7.7	2.3	3.5#



#22
Acetophenone
Concen: 37.212 ng
RT: 9.064 min Scan# 949
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:	105	Resp:	168310
Ion Ratio	Lower	Upper	
105	100		
71	0.9	0.6	0.8#
51	30.0	26.6	40.0
120	21.4	16.2	24.4





#23

Nitrobenzene-d5

Concen: 74.167 ng

RT: 9.411 min Scan# 1008

Delta R.T. 0.000 min

Lab File: BG050794.D

Acq: 1 Nov 2021 11:41

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040





Tgt Ion: 82 Resp: 285038

Ion Ratio Lower Upper

82 100

128 37.8 29.8 44.6

54 50.4 45.8 68.8



#24

Nitrobenzene

Concen: 36.319 ng

RT: 9.458 min Scan# 1016

Delta R.T. 0.000 min

Lab File: BG050794.D

Acq: 1 Nov 2021 11:41




Tgt Ion: 77 Resp: 138788


Ion Ratio Lower Upper

77 100

123 39.3 27.5 41.3

65 13.6 12.6 18.8









#25
Isophorone
Concen: 36.638 ng
RT: 9.975 min Scan# 1108
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 249753
Ion Ratio Lower Upper
82 100
95 7.3 8.5 12.7#
138 14.7 11.8 17.6



#26
2-Nitrophenol
Concen: 42.112 ng
RT: 10.169 min Scan# 1137
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:139 Resp: 59090
Ion Ratio Lower Upper
139 100
109 34.8 37.8 56.6#
65 60.9 51.8 77.8





#27

2,4-Dimethylphenol

Concen: 38.562 ng

RT: 10.216 min Scan# 1145

Delta R.T. 0.000 min

Lab File: BG050794.D

Acq: 1 Nov 2021 11:41

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040





Tgt Ion:122 Resp: 93483

Ion Ratio Lower Upper

122 100

107 121.9 110.1 165.1

121 56.4 49.0 73.6



#28

bis(2-Chloroethoxy)methane

Concen: 37.355 ng

RT: 10.451 min Scan# 1185

Delta R.T. 0.000 min

Lab File: BG050794.D

Acq: 1 Nov 2021 11:41




Tgt Ion: 93 Resp: 145598


Ion Ratio Lower Upper

93 100

95 32.6 22.2 33.4

123 10.3 8.8 13.2









#29
2,4-Dichlorophenol
Concen: 39.692 ng
RT: 10.703 min Scan# 1228
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 98075
Ion Ratio Lower Upper
162 100
164 65.1 39.4 79.4
98 40.0 18.7 58.7



#30
1,2,4-Trichlorobenzene
Concen: 38.627 ng
RT: 10.921 min Scan# 1265
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:180 Resp: 108418
Ion Ratio Lower Upper
180 100
182 93.3 78.9 118.3
145 30.7 26.2 39.4









#31
Naphthalene
Concen: 38.803 ng
RT: 11.115 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 330439
Ion Ratio Lower Upper
128 100
129 11.1 8.5 12.7
127 14.2 10.2 15.4



#32
Benzoic acid
Concen: 34.301 ng
RT: 10.363 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:122 Resp: 61715
Ion Ratio Lower Upper
122 100
105 130.2 92.3 132.3
77 111.6 66.5 106.5#









#33
4-Chloroaniline
Concen: 39.972 ng
RT: 11.220 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:127	Resp: 142816
Ion Ratio	Lower Upper
127 100	
129 31.5	26.7 40.1
65 36.3	32.8 49.2
92 20.0	16.2 24.4



#34
Hexachlorobutadiene
Concen: 37.341 ng
RT: 11.385 min Scan# 1344
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt	Ion:225	Resp:	65799
Ion	Ratio	Lower	Upper
225	100		
223	62.8	49.8	74.8
227	63.7	46.4	69.6









#35
 Caprolactam
 Concen: 40.578 ng
 RT: 11.996 min Scan# 14
 Delta R.T. 0.000 min
 Lab File: BG050794.D
 Acq: 1 Nov 2021 11:41




Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040


Tgt Ion:113 Resp: 38373
 Ion Ratio Lower Upper
 113 100
 55 183.5 113.8 153.8#
 56 138.7 105.6 145.6



#36
 4-Chloro-3-methylphenol
 Concen: 39.541 ng
 RT: 12.319 min Scan# 1503
 Delta R.T. 0.000 min
 Lab File: BG050794.D
 Acq: 1 Nov 2021 11:41

Tgt Ion:107 Resp: 122118
 Ion Ratio Lower Upper
 107 100
 144 23.2 24.8 37.2#
 142 71.1 62.2 93.4





#37

2-Methylnaphthalene

Concen: 38.589 ng

RT: 12.701 min Scan# 1

Delta R.T. 0.000 min

Lab File: BG050794.D

Acq: 1 Nov 2021 11:41

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040





Tgt Ion:142 Resp: 226737

Ion Ratio Lower Upper

142 100

141 91.6 68.8 103.2

115 39.7 36.4 54.6



#38

1-Methylnaphthalene

Concen: 38.604 ng

RT: 12.924 min Scan# 1606

Delta R.T. 0.000 min

Lab File: BG050794.D

Acq: 1 Nov 2021 11:41




Tgt Ion:142 Resp: 219952


Ion Ratio Lower Upper

142 100

141 97.3 70.6 106.0

116 4.3 4.0 6.0









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.863 min Scan# 1936
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:164 Resp: 104720
Ion Ratio Lower Upper
164 100
162 107.0 81.4 122.0
160 44.0 34.9 52.3



#40
1,2,4,5-Tetrachlorobenzene
Concen: 38.425 ng
RT: 13.065 min Scan# 1630
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41


Tgt Ion:216 Resp: 118826
Ion Ratio Lower Upper
216 100
214 79.8 61.4 92.2
179 21.0 17.3 25.9
108 20.6 16.3 24.5








#41
Hexachlorocyclopentadiene
Concen: 44.529 ng
RT: 13.036 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040







Tgt Ion:237 Resp: 79584
Ion Ratio Lower Upper
237 100
235 64.6 36.6 76.6
272 12.3 0.0 32.5



#42
2,4,6-Tribromophenol
Concen: 82.453 ng
RT: 16.350 min Scan# 2189
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:330 Resp: 82358
Ion Ratio Lower Upper
330 100
332 96.8 77.3 115.9
141 51.3 35.6 53.4









#43
2,4,6-Trichlorophenol
Concen: 39.914 ng
RT: 13.300 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:196 Resp: 86978
Ion Ratio Lower Upper
196 100
198 92.3 82.1 123.1
200 30.2 23.8 35.8



#44
2,4,5-Trichlorophenol
Concen: 39.332 ng
RT: 13.377 min Scan# 1683
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:196 Resp: 89610
Ion Ratio Lower Upper
196 100
198 94.3 87.4 131.0
97 52.5 43.3 64.9
132 26.8 21.4 32.2









#45
2-Fluorobiphenyl
Concen: 77.855 ng
RT: 13.482 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:172 Resp: 541717
Ion Ratio Lower Upper
172 100
171 37.5 28.6 43.0
170 25.9 18.6 27.8



#46
1,1'-Biphenyl
Concen: 39.013 ng
RT: 13.694 min Scan# 1737
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:154 Resp: 297996
Ion Ratio Lower Upper
154 100
153 42.4 20.5 60.5
76 16.9 0.0 29.3









#47
2-Chloronaphthalene
Concen: 39.755 ng
RT: 13.747 min Scan# 1746
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 233167
Ion Ratio Lower Upper
162 100
127 40.5 32.7 49.1
164 33.4 27.0 40.4



#48
2-Nitroaniline
Concen: 40.453 ng
RT: 13.947 min Scan# 1780
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion: 65 Resp: 94382
Ion Ratio Lower Upper
65 100
92 63.3 47.2 70.8
138 90.8 68.0 102.0









#49
Acenaphthylene
Concen: 39.545 ng
RT: 14.587 min Scan# 1889
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 380030
Ion Ratio Lower Upper
152 100
151 21.0 18.2 27.4
153 13.8 11.7 17.5



#50
Dimethylphthalate
Concen: 39.349 ng
RT: 14.305 min Scan# 1841
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:163 Resp: 311464
Ion Ratio Lower Upper
163 100
194 3.5 3.0 4.6
164 10.4 8.6 12.8









#51
2,6-Dinitrotoluene
Concen: 41.159 ng
RT: 14.434 min Scan# 1863
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 65254
Ion Ratio Lower Upper
165 100
63 66.0 63.3 94.9
89 66.3 63.4 95.0



#52
Acenaphthene
Concen: 36.126 ng
RT: 14.928 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:154 Resp: 223095
Ion Ratio Lower Upper
154 100
153 116.0 87.0 130.4
152 55.6 40.1 60.1









#53
3-Nitroaniline
Concen: 41.918 ng
RT: 14.763 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 78611
Ion Ratio Lower Upper
138 100
108 10.4 8.3 12.5
92 130.0 96.4 144.6



#54
2,4-Dinitrophenol
Concen: 35.733 ng
RT: 14.975 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:184 Resp: 35145
Ion Ratio Lower Upper
184 100
63 76.7 68.3 102.5
154 62.0 62.6 93.8#









#55
Dibenzofuran
Concen: 39.172 ng
RT: 15.257 min Scan# 20
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 374132
Ion Ratio Lower Upper
168 100
139 40.9 33.4 50.2
169 14.0 12.3 18.5



#56
4-Nitrophenol
Concen: 39.174 ng
RT: 15.063 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:139 Resp: 59095
Ion Ratio Lower Upper
139 100
109 86.1 116.3 156.3#
65 119.6 105.0 145.0





#57

2,4-Dinitrotoluene

Concen: 41.952 ng

RT: 15.216 min Scan# 1998

Delta R.T. 0.000 min

Lab File: BG050794.D

Acq: 1 Nov 2021 11:41

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040

Tgt Ion:165 Resp: 93749





Ion Ratio Lower Upper

165 100

63 51.6 48.0 72.0

89 84.3 65.4 98.0

182 2.5 2.0 3.0



#58

Fluorene

Concen: 39.573 ng

RT: 15.903 min Scan# 2113

Delta R.T. 0.000 min

Lab File: BG050794.D

Acq: 1 Nov 2021 11:41




Tgt Ion:166 Resp: 290314


Ion Ratio Lower Upper

166 100

165 100.1 77.7 116.5

167 13.8 10.8 16.2









#59
2,3,4,6-Tetrachlorophenol
Concen: 36.984 ng
RT: 15.398 min Scan# 2043
Delta R.T. -0.082 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:232 Resp: 73399
Ion Ratio Lower Upper
232 100
131 43.1 37.2 55.8
130 3.0 2.8 4.2
166 26.0 24.0 36.0



#60
Diethylphthalate
Concen: 39.603 ng
RT: 15.656 min Scan# 2071
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:149 Resp: 329366
Ion Ratio Lower Upper
149 100
177 22.4 19.0 28.6
150 12.7 10.2 15.4









#61
4-Chlorophenyl-phenylether
Concen: 39.294 ng
RT: 15.891 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 157727
Ion Ratio Lower Upper
204 100
206 34.4 24.6 37.0
141 58.8 47.8 71.8



#62
4-Nitroaniline
Concen: 42.296 ng
RT: 15.927 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:138 Resp: 82524
Ion Ratio Lower Upper
138 100
92 62.2 32.3 72.3
108 95.1 132.4 172.4#









#63
Azobenzene
Concen: 38.375 ng
RT: 16.185 min Scan# 2164
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 369439
Ion Ratio Lower Upper
77 100
182 21.0 7.9 47.9
105 19.1 0.0 37.9
51 28.8 27.3 67.3



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.607 min Scan# 2403
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:188 Resp: 237777
Ion Ratio Lower Upper
188 100
94 9.7 5.4 8.0#
80 11.9 3.7 5.5#









#65
4,6-Dinitro-2-methylphenol
Concen: 40.624 ng
RT: 15.979 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 58018
Ion Ratio Lower Upper
198 100
51 50.6 29.4 69.4
105 47.4 15.1 55.1



#66
n-Nitrosodiphenylamine
Concen: 38.950 ng
RT: 16.103 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:169 Resp: 263157
Ion Ratio Lower Upper
169 100
168 72.3 56.3 84.5
167 38.0 28.4 42.6









#67
4-Bromophenyl-phenylether
Concen: 40.042 ng
RT: 16.784 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 95942
Ion Ratio Lower Upper
248 100
250 100.1 76.1 114.1
141 79.9 65.4 98.2



#68
Hexachlorobenzene
Concen: 39.092 ng
RT: 16.908 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:284 Resp: 93094
Ion Ratio Lower Upper
284 100
142 40.4 27.0 40.6
249 35.1 28.2 42.2





#69

Atrazine

Concen: 38.520 ng

RT: 17.043 min Scan# 21

Delta R.T. 0.000 min

Lab File: BG050794.D

Acq: 1 Nov 2021 11:41

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040





Tgt Ion:200 Resp: 102519

Ion Ratio Lower Upper

200 100

173 25.9 3.7 43.7

215 48.5 26.9 66.9



#70

Pentachlorophenol

Concen: 36.998 ng

RT: 17.254 min Scan# 2343

Delta R.T. 0.000 min

Lab File: BG050794.D

Acq: 1 Nov 2021 11:41




Tgt Ion:266 Resp: 59979


Ion Ratio Lower Upper

266 100

268 65.7 49.0 73.4

264 62.9 49.0 73.6





#71

Phenanthrene

Concen: 39.088 ng

RT: 17.648 min Scan# 2410

Delta R.T. 0.000 min

Lab File: BG050794.D

Acq: 1 Nov 2021 11:41

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040





Tgt Ion:178 Resp: 489773

Ion Ratio Lower Upper

178 100

176 19.9 16.0 24.0

179 15.0 14.2 21.4



#72

Anthracene

Concen: 39.267 ng

RT: 17.742 min Scan# 2426

Delta R.T. 0.000 min

Lab File: BG050794.D

Acq: 1 Nov 2021 11:41




Tgt Ion:178 Resp: 488917


Ion Ratio Lower Upper

178 100

176 19.5 14.6 21.8

179 15.7 13.4 20.0









#73
 Carbazole
 Concen: 39.694 ng
 RT: 18.006 min Scan# 2474
 Delta R.T. 0.000 min
 Lab File: BG050794.D
 Acq: 1 Nov 2021 11:41




Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040


Tgt Ion:167 Resp: 492580
 Ion Ratio Lower Upper
 167 100
 166 22.7 17.4 26.2
 139 13.5 10.7 16.1



#74
 Di-n-butylphthalate
 Concen: 40.850 ng
 RT: 18.541 min Scan# 2562
 Delta R.T. 0.000 min
 Lab File: BG050794.D
 Acq: 1 Nov 2021 11:41

Tgt Ion:149 Resp: 594901
 Ion Ratio Lower Upper
 149 100
 150 9.7 7.5 11.3
 104 6.1 6.6 10.0#









#75
Fluoranthene
Concen: 37.994 ng
RT: 19.646 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:202 Resp: 585206
Ion Ratio Lower Upper
202 100
101 11.1 0.0 24.7
203 18.1 0.0 38.5



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.908 min Scan# 3135
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:240 Resp: 198926
Ion Ratio Lower Upper
240 100
120 9.7 4.2 6.2#
236 26.3 21.8 32.6









#77
Benzidine
Concen: 37.092 ng
RT: 19.822 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:184 Resp: 239145
Ion Ratio Lower Upper
184 100
185 14.2 10.7 16.1
183 12.0 9.4 14.0



#78
Pyrene
Concen: 41.761 ng
RT: 20.010 min Scan# 2812
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:202 Resp: 589405
Ion Ratio Lower Upper
202 100
200 21.9 15.8 23.8
203 18.4 14.6 21.8









#79
 Terphenyl-d14
 Concen: 84.332 ng
 RT: 20.192 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BG050794.D
 Acq: 1 Nov 2021 11:41




Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040


Tgt Ion:244 Resp: 860816
 Ion Ratio Lower Upper
 244 100
 212 8.3 6.9 10.3
 122 11.3 5.0 7.6#



#80
 Butylbenzylphthalate
 Concen: 42.940 ng
 RT: 20.874 min Scan# 2959
 Delta R.T. 0.000 min
 Lab File: BG050794.D
 Acq: 1 Nov 2021 11:41

Tgt Ion:149 Resp: 256346
 Ion Ratio Lower Upper
 149 100
 91 79.5 63.4 95.2
 206 19.0 18.4 27.6









#81
Benzo(a)anthracene
Concen: 40.427 ng
RT: 21.884 min Scan# 3115
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 532131
Ion Ratio Lower Upper
228 100
226 28.4 22.7 34.1
229 20.7 16.0 24.0



#82
3,3'-Dichlorobenzidine
Concen: 41.521 ng
RT: 21.790 min Scan# 3115
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:252 Resp: 181123
Ion Ratio Lower Upper
252 100
254 63.5 52.4 78.6
126 11.6 5.9 8.9#









#83
Chrysene
Concen: 40.725 ng
RT: 21.955 min Scan# 3109
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 504229
Ion Ratio Lower Upper
228 100
226 31.5 23.1 34.7
229 19.8 15.3 22.9



#84
Bis(2-ethylhexyl)phthalate
Concen: 42.193 ng
RT: 21.755 min Scan# 3109
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:149 Resp: 357903
Ion Ratio Lower Upper
149 100
167 26.2 22.3 33.5
279 3.6 3.8 5.8#









#85
Di-n-octyl phthalate
Concen: 42.800 ng
RT: 23.030 min Scan# 31
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 605534
Ion Ratio Lower Upper
149 100
167 1.6 1.4 2.2
43 11.5 8.6 12.8



#86
Perylene-d12
Concen: 20.000 ng
RT: 25.310 min Scan# 3714
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:264 Resp: 199527
Ion Ratio Lower Upper
264 100
260 24.5 19.4 29.2
265 20.8 16.9 25.3









#87
 Indeno(1,2,3-cd)pyrene
 Concen: 40.292 ng
 RT: 29.235 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BG050794.D
 Acq: 1 Nov 2021 11:41




Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040


Tgt Ion:276 Resp: 559939
 Ion Ratio Lower Upper
 276 100
 138 24.1 5.1 7.7#
 277 25.7 19.8 29.6



#88
 Benzo(b)fluoranthene
 Concen: 40.158 ng
 RT: 24.223 min Scan# 3529
 Delta R.T. 0.000 min
 Lab File: BG050794.D
 Acq: 1 Nov 2021 11:41

Tgt Ion:252 Resp: 490717
 Ion Ratio Lower Upper
 252 100
 253 22.2 16.2 24.4
 125 10.3 3.0 4.4#









#89
Benzo(k)fluoranthene
Concen: 40.564 ng
RT: 24.293 min Scan# 31
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 487640
Ion Ratio Lower Upper
252 100
253 21.7 18.0 27.0
125 9.9 3.1 4.7#



#90
Benzo(a)pyrene
Concen: 40.492 ng
RT: 25.151 min Scan# 3687
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:252 Resp: 420694
Ion Ratio Lower Upper
252 100
253 22.1 16.6 25.0
125 10.5 3.5 5.3#









#91
Dibenzo(a,h)anthracene
Concen: 40.348 ng
RT: 29.305 min Scan# 41
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 463783
Ion Ratio Lower Upper
278 100
139 15.1 4.6 6.8#
279 22.7 17.9 26.9



#92
Benzo(g,h,i)perylene
Concen: 40.201 ng
RT: 30.468 min Scan# 4592
Delta R.T. 0.000 min
Lab File: BG050794.D
Acq: 1 Nov 2021 11:41

Tgt Ion:276 Resp: 452575
Ion Ratio Lower Upper
276 100
277 24.2 18.4 27.6
138 19.6 7.0 10.6#

