

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110124\
 Data File : BG063140.D
 Acq On : 1 Nov 2024 9:41
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 01 13:33:27 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	40.552	-1.4	105	0.00
3	Pyridine	40.000	41.923	-4.8	106	0.00
4	n-Nitrosodimethylamine	40.000	43.480	-8.7	103	0.00
5 S	2-Fluorophenol	80.000	78.158	2.3	107	0.00
6	Aniline	40.000	41.015	-2.5	107	0.00
7 S	Phenol-d6	80.000	83.309	-4.1	109	0.00
8	2-Chlorophenol	40.000	42.166	-5.4	110	0.00
9	Benzaldehyde	40.000	42.529	-6.3	110	0.00
10 C	Phenol	40.000	41.706	-4.3	111	0.00
11	bis(2-Chloroethyl)ether	40.000	40.984	-2.5	109	0.00
12	1,3-Dichlorobenzene	40.000	39.581	1.0	105	0.00
13 C	1,4-Dichlorobenzene	40.000	40.498	-1.2	108	0.00
14	1,2-Dichlorobenzene	40.000	40.526	-1.3	108	0.00
15	Benzyl Alcohol	40.000	43.646	-9.1	112	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	43.129	-7.8	115	0.00
17	2-Methylphenol	40.000	42.796	-7.0	114	0.00
18	Hexachloroethane	40.000	40.644	-1.6	105	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.704	-9.3	115	0.00
20	3+4-Methylphenols	40.000	43.554	-8.9	113	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	115	0.00
22	Acetophenone	40.000	39.630	0.9	113	0.00
23 S	Nitrobenzene-d5	80.000	78.438	2.0	115	0.00
24	Nitrobenzene	40.000	39.388	1.5	116	0.00
25	Isophorone	40.000	39.370	1.6	112	0.00
26 C	2-Nitrophenol	40.000	38.861	2.8	114	0.00
27	2,4-Dimethylphenol	40.000	39.475	1.3	116	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.022	2.4	113	0.00
29 C	2,4-Dichlorophenol	40.000	38.424	3.9	110	0.00
30	1,2,4-Trichlorobenzene	40.000	37.866	5.3	109	0.00
31	Naphthalene	40.000	38.819	3.0	112	0.00
32	Benzoic acid	40.000	40.181	-0.5	112	0.00
33	4-Chloroaniline	40.000	39.912	0.2	113	0.00
34 C	Hexachlorobutadiene	40.000	36.747	8.1	105	0.00
35	Caprolactam	40.000	38.460	3.8	109	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.824	5.4	111	0.00
37	2-Methylnaphthalene	40.000	38.389	4.0	114	0.00
38	1-Methylnaphthalene	40.000	37.921	5.2	109	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	115	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.859	7.9	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.072	4.8	106	0.00
42 S	2,4,6-Tribromophenol	80.000	72.975	8.8	106	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.836	5.4	105	0.00
44	2,4,5-Trichlorophenol	40.000	38.051	4.9	106	0.00
45 S	2-Fluorobiphenyl	80.000	74.850	6.4	107	0.00
46	1,1'-Biphenyl	40.000	38.306	4.2	111	0.00
47	2-Chloronaphthalene	40.000	38.655	3.4	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110124\
 Data File : BG063140.D
 Acq On : 1 Nov 2024 9:41
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 01 13:33:27 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.244	1.9	108	0.00
49	Acenaphthylene	40.000	38.508	3.7	113	0.00
50	Dimethylphthalate	40.000	38.526	3.7	114	0.00
51	2,6-Dinitrotoluene	40.000	38.514	3.7	111	0.00
52 C	Acenaphthene	40.000	38.037	4.9	111	0.00
53	3-Nitroaniline	40.000	39.627	0.9	112	0.00
54 P	2,4-Dinitrophenol	40.000	39.505	1.2	110	0.00
55	Dibenzofuran	40.000	38.181	4.5	112	0.00
56 P	4-Nitrophenol	40.000	40.134	-0.3	114	0.00
57	2,4-Dinitrotoluene	40.000	39.336	1.7	114	0.00
58	Fluorene	40.000	38.059	4.9	110	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.375	4.1	110	0.00
60	Diethylphthalate	40.000	38.049	4.9	110	0.00
61	4-Chlorophenyl-phenylether	40.000	37.763	5.6	110	0.00
62	4-Nitroaniline	40.000	38.206	4.5	112	0.00
63	Azobenzene	40.000	38.448	3.9	112	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	112	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.563	3.6	109	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.418	4.0	111	0.00
67	4-Bromophenyl-phenylether	40.000	39.120	2.2	112	0.00
68	Hexachlorobenzene	40.000	38.260	4.4	109	0.00
69	Atrazine	40.000	32.239	19.4	113	0.00
70 C	Pentachlorophenol	40.000	40.018	-0.0	106	0.00
71	Phenanthrene	40.000	38.341	4.1	111	0.00
72	Anthracene	40.000	39.046	2.4	111	0.00
73	Carbazole	40.000	38.026	4.9	112	0.00
74	Di-n-butylphthalate	40.000	38.086	4.8	112	0.00
75 C	Fluoranthene	40.000	36.689	8.3	110	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	113	0.00
77	Benzydine	40.000	23.400	41.5#	78	0.00
78	Pyrene	40.000	39.687	0.8	109	0.00
79 S	Terphenyl-d14	80.000	78.605	1.7	108	0.00
80	Butylbenzylphthalate	40.000	40.038	-0.1	112	0.00
81	Benzo(a)anthracene	40.000	39.230	1.9	111	0.00
82	3,3'-Dichlorobenzidine	40.000	38.813	3.0	110	0.00
83	Chrysene	40.000	39.394	1.5	110	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.255	-0.6	113	0.00
85 c	Di-n-octyl phthalate	40.000	40.597	-1.5	114	0.00
86 I	Perylene-d12	20.000	20.000	0.0	112	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.508	1.2	112	0.00
88	Benzo(b)fluoranthene	40.000	39.341	1.6	112	0.00
89	Benzo(k)fluoranthene	40.000	40.489	-1.2	113	0.00
90 C	Benzo(a)pyrene	40.000	39.260	1.9	113	0.00
91	Dibenzo(a,h)anthracene	40.000	39.603	1.0	112	0.00
92	Benzo(g,h,i)perylene	40.000	38.393	4.0	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110124\
Data File : BG063140.D
Acq On : 1 Nov 2024 9:41
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 01 13:33:27 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 01 05:13:40 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0