

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110124\
 Data File : BG063148.D
 Acq On : 1 Nov 2024 15:54
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 01 16:31:41 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.507	0.389	23.3	98	0.00
3	Pyridine	1.223	1.033	15.5	105	0.00
4	n-Nitrosodimethylamine	0.760	0.651	14.3	100	0.00
5 S	2-Fluorophenol	1.290	1.168	9.5	123	0.00
6	Aniline	1.520	1.464	3.7	124	0.00
7 S	Phenol-d6	1.690	1.671	1.1	128	0.00
8	2-Chlorophenol	1.198	1.199	-0.1	129	0.00
9	Benzaldehyde	1.050	1.131	-7.7	138	0.00
10 C	Phenol	1.811	1.779	1.8	129	0.00
11	bis(2-Chloroethyl)ether	1.221	1.212	0.7	130	0.00
12	1,3-Dichlorobenzene	1.453	1.392	4.2	125	0.00
13 C	1,4-Dichlorobenzene	1.481	1.407	5.0	125	0.00
14	1,2-Dichlorobenzene	1.428	1.379	3.4	127	0.00
15	Benzyl Alcohol	1.358	1.393	-2.6	129	0.00
16	2,2'-oxybis(1-Chloropropane	2.215	2.253	-1.7	133	0.00
17	2-Methylphenol	1.152	1.102	4.3	126	0.00
18	Hexachloroethane	0.530	0.525	0.9	126	0.00
19 P	n-Nitroso-di-n-propylamine	1.177	1.194	-1.4	131	0.00
20	3+4-Methylphenols	1.536	1.538	-0.1	128	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	131	0.00
22	Acetophenone	0.536	0.535	0.2	129	0.00
23 S	Nitrobenzene-d5	0.460	0.451	2.0	131	0.00
24	Nitrobenzene	0.497	0.488	1.8	132	0.00
25	Isophorone	0.795	0.796	-0.1	130	0.00
26 C	2-Nitrophenol	0.183	0.178	2.7	130	0.00
27	2,4-Dimethylphenol	0.225	0.223	0.9	132	0.00
28	bis(2-Chloroethoxy)methane	0.417	0.421	-1.0	133	0.00
29 C	2,4-Dichlorophenol	0.340	0.331	2.6	127	0.00
30	1,2,4-Trichlorobenzene	0.418	0.402	3.8	126	0.00
31	Naphthalene	1.055	1.036	1.8	129	0.00
32	Benzoic acid	0.196	0.187	4.6	122	0.02
33	4-Chloroaniline	0.356	0.360	-1.1	130	0.00
34 C	Hexachlorobutadiene	0.344	0.325	5.5	124	0.00
35	Caprolactam	0.108	0.104	3.7	124	0.02
36 C	4-Chloro-3-methylphenol	0.391	0.376	3.8	129	0.00
37	2-Methylnaphthalene	0.748	0.747	0.1	136	0.00
38	1-Methylnaphthalene	0.737	0.730	0.9	130	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	132	0.00
40	1,2,4,5-Tetrachlorobenzene	0.719	0.675	6.1	125	0.00
41 P	Hexachlorocyclopentadiene	0.234	0.237	-1.3	131	0.00
42 S	2,4,6-Tribromophenol	0.340	0.316	7.1	125	0.00
43 C	2,4,6-Trichlorophenol	0.430	0.409	4.9	121	0.00
44	2,4,5-Trichlorophenol	0.473	0.462	2.3	126	0.00
45 S	2-Fluorobiphenyl	1.350	1.294	4.1	127	0.00
46	1,1'-Biphenyl	1.357	1.316	3.0	130	0.00
47	2-Chloronaphthalene	1.137	1.100	3.3	131	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110124\
 Data File : BG063148.D
 Acq On : 1 Nov 2024 15:54
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 01 16:31:41 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.399	0.389	2.5	124	0.00
49	Acenaphthylene	1.605	1.588	1.1	134	0.00
50	Dimethylphthalate	1.425	1.382	3.0	132	0.00
51	2,6-Dinitrotoluene	0.304	0.299	1.6	131	0.00
52 C	Acenaphthene	0.994	0.969	2.5	131	0.00
53	3-Nitroaniline	0.289	0.291	-0.7	132	0.00
54 P	2,4-Dinitrophenol	0.186	0.182	2.2	127	0.00
55	Dibenzofuran	1.752	1.675	4.4	129	0.00
56 P	4-Nitrophenol	0.219	0.220	-0.5	131	0.00
57	2,4-Dinitrotoluene	0.435	0.433	0.5	132	0.00
58	Fluorene	1.388	1.357	2.2	130	0.00
59	2,3,4,6-Tetrachlorophenol	0.454	0.447	1.5	131	0.00
60	Diethylphthalate	1.464	1.398	4.5	127	0.00
61	4-Chlorophenyl-phenylether	0.863	0.847	1.9	131	0.00
62	4-Nitroaniline	0.319	0.309	3.1	131	0.00
63	Azobenzene	1.514	1.476	2.5	131	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	127	0.00
65	4,6-Dinitro-2-methylphenol	0.129	0.125	3.1	124	0.00
66 c	n-Nitrosodiphenylamine	0.499	0.494	1.0	130	0.00
67	4-Bromophenyl-phenylether	0.245	0.245	0.0	129	0.00
68	Hexachlorobenzene	0.284	0.282	0.7	128	0.00
69	Atrazine	0.229	0.181	21.0	126	0.00
70 C	Pentachlorophenol	0.141	0.146	-3.5	124	0.00
71	Phenanthrene	0.970	0.931	4.0	126	0.00
72	Anthracene	0.952	0.915	3.9	124	0.00
73	Carbazole	0.896	0.860	4.0	129	0.00
74	Di-n-butylphthalate	1.017	0.962	5.4	127	0.00
75 C	Fluoranthene	1.372	1.268	7.6	126	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	126	0.00
77	Benzydine	0.308	0.270	12.3	130	0.00
78	Pyrene	1.186	1.192	-0.5	123	0.00
79 S	Terphenyl-d14	0.972	0.946	2.7	120	0.00
80	Butylbenzylphthalate	0.382	0.392	-2.6	128	0.00
81	Benzo(a)anthracene	1.259	1.211	3.8	121	0.00
82	3,3'-Dichlorobenzidine	0.467	0.468	-0.2	128	0.00
83	Chrysene	1.183	1.141	3.6	121	0.00
84	Bis(2-ethylhexyl)phthalate	0.554	0.557	-0.5	126	0.00
85 c	Di-n-octyl phthalate	0.927	0.940	-1.4	128	0.00
86 I	Perylene-d12	1.000	1.000	0.0	127	0.00
87	Indeno(1,2,3-cd)pyrene	1.342	1.313	2.2	125	0.00
88	Benzo(b)fluoranthene	1.166	1.151	1.3	127	0.00
89	Benzo(k)fluoranthene	1.112	1.080	2.9	123	0.00
90 C	Benzo(a)pyrene	1.018	0.998	2.0	127	0.00
91	Dibenzo(a,h)anthracene	1.120	1.092	2.5	125	0.00
92	Benzo(g,h,i)perylene	1.108	1.071	3.3	125	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110124\
Data File : BG063148.D
Acq On : 1 Nov 2024 15:54
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 01 16:31:41 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 01 05:13:40 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0