

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110124\
 Data File : BG063148.D
 Acq On : 1 Nov 2024 15:54
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 01 16:31:41 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	129	0.00
2	1,4-Dioxane	40.000	30.685	23.3	98	0.00
3	Pyridine	40.000	33.781	15.5	105	0.00
4	n-Nitrosodimethylamine	40.000	34.258	14.4	100	0.00
5 S	2-Fluorophenol	80.000	72.438	9.5	123	0.00
6	Aniline	40.000	38.531	3.7	124	0.00
7 S	Phenol-d6	80.000	79.098	1.1	128	0.00
8	2-Chlorophenol	40.000	40.032	-0.1	129	0.00
9	Benzaldehyde	40.000	43.068	-7.7	138	0.00
10 C	Phenol	40.000	39.294	1.8	129	0.00
11	bis(2-Chloroethyl)ether	40.000	39.710	0.7	130	0.00
12	1,3-Dichlorobenzene	40.000	38.322	4.2	125	0.00
13 C	1,4-Dichlorobenzene	40.000	37.990	5.0	125	0.00
14	1,2-Dichlorobenzene	40.000	38.639	3.4	127	0.00
15	Benzyl Alcohol	40.000	41.024	-2.6	129	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.680	-1.7	133	0.00
17	2-Methylphenol	40.000	38.279	4.3	126	0.00
18	Hexachloroethane	40.000	39.586	1.0	126	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.567	-1.4	131	0.00
20	3+4-Methylphenols	40.000	40.042	-0.1	128	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	131	0.00
22	Acetophenone	40.000	39.947	0.1	129	0.00
23 S	Nitrobenzene-d5	80.000	78.381	2.0	131	0.00
24	Nitrobenzene	40.000	39.284	1.8	132	0.00
25	Isophorone	40.000	40.011	-0.0	130	0.00
26 C	2-Nitrophenol	40.000	38.902	2.7	130	0.00
27	2,4-Dimethylphenol	40.000	39.644	0.9	132	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.390	-1.0	133	0.00
29 C	2,4-Dichlorophenol	40.000	39.023	2.4	127	0.00
30	1,2,4-Trichlorobenzene	40.000	38.488	3.8	126	0.00
31	Naphthalene	40.000	39.261	1.8	129	0.00
32	Benzoic acid	40.000	38.142	4.6	122	0.02
33	4-Chloroaniline	40.000	40.468	-1.2	130	0.00
34 C	Hexachlorobutadiene	40.000	37.782	5.5	124	0.00
35	Caprolactam	40.000	38.413	4.0	124	0.02
36 C	4-Chloro-3-methylphenol	40.000	38.414	4.0	129	0.00
37	2-Methylnaphthalene	40.000	39.945	0.1	136	0.00
38	1-Methylnaphthalene	40.000	39.617	1.0	130	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	132	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.576	6.1	125	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.644	-1.6	131	0.00
42 S	2,4,6-Tribromophenol	80.000	74.509	6.9	125	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.976	5.1	121	0.00
44	2,4,5-Trichlorophenol	40.000	39.058	2.4	126	0.00
45 S	2-Fluorobiphenyl	80.000	76.659	4.2	127	0.00
46	1,1'-Biphenyl	40.000	38.767	3.1	130	0.00
47	2-Chloronaphthalene	40.000	38.715	3.2	131	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110124\
 Data File : BG063148.D
 Acq On : 1 Nov 2024 15:54
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 01 16:31:41 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.041	2.4	124	0.00
49	Acenaphthylene	40.000	39.585	1.0	134	0.00
50	Dimethylphthalate	40.000	38.794	3.0	132	0.00
51	2,6-Dinitrotoluene	40.000	39.310	1.7	131	0.00
52 C	Acenaphthene	40.000	39.004	2.5	131	0.00
53	3-Nitroaniline	40.000	40.330	-0.8	132	0.00
54 P	2,4-Dinitrophenol	40.000	39.323	1.7	127	0.00
55	Dibenzofuran	40.000	38.250	4.4	129	0.00
56 P	4-Nitrophenol	40.000	40.166	-0.4	131	0.00
57	2,4-Dinitrotoluene	40.000	39.753	0.6	132	0.00
58	Fluorene	40.000	39.111	2.2	130	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.344	1.6	131	0.00
60	Diethylphthalate	40.000	38.213	4.5	127	0.00
61	4-Chlorophenyl-phenylether	40.000	39.278	1.8	131	0.00
62	4-Nitroaniline	40.000	38.808	3.0	131	0.00
63	Azobenzene	40.000	38.989	2.5	131	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	127	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.754	3.1	124	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.587	1.0	130	0.00
67	4-Bromophenyl-phenylether	40.000	39.919	0.2	129	0.00
68	Hexachlorobenzene	40.000	39.715	0.7	128	0.00
69	Atrazine	40.000	31.567	21.1	126	0.00
70 C	Pentachlorophenol	40.000	41.225	-3.1	124	0.00
71	Phenanthrene	40.000	38.395	4.0	126	0.00
72	Anthracene	40.000	38.452	3.9	124	0.00
73	Carbazole	40.000	38.408	4.0	129	0.00
74	Di-n-butylphthalate	40.000	37.848	5.4	127	0.00
75 C	Fluoranthene	40.000	36.956	7.6	126	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	126	0.00
77	Benzydine	40.000	34.988	12.5	130	0.00
78	Pyrene	40.000	40.201	-0.5	123	0.00
79 S	Terphenyl-d14	80.000	77.896	2.6	120	0.00
80	Butylbenzylphthalate	40.000	41.078	-2.7	128	0.00
81	Benzo(a)anthracene	40.000	38.465	3.8	121	0.00
82	3,3'-Dichlorobenzidine	40.000	40.107	-0.3	128	0.00
83	Chrysene	40.000	38.584	3.5	121	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.184	-0.5	126	0.00
85 c	Di-n-octyl phthalate	40.000	40.558	-1.4	128	0.00
86 I	Perylene-d12	20.000	20.000	0.0	127	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.122	2.2	125	0.00
88	Benzo(b)fluoranthene	40.000	39.497	1.3	127	0.00
89	Benzo(k)fluoranthene	40.000	38.850	2.9	123	0.00
90 C	Benzo(a)pyrene	40.000	39.208	2.0	127	0.00
91	Dibenzo(a,h)anthracene	40.000	38.980	2.6	125	0.00
92	Benzo(g,h,i)perylene	40.000	38.695	3.3	125	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110124\
Data File : BG063148.D
Acq On : 1 Nov 2024 15:54
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 01 16:31:41 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 01 05:13:40 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0