

Data Path : Z:\HPCHEM1\BNA G\DATA\BG110215\
 Data File : BG019443.D
 Acq On : 2 Nov 2015 21:57
 Operator : UM/NP
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02026

Quant Time: Nov 03 00:56:06 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 03 00:19:25 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.476	0.494	-3.8	136	0.00
3 S	1,4-Dioxane-d8	0.418	0.426	-1.9	125	0.00
4	Benzaldehyde	1.194	1.519	-27.2#	118	0.00
5 S	Phenol-d5	1.836	1.949	-6.2	119	0.00
6	Phenol	1.956	2.143	-9.6	119	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.164	1.319	-13.3	125	0.00
8	Bis(2-Chloroethyl)ether	1.344	1.413	-5.1	114	0.00
9 S	2-Chlorophenol-d4	1.146	1.208	-5.4	117	0.00
10	2-Chlorophenol	1.180	1.256	-6.4	117	0.00
11	2-Methylphenol	1.446	1.526	-5.5	118	0.00
12	2,2'-oxybis(1-Chloropropane	1.077	1.238	-14.9	128	0.00
13 S	4-Methylphenol-d8	1.462	1.594	-9.0	118	0.00
14	Acetophenone	2.455	2.671	-8.8	116	0.00
15 P	N-Nitroso-di-n-propylamine	1.609	1.800	-11.9	127	0.00
16	4-Methylphenol	1.580	1.748	-10.6	120	0.00
17	Hexachloroethane	0.616	0.638	-3.6	126	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
19 S	Nitrobenzene-d5	0.147	0.157	-6.8	123	0.00
20	Nitrobenzene	0.575	0.600	-4.3	119	0.00
21	Isophorone	1.041	1.150	-10.5	128	0.00
22 S	2-Nitrophenol-d4	0.174	0.192	-10.3	121	0.00
23 C	2-Nitrophenol	0.198	0.198	0.0	116	0.00
24	2,4-Dimethylphenol	0.518	0.531	-2.5	118	0.00
25	Bis(2-Chloroethoxy)methane	0.493	0.529	-7.3	122	0.00
26 S	2,4-Dichlorophenol-d3	0.379	0.387	-2.1	115	0.00
27 C	2,4-Dichlorophenol	0.362	0.376	-3.9	113	0.00
28	Naphthalene	1.001	1.003	-0.2	113	0.00
29 S	4-Chloroaniline-d4	0.383	0.454	-18.5	116	0.00
30	4-Chloroaniline	0.401	0.465	-16.0	116	0.00
31 C	Hexachlorobutadiene	0.370	0.360	2.7	109	0.00
32	Caprolactam	0.130	0.165	-26.9#	142	0.00
33 C	4-Chloro-3-methylphenol	0.486	0.531	-9.3	122	0.00
34	2-Methylnaphthalene	0.804	0.789	1.9	110	0.00
35	1-Methylnaphthalene	0.761	0.781	-2.6	113	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
37	1,2,4,5-Tetrachlorobenzene	0.824	0.787	4.5	107	0.00
38	Hexachlorocyclopentadiene	0.612	0.537	12.3	99	0.00
39 C	2,4,6-Trichlorophenol	0.487	0.520	-6.8	120	0.00
40	2,4,5-Trichlorophenol	0.535	0.566	-5.8	122	0.00
41	1,1'-Biphenyl	1.414	1.418	-0.3	111	0.00
42	2-Chloronaphthalene	1.103	1.125	-2.0	115	0.00
43	2-Nitroaniline	0.472	0.513	-8.7	121	0.00
44 S	Dimethylphthalate-d6	1.649	1.798	-9.0	123	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG110215\
 Data File : BG019443.D
 Acq On : 2 Nov 2015 21:57
 Operator : UM/NP
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02026

Quant Time: Nov 03 00:56:06 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 03 00:19:25 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.686	1.808	-7.2	123	0.00
46	2,6-Dinitrotoluene	0.341	0.367	-7.6	121	0.00
47 S	Acenaphthylene-d8	1.826	1.892	-3.6	116	0.00
48	Acenaphthylene	1.871	1.901	-1.6	115	0.00
49	3-Nitroaniline	0.294	0.342	-16.3	126	0.00
50 C	Acenaphthene	1.263	1.291	-2.2	116	0.00
51	2,4-Dinitrophenol	0.257	0.179	30.4#	83	0.00
52 S	4-Nitrophenol-d4	0.262	0.305	-16.4	130	0.00
53	4-Nitrophenol	0.409	0.440	-7.6	120	0.00
54	Dibenzofuran	1.861	1.892	-1.7	114	0.00
55	2,4-Dinitrotoluene	0.539	0.578	-7.2	123	0.00
56	2,3,4,6-Tetrachlorophenol	0.559	0.600	-7.3	120	0.00
57	Diethylphthalate	1.665	1.822	-9.4	124	0.00
58 S	Fluorene-d10	1.525	1.556	-2.0	113	0.00
59	Fluorene	1.623	1.668	-2.8	116	0.00
60	4-Chlorophenyl-phenylether	0.965	0.973	-0.8	114	0.00
61	4-Nitroaniline	0.338	0.385	-13.9	126	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.129	0.098	24.0#	101	0.00
64	4,6-Dinitro-2-methylphenol	0.140	0.100	28.6#	95	0.00
65	N-Nitrosodiphenylamine	0.517	0.465	10.1	118	0.00
66	4-Bromophenyl-phenylether	0.238	0.214	10.1	117	0.00
67	Hexachlorobenzene	0.252	0.223	11.5	115	0.00
68	Atrazine	0.256	0.254	0.8	128	0.00
69 C	Pentachlorophenol	0.149	0.122	18.1	113	0.00
70	Phenanthrene	1.009	0.913	9.5	119	0.00
71 S	Anthracene-d10	0.912	0.834	8.6	121	0.00
72	Anthracene	1.029	0.945	8.2	121	0.00
73	1,2,3,4-Tetrachlorobenzene	0.290	0.249	14.1	110	0.00
74	Pentachlorobenzene	0.292	0.246	15.8	109	0.00
75	Carbazole	0.821	0.773	5.8	119	0.00
76	Di-n-butylphthalate	1.018	0.941	7.6	123	0.00
77 C	Fluoranthene	1.307	1.282	1.9	122	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	126	0.00
79 S	Pyrene-d10	0.875	0.854	2.4	123	0.00
80	Pyrene	1.122	1.077	4.0	121	0.00
81	Butylbenzylphthalate	0.365	0.364	0.3	127	0.00
82	3,3'-Dichlorobenzidine	0.392	0.387	1.3	117	0.00
83	Benzo(a)anthracene	1.165	1.154	0.9	126	0.00
84	Bis(2-ethylhexyl)phthalate	0.518	0.505	2.5	125	0.00
85	Chrysene	1.093	1.063	2.7	124	0.00
86 I	Perylene-d12	1.000	1.000	0.0	125	0.00
87	Di-n-octyl phthalate	0.906	0.930	-2.6	125	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG110215\
Data File : BG019443.D
Acq On : 2 Nov 2015 21:57
Operator : UM/NP
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02026

Quant Time: Nov 03 00:56:06 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 03 00:19:25 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.180	1.171	0.8	126	0.00
89	Benzo(k)fluoranthene	1.135	1.154	-1.7	129	0.00
90 S	Benzo(a)pyrene-d12	1.004	1.006	-0.2	128	0.00
91 C	Benzo(a)pyrene	1.133	1.133	0.0	127	0.00
92	Indeno(1,2,3-cd)pyrene	1.277	1.281	-0.3	128	0.00
93	Dibenzo(a,h)anthracene	1.066	1.058	0.8	128	0.00
94	Benzo(a,h,i)perylene	0.969	1.070	-10.4	130	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0