


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110220\
 Data File : BG047191.D
 Acq On : 2 Nov 2020 9:54
 Operator : CG/JU
 Sample : DFTPP48
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP48

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG102220MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Nov 02 10:55:23 2020



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.5	40664	PASS
68	69	0.00	2	0.7	276	PASS
69	198	0.00	100	35.7	40970	PASS
70	69	0.00	2	0.6	254	PASS
127	198	10	80	41.5	47512	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	114605	PASS
199	198	5	9	6.8	7804	PASS
275	198	10	60	28.0	32088	PASS
365	198	1	100	3.7	4205	PASS
441	198	0.01	100	11.8	13516	PASS
442	198	50	100	73.7	84440	PASS
443	442	15	24	18.8	15897	PASS

Scan 2465 (17.824 min): BG047191.D
DFTPP48

m/z	Abundance
40.00	244.0
43.10	220.0
44.00	994.0
81.00	173.0
207.00	635.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

m/z	Abundance
38.00	676.0
39.10	3137.0
40.00	281.0
42.10	151.0
44.00	733.0
49.10	333.0
50.10	10904.0
51.10	39080.0
52.00	1686.0
56.00	1094.0
57.10	2557.0
61.10	670.0
62.10	509.0
63.10	1563.0
64.10	213.0
65.10	828.0
66.00	228.0
67.10	401.0
69.00	37800.0
70.00	189.0
72.90	522.0
74.10	3751.0
75.10	6005.0
77.10	41256.0
78.10	2647.0
79.00	2792.0
80.00	2478.0
81.10	3398.0
82.00	1133.0
83.10	414.0
84.90	441.0
86.10	498.0
87.10	721.0
91.10	560.0
92.10	567.0
93.10	4922.0
94.00	555.0
95.10	170.0
96.20	358.0
98.00	4008.0
99.10	3296.0
99.80	185.0
101.00	1670.0
101.80	227.0
103.00	480.0
104.10	1662.0
105.00	1081.0
106.10	433.0
107.00	12681.0
108.00	1810.0
110.00	23624.0
111.00	3246.0
112.00	605.0
116.10	1228.0
117.00	10451.0
118.00	714.0
118.90	184.0
119.90	250.0
122.00	1035.0
123.10	1840.0
124.00	588.0
125.10	344.0
125.80	283.0
127.00	41360.0
128.00	3284.0
129.00	18264.0
130.00	2008.0
131.20	355.0
132.10	331.0
133.00	195.0
134.10	621.0
135.00	1616.0
136.00	938.0
137.10	971.0
140.20	385.0
140.90	2426.0
142.00	1263.0
143.10	699.0
143.90	314.0
146.00	587.0
147.00	1564.0
148.00	2366.0
149.00	437.0
149.80	158.0
150.90	175.0
151.10	167.0
151.90	193.0
153.10	744.0
154.00	654.0
155.00	1527.0
156.10	1813.0
157.00	315.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

157.90	529.0
159.00	555.0
160.00	803.0
161.10	1029.0
162.00	402.0
163.20	162.0
164.90	761.0
166.00	794.0
167.00	4165.0
168.10	1587.0
169.10	457.0
169.80	323.0
170.90	222.0
171.90	462.0
173.00	607.0
174.10	1195.0
175.10	2173.0
176.00	589.0
177.00	1242.0
178.00	292.0
179.00	4025.0
180.00	2245.0
181.00	1516.0
182.00	178.0
183.90	211.0
185.10	2301.0
186.00	12944.0
187.00	4253.0
187.90	376.0
189.00	1025.0
191.00	491.0
192.00	1294.0
193.00	1554.0
193.80	347.0
194.20	256.0
196.00	2359.0
198.00	92712.0
199.00	6367.0
200.00	723.0
201.10	172.0
201.60	373.0
202.10	216.0
202.90	663.0
204.00	3937.0
205.00	6479.0
206.00	27896.0
207.10	4000.0
208.00	1049.0
209.00	514.0
209.80	324.0
210.30	508.0
211.00	1153.0
211.60	331.0
211.90	317.0
215.00	552.0
216.10	607.0
217.00	7328.0
218.00	995.0
220.00	199.0
221.00	3710.0
222.00	726.0
223.00	1670.0
224.10	14191.0
225.00	4033.0
226.00	498.0
227.00	7151.0
227.90	769.0
229.00	1290.0
229.90	217.0
231.00	513.0
234.00	526.0
235.00	587.0
236.00	356.0
237.00	670.0
238.10	191.0
238.80	206.0
240.20	349.0
241.10	497.0
242.10	943.0
243.10	794.0
244.10	9873.0
245.00	2198.0
246.00	2352.0
247.10	612.0
249.00	451.0
251.00	337.0
252.00	169.0
255.00	60136.0
256.10	8342.0
257.00	720.0
258.10	3549.0
258.90	454.0
260.00	290.0
265.00	1489.0
266.00	234.0
271.00	249.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

273.00	2208.0
274.00	5117.0
275.00	23920.0
276.00	2902.0
277.10	2702.0
278.00	433.0
283.00	336.0
284.90	430.0
291.90	191.0
292.90	486.0
296.00	7024.0
297.10	1059.0
302.10	197.0
303.00	1239.0
312.80	226.0
314.10	314.0
315.10	826.0
315.90	345.0
323.10	2483.0
324.10	481.0
326.10	252.0
326.90	293.0
328.00	274.0
334.00	1127.0
335.00	274.0
341.00	288.0
345.90	311.0
351.90	645.0
352.90	351.0
354.00	588.0
365.00	3022.0
366.00	375.0
372.00	1238.0
382.90	398.0
391.10	167.0
402.00	330.0
403.10	845.0
404.10	217.0
421.00	583.0
422.00	379.0
423.10	2375.0
424.00	381.0
441.10	8441.0
442.10	52824.0
443.10	10197.0
444.10	860.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

m/z	Abundance
38.20	628.0
39.10	3209.0
40.00	454.0
41.30	241.0
44.10	870.0
49.10	366.0
50.10	11045.0
51.10	47856.0
52.10	2477.0
55.00	245.0
56.00	1492.0
57.10	2894.0
61.00	601.0
62.10	1042.0
63.10	2792.0
64.00	464.0
65.10	956.0
67.40	191.0
69.00	49824.0
70.00	262.0
73.10	540.0
74.10	6062.0
75.10	8274.0
76.10	2168.0
77.10	55928.0
78.10	3523.0
79.00	4316.0
80.00	3237.0
81.00	4219.0
82.00	1237.0
83.00	940.0
84.10	228.0
85.10	710.0
86.00	1005.0
87.20	406.0
87.90	512.0
88.90	354.0
91.00	841.0
92.00	1254.0
93.00	6787.0
94.10	584.0
95.10	155.0
96.00	418.0
98.00	5446.0
99.10	4093.0
100.10	438.0
101.00	2915.0
102.10	226.0
103.00	951.0
104.10	1866.0
105.00	1463.0
106.00	431.0
107.00	17536.0
108.10	3376.0
110.00	30416.0
111.10	5475.0
112.10	716.0
117.10	15438.0
117.90	731.0
120.10	199.0
122.00	1371.0
123.10	1974.0
124.10	860.0
125.10	732.0
127.10	56160.0
128.10	5212.0
129.00	24592.0
130.00	2465.0
131.10	285.0
132.30	293.0
134.10	991.0
135.00	1532.0
136.00	527.0
137.10	1102.0
137.90	337.0
139.10	160.0
139.90	279.0
141.00	3717.0
142.00	1052.0
142.90	657.0
144.90	255.0
145.90	740.0
147.10	2120.0
148.00	3287.0
149.00	861.0
150.10	159.0
151.20	242.0
152.00	264.0
153.00	1156.0
154.00	861.0
155.10	2169.0
156.10	2696.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

157.00	640.0
157.90	592.0
159.10	377.0
160.10	1206.0
161.00	1556.0
162.10	233.0
163.00	264.0
164.10	222.0
165.00	1242.0
166.00	873.0
167.10	5860.0
168.10	2806.0
168.90	610.0
169.90	402.0
171.00	572.0
171.90	476.0
173.00	935.0
174.00	1685.0
175.10	2180.0
176.00	677.0
177.00	1236.0
177.70	388.0
178.10	600.0
179.00	5238.0
180.00	3822.0
180.90	1697.0
183.10	179.0
184.00	500.0
185.10	3105.0
186.00	20064.0
187.00	5307.0
188.10	722.0
189.00	1754.0
190.30	389.0
191.10	967.0
192.00	1760.0
193.00	1800.0
193.90	727.0
194.90	441.0
196.10	2975.0
198.00	140608.0
199.00	9663.0
200.10	577.0
201.30	623.0
202.10	480.0
203.00	861.0
204.10	5122.0
205.00	8725.0
206.00	40464.0
207.10	6208.0
208.00	1328.0
209.10	269.0
210.10	878.0
211.00	1699.0
211.80	496.0
214.90	312.0
216.00	1101.0
217.00	11745.0
218.00	1645.0
220.20	474.0
221.00	5495.0
221.90	2408.0
223.10	2100.0
224.10	22040.0
225.00	6911.0
226.10	560.0
227.00	10412.0
227.90	1436.0
229.00	1771.0
230.10	201.0
231.00	952.0
234.00	1001.0
235.10	824.0
236.00	489.0
237.00	618.0
238.90	506.0
240.10	445.0
241.00	642.0
242.00	1966.0
243.10	1823.0
244.10	17856.0
245.10	2217.0
246.00	3343.0
247.10	1156.0
248.00	213.0
248.80	514.0
251.80	201.0
253.10	669.0
255.00	91280.0
256.00	13527.0
256.90	873.0
258.00	5915.0
259.00	752.0
260.30	170.0
260.90	177.0
265.00	2306.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

266.00	318.0
271.00	463.0
272.00	233.0
273.00	3012.0
274.00	7044.0
275.00	38096.0
276.10	4859.0
277.00	3982.0
278.00	514.0
279.00	167.0
281.10	310.0
283.10	447.0
283.90	327.0
285.10	1017.0
286.00	298.0
288.00	192.0
290.20	182.0
291.20	202.0
292.10	213.0
292.90	884.0
294.00	246.0
294.90	223.0
296.00	12550.0
297.00	1347.0
301.00	310.0
303.10	1439.0
304.00	318.0
304.80	164.0
307.90	371.0
309.90	312.0
313.90	542.0
314.90	1198.0
316.00	395.0
317.00	219.0
321.00	238.0
321.90	313.0
323.00	4250.0
324.10	864.0
326.00	151.0
327.00	815.0
328.10	221.0
332.10	296.0
333.10	239.0
334.00	2301.0
335.10	487.0
336.00	197.0
341.10	375.0
342.00	259.0
345.10	169.0
346.00	767.0
352.10	1432.0
353.00	940.0
354.10	962.0
355.00	225.0
365.00	5466.0
366.00	798.0
371.10	619.0
372.00	1866.0
373.00	344.0
383.00	391.0
384.40	185.0
390.10	246.0
391.20	188.0
392.10	173.0
400.90	170.0
402.00	464.0
403.10	1208.0
404.00	421.0
415.10	181.0
421.00	673.0
422.00	713.0
423.10	5047.0
424.10	998.0
437.10	153.0
441.10	14797.0
442.10	93536.0
443.10	17840.0
444.10	2101.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

m/z	Abundance
38.00	329.0
39.10	2825.0
40.00	287.0
44.00	786.0
49.10	254.0
50.10	10142.0
51.10	35056.0
52.00	1090.0
54.80	258.0
56.10	1498.0
57.00	1879.0
61.10	527.0
61.90	590.0
63.10	1352.0
64.00	244.0
65.10	795.0
68.00	828.0
69.00	35288.0
70.00	312.0
73.10	326.0
74.10	3416.0
75.00	6187.0
76.00	2272.0
77.10	43336.0
78.10	2529.0
79.10	2695.0
80.00	2552.0
81.00	3870.0
82.00	769.0
83.00	621.0
85.20	368.0
86.00	902.0
87.20	558.0
89.00	229.0
91.00	774.0
92.00	1040.0
93.10	6583.0
94.10	739.0
98.00	4048.0
99.10	3407.0
100.00	524.0
101.00	1956.0
102.30	218.0
103.00	1146.0
104.00	1271.0
105.00	1143.0
107.00	13595.0
108.10	2654.0
110.00	22976.0
111.00	3601.0
112.10	417.0
116.10	846.0
117.10	11160.0
118.00	852.0
119.00	229.0
120.10	313.0
120.90	162.0
122.00	927.0
122.90	1951.0
124.00	795.0
125.00	515.0
127.00	45016.0
128.00	3474.0
129.00	19328.0
130.00	1653.0
131.00	437.0
133.80	503.0
134.90	1652.0
136.00	763.0
137.00	946.0
137.70	289.0
138.90	155.0
140.10	196.0
141.00	3074.0
141.90	1091.0
143.00	520.0
144.90	218.0
146.00	728.0
147.00	1286.0
148.00	2383.0
149.10	669.0
150.10	200.0
151.00	395.0
153.00	985.0
153.90	1381.0
155.00	1539.0
156.10	2389.0
156.90	362.0
158.10	342.0
159.00	208.0
159.90	827.0
161.10	1362.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

162.00	165.0
165.00	1397.0
165.90	705.0
167.00	5287.0
168.10	2325.0
169.10	670.0
170.00	356.0
170.90	341.0
172.10	511.0
173.10	638.0
174.10	1492.0
175.00	1581.0
176.00	628.0
177.00	1500.0
177.80	603.0
179.00	4469.0
180.10	3177.0
181.00	1454.0
181.80	152.0
184.00	320.0
185.10	2547.0
186.10	15033.0
187.10	5001.0
188.00	516.0
189.00	1254.0
189.90	239.0
190.90	405.0
192.00	2183.0
193.10	1879.0
193.90	727.0
194.90	248.0
196.00	3863.0
198.00	110496.0
199.00	7383.0
200.00	366.0
200.80	329.0
201.70	571.0
203.00	1123.0
204.10	4542.0
205.00	8953.0
206.00	32768.0
207.00	4079.0
208.10	1237.0
208.90	195.0
210.20	646.0
210.70	1048.0
211.10	997.0
215.10	676.0
215.90	715.0
217.00	9927.0
218.00	1108.0
219.70	372.0
221.10	4985.0
222.10	1105.0
223.00	2300.0
224.10	20648.0
225.00	4704.0
226.00	609.0
227.00	8056.0
227.90	907.0
229.00	1939.0
231.10	564.0
231.90	177.0
234.00	593.0
235.00	423.0
236.00	601.0
237.00	902.0
238.10	175.0
239.20	170.0
240.00	393.0
241.00	733.0
242.00	1117.0
244.00	14005.0
245.00	2113.0
246.00	3456.0
247.00	914.0
248.10	260.0
248.90	547.0
250.10	173.0
251.90	260.0
252.60	288.0
253.10	550.0
255.00	82176.0
256.00	11142.0
257.00	1189.0
258.00	4628.0
259.00	913.0
259.90	261.0
263.90	411.0
265.00	1978.0
266.10	511.0
272.00	169.0
273.00	2558.0
274.00	6461.0
275.00	34248.0
276.00	4175.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

277.00	3111.0
278.00	354.0
282.10	161.0
283.00	329.0
283.90	413.0
285.10	896.0
293.00	895.0
293.90	305.0
295.10	215.0
296.00	12320.0
297.10	1532.0
298.10	155.0
301.20	323.0
302.00	438.0
303.00	1732.0
304.10	325.0
309.00	156.0
311.10	238.0
312.80	196.0
313.10	170.0
314.00	503.0
315.00	1245.0
316.00	585.0
321.00	292.0
322.10	558.0
323.00	3850.0
324.10	521.0
327.00	512.0
328.00	395.0
331.80	224.0
333.10	363.0
334.10	1785.0
335.00	643.0
335.90	150.0
341.10	397.0
346.10	571.0
352.10	1193.0
353.10	860.0
354.10	765.0
365.00	4127.0
366.00	602.0
371.10	236.0
372.00	2066.0
373.00	328.0
383.10	569.0
384.00	264.0
389.90	338.0
402.00	613.0
403.20	899.0
404.00	505.0
421.00	581.0
422.00	987.0
423.10	5966.0
424.10	1271.0
425.10	238.0
441.10	17312.0
442.10	106960.0
443.10	19656.0
444.00	1632.0
445.00	211.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48