

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110221\
 Data File : BG050805.D
 Acq On : 1 Nov 2021 20:49
 Operator : CG/JU
 Sample : SSTD04057
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD040408

Quant Time: Nov 02 00:52:33 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110221.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 02 00:48:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.236	152	29397	20.000	ng/u1	0.00
20) Naphthalene-d8	11.062	136	136547	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.857	164	91428	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.601	188	211073	20.000	ng/u1	0.00
79) Chrysene-d12	21.902	240	175520	20.000	ng/u1	0.00
88) Perylene-d12	25.298	264	174963	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.588	96	14129	18.937	ng/uL	0.00
4) Pyridine-d5	4.017	84	105005	44.443	ng/u1	0.00
7) Phenol-d5	7.378	99	126194	46.195	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.548	67	81701	46.486	ng/u1	0.00
11) 2-Chlorophenol-d4	7.760	132	91080	47.466	ng/u1	0.00
15) 4-Methylphenol-d8	8.935	113	98464	46.683	ng/u1	0.00
21) Nitrobenzene-d5	9.405	128	48758	46.782	ng/u1	0.00
24) 2-Nitrophenol-d4	10.134	143	54794	46.620	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.674	165	90119	43.661	ng/u1	0.00
31) 4-Chloroaniline-d4	11.191	131	131980	43.568	ng/u1	0.00
46) Dimethylphthalate-d6	14.252	166	291607	45.162	ng/u1	0.00
49) Acenaphthylene-d8	14.558	160	359840	44.874	ng/u1	0.00
54) 4-Nitrophenol-d4	15.045	143	53448	45.738	ng/u1	0.00
60) Fluorene-d10	15.845	176	254541	45.872	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.962	200	54210	43.888	ng/u1	0.00
73) Anthracene-d10	17.701	188	402622	44.489	ng/u1	0.00
81) Pyrene-d10	19.975	212	464789	46.468	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.069	264	378969	42.098	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.624	88	15873	17.150	ng/uL	98
5) Pyridine	4.035	79	113312	46.122	ng/u1	98
6) Benzaldehyde	7.366	77	96104	70.507	ng/u1	98
8) Phenol	7.407	94	129745	45.529	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.642	93	98768	45.940	ng/u1	96
12) 2-Chlorophenol	7.795	128	90590	46.970	ng/u1	99
13) 2-Methylphenol	8.671	108	96756	45.858	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.753	45	155015	45.670	ng/u1	98
16) Acetophenone	9.058	105	152112	46.048	ng/u1	96
17) N-Nitroso-di-n-propyla...	9.041	70	92932	46.514	ng/u1	96
18) 4-Methylphenol	9.000	108	101821	45.757	ng/u1	92
19) Hexachloroethane	9.323	117	37872	45.504	ng/u1	88
22) Nitrobenzene	9.452	77	132340	43.690	ng/u1	98
23) Isophorone	9.969	82	253551	44.032	ng/u1	100
25) 2-Nitrophenol	10.163	139	54818	44.759	ng/u1	95
26) 2,4-Dimethylphenol	10.210	107	117999	44.071	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.445	93	135733	43.474	ng/u1	99
29) 2,4-Dichlorophenol	10.704	162	88981	44.620	ng/u1	99
30) Naphthalene	11.115	128	302304	43.400	ng/u1	99
32) 4-Chloroaniline	11.215	127	132243	43.467	ng/u1	99
33) Hexachlorobutadiene	11.379	225	59033	43.680	ng/u1	96
34) Caprolactam	11.978	113	19922	24.068	ng/u1	88
35) 4-Chloro-3-methylphenol	12.319	107	113163	45.579	ng/u1	99
36) 2-Methylnaphthalene	12.701	142	205387	44.087	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110221\
 Data File : BG050805.D
 Acq On : 1 Nov 2021 20:49
 Operator : CG/JU
 Sample : SST04057
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST040408

Quant Time: Nov 02 00:52:33 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110221.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 02 00:48:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.919	142	208471	44.305	ng/ul	94
39) 1,2,4,5-Tetrachloroben...	13.060	216	112411	44.258	ng/ul	97
40) Hexachlorocyclopentadiene	13.030	237	55781	40.860	ng/ul	96
41) 2,4,6-Trichlorophenol	13.295	196	76273	45.622	ng/ul	95
42) 2,4,5-Trichlorophenol	13.371	196	78587	44.540	ng/ul	94
43) 1,1'-Biphenyl	13.694	154	274468	43.328	ng/ul	99
44) 2-Chloronaphthalene	13.741	162	215096	43.846	ng/ul	97
45) 2-Nitroaniline	13.941	65	86640	44.282	ng/ul	99
47) Dimethylphthalate	14.299	163	287509	44.342	ng/ul	99
48) 2,6-Dinitrotoluene	14.429	165	61851	46.293	ng/ul	95
50) Acenaphthylene	14.587	152	355265	43.335	ng/ul	99
51) 3-Nitroaniline	14.763	138	67967	52.644	ng/ul#	93
52) Acenaphthene	14.922	153	234150	43.719	ng/ul	98
53) 2,4-Dinitrophenol	14.969	184	34250	45.961	ng/ul	93
55) 4-Nitrophenol	15.063	109	49418	43.977	ng/ul	99
56) Dibenzofuran	15.257	168	334392	44.158	ng/ul	99
57) 2,4-Dinitrotoluene	15.216	165	89419	46.645	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.474	232	64542	46.929	ng/ul#	93
59) Diethylphthalate	15.651	149	310377	44.810	ng/ul	99
61) Fluorene	15.903	166	263310	43.806	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.886	204	141226	44.549	ng/ul	98
63) 4-Nitroaniline	15.921	138	68853	56.971	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.980	198	51886	43.485	ng/ul#	93
67) N-Nitrosodiphenylamine	16.097	169	240161	42.821	ng/ul	98
68) 4-Bromophenyl-phenylether	16.779	248	87703	44.102	ng/ul	97
69) Hexachlorobenzene	16.902	284	88378	43.514	ng/ul	97
70) Atrazine	17.037	200	104579	43.556	ng/ul	99
71) Pentachlorophenol	17.249	266	42039	40.617	ng/ul	95
72) Phenanthrene	17.642	178	453880	43.016	ng/ul	99
74) Anthracene	17.736	178	451375	42.884	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.665	216	118520	41.996	ng/ul	99
76) Pentachlorobenzene	15.175	250	109420	41.980	ng/ul	98
77) Carbazole	18.007	167	428943	43.889	ng/ul	99
78) Di-n-butylphthalate	18.535	149	542764	43.972	ng/ul	100
80) Fluoranthene	19.646	202	550405	43.175	ng/ul	99
82) Pyrene	20.004	202	536453	43.034	ng/ul	99
83) Butylbenzylphthalate	20.868	149	228548	43.248	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.785	252	182981	47.000	ng/ul	97
85) Benzo(a)anthracene	21.879	228	484744	43.546	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.749	149	327382	43.532	ng/ul	100
87) Chrysene	21.949	228	467531	44.111	ng/ul	98
89) Di-n-octyl phthalate	23.024	149	555395	43.601	ng/ul	100
90) Benzo(b)fluoranthene	24.217	252	486670	43.579	ng/ul	99
91) Benzo(k)fluoranthene	24.288	252	454045	43.625	ng/ul	99
93) Benzo(a)pyrene	25.145	252	467178	43.645	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	29.223	276	519171	43.891	ng/ul	99
95) Dibenzo(a,h)anthracene	29.282	278	441337	43.381	ng/ul	99
96) Benzo(g,h,i)perylene	30.457	276	435150	44.036	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 17.150 ng/uL
RT: 3.624 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion: 88 Resp: 15873
Ion Ratio Lower Upper
88 100
43 30.5 27.2 40.8
58 72.2 58.2 87.2



#5
Pyridine
Concen: 46.122 ng/uL
RT: 4.035 min Scan# 93
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion: 79 Resp: 113312
Ion Ratio Lower Upper
79 100
52 78.9 65.1 97.7
51 36.2 29.6 44.4









#6
Benzaldehyde
Concen: 70.507 ng/ul
RT: 7.366 min Scan# 660
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion: 77 Resp: 96104
Ion Ratio Lower Upper
77 100
105 86.0 69.8 104.6
106 82.0 67.1 100.7



#8
Phenol
Concen: 45.529 ng/ul
RT: 7.407 min Scan# 667
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion: 94 Resp: 129745
Ion Ratio Lower Upper
94 100
65 30.9 26.2 39.4
66 39.4 32.2 48.4





#10

Bis(2-Chloroethyl)ether

Concen: 45.940 ng/ul

RT: 7.642 min Scan# 707

Delta R.T. -0.001 min

Lab File: BG050805.D

Acq: 1 Nov 2021 20:49

Instrument :

BNA_G

ClientSampleId :

SSTD040408





Tgt Ion: 93 Resp: 98768

Ion Ratio Lower Upper

93 100

63 89.2 68.2 102.4

95 33.1 27.7 41.5



#12

2-Chlorophenol

Concen: 46.970 ng/ul

RT: 7.795 min Scan# 733

Delta R.T. -0.001 min

Lab File: BG050805.D

Acq: 1 Nov 2021 20:49




Tgt Ion:128 Resp: 90590


Ion Ratio Lower Upper

128 100

64 57.7 47.2 70.8

130 32.3 25.7 38.5





#13

2-Methylphenol

Concen: 45.858 ng/ul

RT: 8.671 min Scan# 881

Delta R.T. 0.005 min

Lab File: BG050805.D

Acq: 1 Nov 2021 20:49

Instrument :

BNA_G

ClientSampleId :





SSTD040408

Tgt Ion:108 Resp: 96756

Ion Ratio Lower Upper

108 100

107 88.9 72.2 108.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 45.670 ng/ul

RT: 8.753 min Scan# 896

Delta R.T. -0.007 min

Lab File: BG050805.D

Acq: 1 Nov 2021 20:49




Tgt Ion: 45 Resp: 155015


Ion Ratio Lower Upper

45 100

77 14.4 12.1 18.1

79 10.0 8.5 12.7









#16
Acetophenone
Concen: 46.048 ng/ul
RT: 9.058 min Scan# 948
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:105 Resp: 152112
Ion Ratio Lower Upper
105 100
77 88.6 67.4 101.0
51 29.3 22.8 34.2



#17
N-Nitroso-di-n-propylamine
Concen: 46.514 ng/ul
RT: 9.041 min Scan# 945
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion: 70 Resp: 92932
Ion Ratio Lower Upper
70 100
42 55.6 47.6 71.4
101 8.3 6.1 9.1
130 16.3 12.2 18.2









#18
4-Methylphenol
Concen: 45.757 ng/ul
RT: 9.000 min Scan# 911
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:108 Resp: 101821
Ion Ratio Lower Upper
108 100
107 113.1 97.8 146.6



#19
Hexachloroethane
Concen: 45.504 ng/ul
RT: 9.323 min Scan# 993
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:117 Resp: 37872
Ion Ratio Lower Upper
117 100
201 92.3 86.2 129.2
199 58.7 51.8 77.8









#22
Nitrobenzene
Concen: 43.690 ng/ul
RT: 9.452 min Scan# 1014
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion: 77 Resp: 132340
Ion Ratio Lower Upper
77 100
123 37.5 31.0 46.6
65 14.0 10.8 16.2



#23
Isophorone
Concen: 44.032 ng/ul
RT: 9.969 min Scan# 1103
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion: 82 Resp: 253551
Ion Ratio Lower Upper
82 100
95 6.9 5.5 8.3
138 15.3 12.4 18.6









#25
2-Nitrophenol
Concen: 44.759 ng/ul
RT: 10.163 min Scan# 1136
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:139 Resp: 54818
Ion Ratio Lower Upper
139 100
65 63.5 46.7 70.1
109 34.4 29.0 43.6



#26
2,4-Dimethylphenol
Concen: 44.071 ng/ul
RT: 10.210 min Scan# 1144
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:107 Resp: 117999
Ion Ratio Lower Upper
107 100
121 48.7 39.1 58.7
122 78.8 64.3 96.5





#27

Bis(2-Chloroethoxy)methane

Concen: 43.474 ng/ul

RT: 10.445 min Scan# 1184

Delta R.T. -0.001 min

Lab File: BG050805.D

Acq: 1 Nov 2021 20:49

Instrument :

BNA_G

ClientSampleId :

SSTD040408





Tgt Ion: 93 Resp: 135733

Ion Ratio Lower Upper

93 100

95 33.8 26.8 40.2

123 10.1 7.6 11.4



#29

2,4-Dichlorophenol

Concen: 44.620 ng/ul

RT: 10.704 min Scan# 1228

Delta R.T. 0.005 min

Lab File: BG050805.D

Acq: 1 Nov 2021 20:49




Tgt Ion:162 Resp: 88981


Ion Ratio Lower Upper

162 100

164 64.8 52.2 78.4

98 38.2 30.9 46.3









#30
Naphthalene
Concen: 43.400 ng/ul
RT: 11.115 min Scan# 1115
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:128 Resp: 302304
Ion Ratio Lower Upper
128 100
129 11.2 8.5 12.7
127 13.7 10.7 16.1



#32
4-Chloroaniline
Concen: 43.467 ng/ul
RT: 11.215 min Scan# 1315
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:127 Resp: 132243
Ion Ratio Lower Upper
127 100
129 32.9 25.8 38.8









#33
Hexachlorobutadiene
Concen: 43.680 ng/ul
RT: 11.379 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:225 Resp: 59033
Ion Ratio Lower Upper
225 100
223 65.4 53.7 80.5
227 68.5 50.9 76.3



#34
Caprolactam
Concen: 24.068 ng/ul
RT: 11.978 min Scan# 1445
Delta R.T. 0.011 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:113 Resp: 19922
Ion Ratio Lower Upper
113 100
55 200.7 143.9 215.9
56 139.6 104.0 156.0









#35
4-Chloro-3-methylphenol
Concen: 45.579 ng/ul
RT: 12.319 min Scan# 11
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:107 Resp: 113163
Ion Ratio Lower Upper
107 100
144 23.3 18.6 27.8
142 70.8 57.8 86.6



#36
2-Methylnaphthalene
Concen: 44.087 ng/ul
RT: 12.701 min Scan# 1568
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:142 Resp: 205387
Ion Ratio Lower Upper
142 100
141 90.9 71.1 106.7









#37
1-Methylnaphthalene
Concen: 44.305 ng/ul
RT: 12.919 min Scan# 1
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:142 Resp: 208471
Ion Ratio Lower Upper
142 100
141 97.1 73.0 109.4
116 4.7 3.6 5.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 44.258 ng/ul
RT: 13.060 min Scan# 1629
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:216 Resp: 112411
Ion Ratio Lower Upper
216 100
214 76.0 63.3 94.9
179 21.6 16.5 24.7
108 18.4 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 40.860 ng/ul
RT: 13.030 min Scan# 1624
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:237 Resp: 55781
Ion Ratio Lower Upper
237 100
235 60.9 46.4 69.6
272 10.0 9.6 14.4



#41
2,4,6-Trichlorophenol
Concen: 45.622 ng/ul
RT: 13.295 min Scan# 1669
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:196 Resp: 76273
Ion Ratio Lower Upper
196 100
198 92.2 78.5 117.7
200 30.6 25.5 38.3









#42
2,4,5-Trichlorophenol
Concen: 44.540 ng/ul
RT: 13.371 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:196 Resp: 78587
Ion Ratio Lower Upper
196 100
198 96.2 72.2 108.4
200 31.4 23.1 34.7



#43
1,1'-Biphenyl
Concen: 43.328 ng/ul
RT: 13.694 min Scan# 1737
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:154 Resp: 274468
Ion Ratio Lower Upper
154 100
153 42.1 34.0 51.0
76 16.1 11.9 17.9









#44
2-Chloronaphthalene
Concen: 43.846 ng/ul
RT: 13.741 min Scan# 1745
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:162 Resp: 215096
Ion Ratio Lower Upper
162 100
164 31.7 27.2 40.8
127 40.5 33.4 50.0



#45
2-Nitroaniline
Concen: 44.282 ng/ul
RT: 13.941 min Scan# 1779
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion: 65 Resp: 86640
Ion Ratio Lower Upper
65 100
92 64.0 51.8 77.6
138 88.7 70.6 105.8









#47
Dimethylphthalate
Concen: 44.342 ng/ul
RT: 14.299 min Scan# 1839
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:163 Resp: 287509
Ion Ratio Lower Upper
163 100
194 3.5 3.3 4.9
164 10.4 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 46.293 ng/ul
RT: 14.429 min Scan# 1862
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:165 Resp: 61851
Ion Ratio Lower Upper
165 100
89 71.3 53.0 79.6
121 23.1 18.1 27.1









#50
Acenaphthylene
Concen: 43.335 ng/ul
RT: 14.587 min Scan# 1888
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:152 Resp: 355265
Ion Ratio Lower Upper
152 100
151 20.9 16.6 25.0
153 13.2 11.4 17.2



#51
3-Nitroaniline
Concen: 52.644 ng/ul
RT: 14.763 min Scan# 1919
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:138 Resp: 67967
Ion Ratio Lower Upper
138 100
108 10.2 10.4 15.6#
92 124.4 106.2 159.2









#52
Acenaphthene
Concen: 43.719 ng/ul
RT: 14.922 min Scan# 1946
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:153 Resp: 234150
Ion Ratio Lower Upper
153 100
152 47.2 37.8 56.6
154 82.2 68.6 102.8



#53
2,4-Dinitrophenol
Concen: 45.961 ng/ul
RT: 14.969 min Scan# 1954
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:184 Resp: 34250
Ion Ratio Lower Upper
184 100
63 75.6 54.5 81.7
154 65.7 49.6 74.4









#55
4-Nitrophenol
Concen: 43.977 ng/ul
RT: 15.063 min Scan# 1969
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:109 Resp: 49418
Ion Ratio Lower Upper
109 100
139 109.6 86.8 130.2
65 135.2 108.7 163.1



#56
Dibenzofuran
Concen: 44.158 ng/ul
RT: 15.257 min Scan# 2003
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:168 Resp: 334392
Ion Ratio Lower Upper
168 100
139 40.6 32.0 48.0









#57
2,4-Dinitrotoluene
Concen: 46.645 ng/ul
RT: 15.216 min Scan# 1995
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:165 Resp: 89419
Ion Ratio Lower Upper
165 100
63 49.4 43.0 64.6
182 2.5 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 46.929 ng/ul
RT: 15.474 min Scan# 2040
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49


Tgt Ion:232 Resp: 64542
Ion Ratio Lower Upper
232 100
131 52.1 37.4 56.2
130 0.0 2.3 3.5#
166 32.9 24.4 36.6








#59
Diethylphthalate
Concen: 44.810 ng/ul
RT: 15.651 min Scan# 2070
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49


Instrument :
BNA_G
ClientSampleId :
SSTD040408





Tgt Ion:149 Resp: 310377
Ion Ratio Lower Upper
149 100
177 21.5 16.3 24.5
150 12.8 10.3 15.5




#61
Fluorene
Concen: 43.806 ng/ul
RT: 15.903 min Scan# 2113
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49



Tgt Ion:166 Resp: 263310
Ion Ratio Lower Upper
166 100
165 100.7 78.3 117.5
167 14.2 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 44.549 ng/ul
RT: 15.886 min Scan# 2110
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:204 Resp: 141226
Ion Ratio Lower Upper
204 100
206 31.3 26.6 40.0
141 56.9 45.8 68.8



#63
4-Nitroaniline
Concen: 56.971 ng/ul
RT: 15.921 min Scan# 2116
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:138 Resp: 68853
Ion Ratio Lower Upper
138 100
92 58.6 43.4 65.0
108 96.8 75.7 113.5









#66
 4,6-Dinitro-2-methylphenol
 Concen: 43.485 ng/ul
 RT: 15.980 min Scan# 2126
 Delta R.T. 0.005 min
 Lab File: BG050805.D
 Acq: 1 Nov 2021 20:49




Instrument :
 BNA_G
 ClientSampleId :
 SST040408


Tgt Ion:198 Resp: 51886
 Ion Ratio Lower Upper
 198 100
 182 3.0 3.7 5.5#
 77 28.0 25.6 38.4



#67
 N-Nitrosodiphenylamine
 Concen: 42.821 ng/ul
 RT: 16.097 min Scan# 2146
 Delta R.T. -0.001 min
 Lab File: BG050805.D
 Acq: 1 Nov 2021 20:49

Tgt Ion:169 Resp: 240161
 Ion Ratio Lower Upper
 169 100
 168 70.9 58.9 88.3
 167 38.4 30.5 45.7









#68
 4-Bromophenyl-phenylether
 Concen: 44.102 ng/ul
 RT: 16.779 min Scan# 21
 Delta R.T. -0.001 min
 Lab File: BG050805.D
 Acq: 1 Nov 2021 20:49




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040408


Tgt Ion:248 Resp: 87703
 Ion Ratio Lower Upper
 248 100
 250 93.3 78.6 118.0
 141 80.8 63.7 95.5



#69
 Hexachlorobenzene
 Concen: 43.514 ng/ul
 RT: 16.902 min Scan# 2283
 Delta R.T. -0.001 min
 Lab File: BG050805.D
 Acq: 1 Nov 2021 20:49

Tgt Ion:284 Resp: 88378
 Ion Ratio Lower Upper
 284 100
 142 42.8 32.3 48.5
 249 35.7 28.0 42.0









#70
 Atrazine
 Concen: 43.556 ng/ul
 RT: 17.037 min Scan# 21
 Delta R.T. -0.001 min
 Lab File: BG050805.D
 Acq: 1 Nov 2021 20:49




Instrument :
 BNA_G
 ClientSampleId :
 SST040408


Tgt Ion:200 Resp: 104579
 Ion Ratio Lower Upper
 200 100
 173 26.1 20.4 30.6
 215 45.6 37.2 55.8



#71
 Pentachlorophenol
 Concen: 40.617 ng/ul
 RT: 17.249 min Scan# 2342
 Delta R.T. -0.001 min
 Lab File: BG050805.D
 Acq: 1 Nov 2021 20:49

Tgt Ion:266 Resp: 42039
 Ion Ratio Lower Upper
 266 100
 264 63.7 56.3 84.5
 268 64.6 53.2 79.8









#72
Phenanthrene
Concen: 43.016 ng/ul
RT: 17.642 min Scan# 2409
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:178 Resp: 453880
Ion Ratio Lower Upper
178 100
179 15.9 13.0 19.6
176 19.5 15.9 23.9



#74
Anthracene
Concen: 42.884 ng/ul
RT: 17.736 min Scan# 2425
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:178 Resp: 451375
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.4
176 19.0 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 41.996 ng/uL
RT: 13.665 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:216 Resp: 118520
Ion Ratio Lower Upper
216 100
214 77.3 60.5 90.7
218 48.8 39.3 58.9



#76
Pentachlorobenzene
Concen: 41.980 ng/uL
RT: 15.175 min Scan# 1989
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49


Tgt Ion:250 Resp: 109420
Ion Ratio Lower Upper
250 100
248 63.3 51.8 77.8
252 65.1 50.3 75.5








#77
 Carbazole
 Concen: 43.889 ng/ul
 RT: 18.007 min Scan# 2470
 Delta R.T. 0.005 min
 Lab File: BG050805.D
 Acq: 1 Nov 2021 20:49

Instrument :
 BNA_G
 ClientSampleId :
 SSTD040408







Tgt Ion:167 Resp: 428943
 Ion Ratio Lower Upper
 167 100
 166 22.6 17.9 26.9
 139 13.6 10.4 15.6



#78
 Di-n-butylphthalate
 Concen: 43.972 ng/ul
 RT: 18.535 min Scan# 2561
 Delta R.T. -0.001 min
 Lab File: BG050805.D
 Acq: 1 Nov 2021 20:49

Tgt Ion:149 Resp: 542764
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.5 11.3
 104 6.1 5.0 7.6









#80
 Fluoranthene
 Concen: 43.175 ng/ul
 RT: 19.646 min Scan# 21
 Delta R.T. 0.005 min
 Lab File: BG050805.D
 Acq: 1 Nov 2021 20:49




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040408


Tgt Ion:202 Resp: 550405
 Ion Ratio Lower Upper
 202 100
 101 10.8 8.2 12.4
 100 8.7 7.2 10.8



#82
 Pyrene
 Concen: 43.034 ng/ul
 RT: 20.004 min Scan# 2811
 Delta R.T. -0.001 min
 Lab File: BG050805.D
 Acq: 1 Nov 2021 20:49

Tgt Ion:202 Resp: 536453
 Ion Ratio Lower Upper
 202 100
 101 13.3 10.6 16.0
 100 11.8 9.0 13.6









#83
Butylbenzylphthalate
Concen: 43.248 ng/ul
RT: 20.868 min Scan# 2958
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:149 Resp: 228548
Ion Ratio Lower Upper
149 100
91 83.8 63.9 95.9
206 19.7 15.4 23.2



#84
3,3'-Dichlorobenzidine
Concen: 47.000 ng/ul
RT: 21.785 min Scan# 3114
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:252 Resp: 182981
Ion Ratio Lower Upper
252 100
254 66.5 51.4 77.0
126 10.9 7.6 11.4









#85
Benzo(a)anthracene
Concen: 43.546 ng/ul
RT: 21.879 min Scan# 3130
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:228 Resp: 484744
Ion Ratio Lower Upper
228 100
229 20.8 15.2 22.8
226 28.7 22.1 33.1



#86
Bis(2-ethylhexyl)phthalate
Concen: 43.532 ng/ul
RT: 21.749 min Scan# 3108
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:149 Resp: 327382
Ion Ratio Lower Upper
149 100
167 25.8 20.8 31.2
279 3.6 2.9 4.3









#87
Chrysene
Concen: 44.111 ng/ul
RT: 21.949 min Scan# 3142
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:228 Resp: 467531
Ion Ratio Lower Upper
228 100
226 31.0 24.0 36.0
229 20.8 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 43.601 ng/ul
RT: 23.024 min Scan# 3325
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:149 Resp: 555395









#90
Benzo(b)fluoranthene
Concen: 43.579 ng/ul
RT: 24.217 min Scan# 3527
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:252 Resp: 486670
Ion Ratio Lower Upper
252 100
253 22.8 17.8 26.8
125 10.2 8.1 12.1



#91
Benzo(k)fluoranthene
Concen: 43.625 ng/ul
RT: 24.288 min Scan# 3540
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:252 Resp: 454045
Ion Ratio Lower Upper
252 100
253 21.8 17.0 25.6
125 10.0 7.3 10.9









#93
Benzo(a)pyrene
Concen: 43.645 ng/ul
RT: 25.145 min Scan# 3
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49




Instrument :
BNA_G
ClientSampleId :
SSTD040408


Tgt Ion:252 Resp: 467178
Ion Ratio Lower Upper
252 100
253 21.9 18.0 27.0
125 11.3 8.0 12.0



#94
Indeno(1,2,3-cd)pyrene
Concen: 43.891 ng/ul
RT: 29.223 min Scan# 4380
Delta R.T. 0.005 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49


Tgt Ion:276 Resp: 519171
Ion Ratio Lower Upper
276 100
138 19.6 14.9 22.3
277 24.1 19.1 28.7








#95
Dibenzo(a,h)anthracene
Concen: 43.381 ng/ul
RT: 29.282 min Scan# 41
Delta R.T. -0.001 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Instrument :
BNA_G
ClientSampleId :
SSTD040408



Tgt Ion:278 Resp: 441337
Ion Ratio Lower Upper
278 100
139 14.3 12.0 18.0
279 24.3 19.1 28.7



#96
Benzo(g,h,i)perylene
Concen: 44.036 ng/ul
RT: 30.457 min Scan# 4590
Delta R.T. 0.011 min
Lab File: BG050805.D
Acq: 1 Nov 2021 20:49

Tgt Ion:276 Resp: 435150
Ion Ratio Lower Upper
276 100
138 19.0 14.8 22.2
277 22.4 18.2 27.2

