

Data Path : Z:\HPCHEM1\BNA G\DATA\BG110315\
 Data File : BG019462.D
 Acq On : 3 Nov 2015 21:24
 Operator : UM/NP
 Sample : G4211-05
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BC7E5

Quant Time: Nov 04 01:00:57 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 04 00:13:26 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.18	152	27753	20.00	ng/ul	0.00
18) Naphthalene-d8	11.01	136	117447	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.81	164	95938	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.55	188	281952	20.00	ng/ul	0.00
78) Chrysene-d12	21.83	240	390731	20.00	ng/ul	0.00
86) Perylene-d12	25.18	264	378013	20.00	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.48	96	3092	5.34	ng/uL	0.00
5) Phenol-d5	7.33	99	75991	29.83	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.49	67	54633	33.83	ng/ul	0.00
9) 2-Chlorophenol-d4	7.71	132	49127	30.89	ng/ul	0.00
13) 4-Methylphenol-d8	8.89	113	57399	28.29	ng/ul	0.00
19) Nitrobenzene-d5	9.35	128	26542	30.75	ng/ul	0.00
22) 2-Nitrophenol-d4	10.08	143	30776	30.04	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.62	165	64807	29.15	ng/ul	0.00
29) 4-Chloroaniline-d4	11.14	131	73653	32.75	ng/ul	0.00
44) Dimethylphthalate-d6	14.21	166	273010	34.51	ng/ul	0.00
47) Acenaphthylene-d8	14.51	160	280863	32.06	ng/ul	0.00
52) 4-Nitrophenol-d4	14.99	143	39302	31.28	ng/ul	0.00
58) Fluorene-d10	15.79	176	234521	32.05	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.91	200	50412	27.62	ng/ul	0.00
71) Anthracene-d10	17.65	188	416336	32.39	ng/ul	0.00
79) Pyrene-d10	19.92	212	542284	31.71	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.95	264	623776	32.88	ng/ul	-0.01


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
45) Dimethylphthalate	14.25	163	149597	18.49	ng/ul#	98
84) Bis(2-ethylhexyl)phthalate	21.69	149	17128	1.69	ng/ul	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG110315\
Data File : BG019462.D
Acq On : 3 Nov 2015 21:24
Operator : UM/NP
Sample : G4211-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BC7E5

Quant Time: Nov 04 01:00:57 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 04 00:13:26 2015
Response via : Initial Calibration









#45
Dimethylphthalate
Concen: 18.49 ng/ul
RT: 14.25 min Scan# 1943
Delta R.T. -0.00 min
Lab File: BG019462.D
Acq: 3 Nov 2015 21:24

Instrument :
BNA_G
ClientSampleId :
BC7E5

Tot Ion:163 Resp: 149597
Ion Ratio Lower Upper
163 100
194 3.6 4.3 6.5#
164 9.6 7.5 11.3




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E



#84
Bis(2-ethylhexyl)phthalate
Concen: 1.69 ng/ul
RT: 21.69 min Scan# 3209
Delta R.T. -0.01 min
Lab File: BG019462.D
Acq: 3 Nov 2015 21:24

Tgt Ion:149 Resp: 17128
Ion Ratio Lower Upper
149 100
167 25.4 23.2 34.8
279 5.5 5.0 7.4



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E

