

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG110316\
 Data File : BG024651.D
 Acq On : 3 Nov 2016 13:51
 Operator : UM/SJ
 Sample : H5502-30
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 GW-88-7.6-20.0

Manual Integrations
 APPROVED

Sohil
 11/4/2016 10:04:13 AM

Quant Time: Nov 03 19:05:07 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 26 11:14:13 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.27	152	122495	20.00	ng	0.00
21) Naphthalene-d8	11.09	136	449951	20.00	ng	0.00
38) Acenaphthene-d10	14.88	164	331835	20.00	ng	0.00
63) Phenanthrene-d10	17.63	188	792499	20.00	ng	0.00
75) Chrysene-d12	21.92	240	1103648	20.00	ng	0.00
86) Perylene-d12	25.35	264	1136170	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.80	112	374761	50.14	ng	0.00
7) Phenol-d6	7.40	99	339902	33.15	ng	0.00
23) Nitrobenzene-d5	9.44	82	727652	73.63	ng	0.00
41) 2,4,6-Tribromophenol	16.36	330	585535	118.11	ng	0.00
44) 2-Fluorobiphenyl	13.51	172	1519120	76.09	ng	0.00
78) Terphenyl-d14	20.21	244	2402097	65.04	ng	0.00
Target Compounds						
31) Naphthalene	11.15	128	48073	2.14	ng	Qvalue # 94
37) 2-Methylnaphthalene	12.73	142	33797	2.06	ng	90
51) Acenaphthene	14.95	154	122046	6.09	ng	99
54) Dibenzofuran	15.28	168	252780	9.12	ng	# 94
57) Fluorene	15.93	166	46318	2.02	ng	89
70) Phenanthrene	17.67	178	1205934	29.85	ng	96
71) Anthracene	17.76	178	85985m	2.11	ng	
72) Carbazole	18.03	167	80314	2.16	ng	95
74) Fluoranthene	19.66	202	401606	7.86	ng	91
77) Pyrene	20.02	202	258250	4.79	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG110316\
Data File : BG024651.D
Acq On : 3 Nov 2016 13:51
Operator : UM/SJ
Sample : H5502-30
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
GW-88-7.6-20.0

Manual Integrations
APPROVED

Sohil
11/4/2016 10:04:13 AM

Quant Time: Nov 03 19:05:07 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG102516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 26 11:14:13 2016
Response via : Initial Calibration

