


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110321\
 Data File : BG050809.D
 Acq On : 2 Nov 2021 8:46
 Operator : CG/JU
 Sample : DFTPP428
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP428

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110321.M
 Title : SVOA CALIBRATION
 Last Update : Tue Nov 02 14:49:05 2021



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.9	36400	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	47.8	37104	PASS
70	69	0.00	2	0.6	222	PASS
127	198	10	80	53.7	41712	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	77653	PASS
199	198	5	9	6.9	5349	PASS
275	198	10	60	26.3	20448	PASS
365	198	1	100	3.3	2563	PASS
441	198	0.01	100	8.2	6391	PASS
442	198	50	100	53.4	41493	PASS
443	442	15	24	19.9	8267	PASS

m/z	Abundance
40.00	409.0
44.00	726.0
207.00	300.0

Instrument :
BNA_G
ClientSampleId :
DFTPP428

m/z	Abundance
37.00	176.0
38.10	479.0
39.00	3050.0
40.00	557.0
40.90	197.0
44.00	498.0
49.10	353.0
50.00	9291.0
51.00	37384.0
52.00	1903.0
54.90	225.0
56.00	1150.0
57.00	2691.0
61.00	521.0
62.00	740.0
63.00	1675.0
64.00	205.0
65.00	666.0
69.00	37080.0
69.90	256.0
73.00	449.0
74.00	3907.0
75.00	5971.0
76.00	1972.0
77.00	42360.0
78.00	2844.0
79.00	2647.0
80.00	2469.0
81.00	3089.0
82.00	680.0
83.00	678.0
84.10	152.0
85.00	446.0
86.00	951.0
86.90	337.0
88.00	157.0
91.00	694.0
91.90	789.0
93.00	4830.0
94.00	492.0
95.00	185.0
96.00	274.0
98.00	3681.0
99.00	2765.0
100.00	248.0
101.00	1684.0
103.00	637.0
104.00	1181.0
105.00	1196.0
105.90	333.0
107.00	11925.0
108.00	1867.0
110.00	21248.0
110.90	3548.0
111.90	345.0
113.00	312.0
116.10	988.0
117.00	11438.0
117.90	656.0
122.00	729.0
123.00	1242.0
123.90	650.0
124.90	558.0
127.00	39672.0
128.00	3368.0
129.00	15681.0
130.00	1282.0
131.00	337.0
131.90	176.0
133.90	519.0
135.00	1371.0
135.90	635.0
137.00	560.0
140.00	168.0
141.00	2026.0
142.00	576.0
143.00	477.0
145.90	305.0
147.00	895.0
148.00	2666.0
149.00	599.0
151.00	371.0
152.90	747.0
153.90	489.0
155.00	1068.0
155.90	1511.0
157.10	408.0
158.00	547.0
159.00	268.0
159.90	657.0
161.00	893.0
161.90	250.0

Instrument :

BNA_G

ClientSampleId :

DFTPP428

164.90	957.0
165.90	792.0
167.00	3898.0
168.00	1804.0
169.00	292.0
171.00	157.0
171.90	348.0
172.90	378.0
174.00	1052.0
175.00	1604.0
176.00	430.0
176.90	729.0
178.10	330.0
178.90	3329.0
180.00	2236.0
181.00	1019.0
183.90	249.0
185.00	1606.0
186.00	11580.0
187.00	3497.0
188.00	350.0
189.00	777.0
190.10	207.0
191.00	431.0
192.00	916.0
193.00	1039.0
194.00	285.0
194.90	171.0
196.00	2372.0
197.90	70760.0
199.00	5028.0
199.90	385.0
201.50	417.0
203.00	648.0
204.00	3046.0
205.00	5100.0
206.00	20792.0
207.00	2936.0
208.00	775.0
209.00	306.0
210.10	417.0
211.00	856.0
211.70	288.0
215.00	305.0
216.00	572.0
217.00	5434.0
218.00	650.0
221.00	2607.0
221.90	889.0
223.00	1411.0
224.00	10886.0
225.00	3091.0
226.00	435.0
227.00	5190.0
228.00	790.0
229.00	1005.0
231.00	578.0
234.00	313.0
235.00	347.0
236.00	261.0
237.00	467.0
240.90	371.0
242.00	679.0
243.00	618.0
244.00	8346.0
245.00	1031.0
246.00	1620.0
247.00	280.0
249.00	378.0
253.10	237.0
255.00	41200.0
256.00	6791.0
257.00	570.0
258.00	2780.0
259.00	354.0
259.80	155.0
265.00	1100.0
266.00	292.0
272.00	215.0
273.00	1332.0
274.00	3156.0
275.00	17024.0
276.00	2517.0
277.10	1451.0
277.90	293.0
283.10	256.0
285.00	283.0
293.00	349.0
296.00	4995.0
297.00	840.0
301.00	185.0
303.00	678.0
304.00	200.0
308.00	159.0
314.00	342.0
315.00	649.0

Instrument :
BNA_G
ClientSampleId :
DFTPP428

316.00	426.0
323.00	1505.0
324.10	256.0
327.00	440.0
334.10	925.0
335.00	326.0
341.10	171.0
346.00	188.0
351.90	301.0
353.00	308.0
354.10	452.0
365.00	2094.0
366.00	225.0
372.10	551.0
373.00	194.0
383.00	186.0
403.00	163.0
403.90	188.0
421.00	313.0
422.00	202.0
423.10	1614.0
424.10	365.0
441.10	4436.0
442.10	29888.0
443.10	6354.0
444.10	675.0

Instrument :
BNA_G
ClientSampleId :
DFTPP428

m/z	Abundance
37.00	212.0
38.00	422.0
39.00	2983.0
39.90	618.0
44.00	448.0
45.00	188.0
50.00	10761.0
51.00	41744.0
52.00	2254.0
55.00	228.0
56.00	1253.0
57.00	3284.0
58.10	189.0
60.90	617.0
62.00	604.0
63.00	1701.0
64.00	276.0
65.00	665.0
69.00	43384.0
69.90	186.0
72.90	444.0
74.00	4790.0
75.00	7178.0
76.00	2537.0
77.00	50584.0
78.00	2846.0
79.00	3211.0
80.00	2516.0
81.00	3747.0
82.00	832.0
83.00	913.0
84.90	705.0
86.00	1021.0
87.00	607.0
88.00	191.0
91.00	899.0
92.00	814.0
93.00	6017.0
94.00	325.0
96.00	295.0
98.00	4833.0
99.00	2828.0
99.90	341.0
101.00	2160.0
103.00	731.0
104.00	1275.0
105.00	1186.0
106.00	503.0
107.00	14822.0
108.00	2109.0
110.00	25056.0
110.90	3573.0
112.00	489.0
112.90	154.0
116.10	926.0
117.00	13077.0
118.00	867.0
120.00	235.0
121.90	984.0
123.00	1724.0
123.90	695.0
125.00	642.0
127.00	47864.0
128.00	3423.0
129.00	20088.0
130.00	1688.0
130.90	356.0
132.00	332.0
133.90	667.0
135.00	1524.0
136.00	634.0
137.10	806.0
137.90	168.0
138.90	161.0
140.00	352.0
141.00	2424.0
141.90	861.0
143.00	666.0
144.00	228.0
145.10	153.0
145.90	426.0
147.00	1416.0
148.00	3319.0
149.00	555.0
150.00	159.0
151.00	272.0
151.80	204.0
153.00	924.0
154.00	650.0
155.00	1579.0
156.00	2209.0
157.00	325.0

Instrument :
BNA_G
ClientSampleId :
DFTPP428

158.00	546.0
159.00	382.0
160.00	938.0
161.00	1224.0
162.00	243.0
164.90	1057.0
166.10	834.0
167.00	4634.0
168.00	2546.0
168.90	482.0
170.00	295.0
171.10	314.0
172.00	477.0
173.00	570.0
174.00	1116.0
175.00	1915.0
176.00	529.0
176.90	937.0
178.10	491.0
179.00	4037.0
180.00	2449.0
181.00	1017.0
182.00	160.0
182.80	153.0
183.20	168.0
184.00	408.0
185.00	1929.0
186.00	14899.0
187.00	3969.0
188.10	302.0
188.90	731.0
190.00	226.0
191.00	424.0
192.00	1532.0
193.00	1483.0
194.00	276.0
195.00	159.0
196.00	2799.0
198.00	89600.0
198.90	6138.0
200.00	395.0
201.40	394.0
202.90	1031.0
204.00	4002.0
205.00	6414.0
206.00	26528.0
207.00	3826.0
208.00	983.0
208.90	344.0
210.00	386.0
211.00	949.0
211.90	176.0
214.90	343.0
215.90	660.0
217.00	7643.0
218.00	969.0
221.00	3497.0
221.80	1343.0
223.00	1466.0
224.00	14368.0
225.00	4027.0
226.00	297.0
227.00	6896.0
228.00	916.0
229.00	1523.0
230.10	152.0
231.00	483.0
233.90	446.0
235.00	533.0
236.10	260.0
237.00	729.0
239.00	160.0
241.10	354.0
242.00	709.0
243.00	959.0
244.00	10605.0
245.00	1392.0
246.00	2457.0
246.90	437.0
248.90	482.0
251.00	155.0
251.90	152.0
253.00	397.0
255.00	56904.0
256.00	8516.0
257.00	659.0
258.00	3586.0
259.00	628.0
264.10	175.0
265.00	1789.0
266.00	201.0
271.00	158.0
272.00	282.0
273.00	1807.0
274.00	3876.0
275.00	24056.0

Instrument :
BNA_G
ClientSampleId :
DFTPP428

276.00	3452.0
277.00	2350.0
277.90	424.0
283.10	263.0
285.10	336.0
293.00	640.0
294.90	157.0
296.00	7493.0
297.10	920.0
302.10	166.0
303.00	790.0
304.10	177.0
314.00	266.0
315.00	841.0
316.10	455.0
321.10	169.0
323.00	2086.0
324.00	459.0
327.00	350.0
328.00	212.0
332.00	347.0
333.00	275.0
334.10	1374.0
335.10	391.0
341.10	289.0
346.00	432.0
352.00	629.0
353.00	339.0
354.00	573.0
355.00	174.0
365.00	2729.0
366.10	599.0
372.10	950.0
373.10	320.0
383.10	167.0
402.10	432.0
403.10	575.0
404.10	283.0
421.00	458.0
422.10	411.0
423.10	3080.0
424.10	662.0
441.10	6667.0
442.10	47784.0
443.10	9470.0
444.10	822.0

Instrument :
BNA_G
ClientSampleId :
DFTPP428

m/z	Abundance
37.90	299.0
39.00	2505.0
39.90	608.0
40.90	198.0
44.00	499.0
49.00	181.0
50.00	7813.0
51.00	30072.0
52.00	1755.0
55.10	181.0
56.00	1037.0
57.00	1929.0
61.00	345.0
62.00	476.0
63.00	1047.0
65.00	596.0
69.00	30848.0
70.00	224.0
72.90	339.0
74.00	3580.0
75.00	4871.0
76.10	1790.0
77.00	37800.0
78.00	2241.0
79.00	2132.0
80.00	1820.0
80.90	2864.0
82.00	789.0
83.00	622.0
84.90	411.0
86.00	743.0
86.90	390.0
91.00	660.0
92.00	576.0
93.00	4858.0
93.90	266.0
94.90	181.0
95.90	211.0
97.00	155.0
98.00	3016.0
99.00	2245.0
100.10	199.0
100.90	1482.0
102.90	566.0
104.00	1150.0
105.00	1094.0
106.00	292.0
107.00	11496.0
108.00	1850.0
110.00	18584.0
111.00	3252.0
111.90	479.0
115.10	156.0
116.00	658.0
117.00	8833.0
118.00	806.0
118.90	229.0
119.90	183.0
122.00	753.0
123.00	1116.0
123.90	529.0
124.90	459.0
127.00	37600.0
128.00	2565.0
129.00	14458.0
130.00	1361.0
130.90	331.0
134.00	586.0
135.00	1068.0
135.90	502.0
137.00	684.0
139.90	238.0
141.00	2002.0
142.00	640.0
142.90	316.0
146.00	382.0
147.00	1241.0
148.00	2604.0
148.90	321.0
149.90	153.0
150.90	172.0
152.90	850.0
153.90	484.0
155.00	1258.0
156.00	1816.0
157.00	350.0
157.90	533.0
158.90	353.0
159.90	893.0
161.00	899.0
162.00	316.0
165.00	874.0

Instrument :
BNA_G
ClientSampleId :
DFTPP428

166.10	590.0
167.00	3548.0
168.00	1843.0
169.10	365.0
171.00	216.0
172.00	520.0
173.00	557.0
174.00	1089.0
175.00	1706.0
176.00	542.0
176.90	818.0
178.00	393.0
179.00	3221.0
180.00	2216.0
181.00	1055.0
183.90	272.0
185.00	1543.0
186.00	11075.0
187.10	3427.0
188.00	219.0
189.00	861.0
190.90	409.0
191.90	1010.0
193.00	1208.0
194.00	207.0
195.90	2202.0
198.00	72600.0
199.00	4881.0
199.90	404.0
201.60	352.0
203.00	723.0
204.00	2937.0
205.00	5792.0
206.00	21336.0
207.00	2768.0
208.00	670.0
209.00	272.0
210.10	468.0
211.00	688.0
215.00	248.0
216.00	487.0
217.00	5995.0
218.00	901.0
221.00	2625.0
221.90	749.0
223.00	1552.0
224.00	12087.0
225.00	3196.0
225.90	351.0
227.00	5437.0
228.00	525.0
229.00	935.0
231.00	569.0
232.10	156.0
233.90	452.0
235.00	376.0
236.20	278.0
236.90	308.0
239.00	277.0
240.10	185.0
240.90	304.0
242.00	650.0
243.10	858.0
244.00	9450.0
245.00	1219.0
246.00	2759.0
246.90	279.0
249.00	405.0
251.00	187.0
253.00	179.0
255.00	49968.0
256.00	7155.0
257.00	531.0
258.00	3382.0
259.00	483.0
265.00	1510.0
266.00	163.0
273.00	1420.0
274.00	3582.0
275.00	20264.0
276.00	2781.0
277.00	1963.0
277.90	455.0
283.10	230.0
284.10	186.0
285.10	433.0
292.90	366.0
295.00	237.0
296.00	6564.0
297.00	1020.0
303.00	719.0
304.00	187.0
308.00	159.0
314.10	406.0
315.00	982.0
316.00	280.0

Instrument :
BNA_G
ClientSampleId :
DFTPP428

321.00	174.0
323.10	1809.0
324.10	302.0
326.90	390.0
328.00	162.0
333.00	254.0
334.00	1052.0
335.00	462.0
341.00	235.0
346.10	340.0
352.00	499.0
353.00	395.0
354.00	415.0
354.90	160.0
365.00	2866.0
366.00	315.0
371.00	179.0
372.00	879.0
373.00	214.0
383.10	280.0
402.00	307.0
403.10	498.0
421.00	455.0
422.10	324.0
423.00	2320.0
424.10	673.0
441.10	8069.0
442.10	46808.0
443.10	8977.0
444.10	848.0

Instrument :
BNA_G
ClientSampleId :
DFTPP428