

Data Path : Z:\HPCHEM1\BNA G\DATA\BG110416\
 Data File : BG024685.D
 Acq On : 4 Nov 2016 21:18
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 99 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 00:38:55 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 26 11:14:13 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
2	1,4-Dioxane	0.499	0.526	-5.4	85	0.00
3	Pyridine	1.493	1.631	-9.2	88	0.00
4	n-Nitrosodimethylamine	0.773	0.823	-6.5	84	0.00
5 S	2-Fluorophenol	1.220	1.265	-3.7	85	0.00
6	Aniline	2.121	2.232	-5.2	84	0.00
7 S	Phenol-d6	1.674	1.822	-8.8	87	0.00
8	2-Chlorophenol	1.241	1.354	-9.1	90	0.00
9	Benzaldehyde	1.034	1.041	-0.7	81	0.00
10 C	Phenol	1.725	1.857	-7.7	87	0.00
11	bis(2-Chloroethyl)ether	1.296	1.407	-8.6	90	0.00
12	1,3-Dichlorobenzene	1.536	1.637	-6.6	89	0.00
13 C	1,4-Dichlorobenzene	1.517	1.626	-7.2	87	0.00
14	1,2-Dichlorobenzene	1.459	1.519	-4.1	85	0.00
15	Benzyl Alcohol	1.557	1.639	-5.3	85	0.00
16	2,2'-oxybis(1-Chloropropane	2.405	2.447	-1.7	82	0.00
17	2-Methylphenol	1.143	1.208	-5.7	86	0.00
18	Hexachloroethane	0.583	0.592	-1.5	86	0.00
19 P	n-Nitroso-di-n-propylamine	1.234	1.241	-0.6	79	0.00
20	3+4-Methylphenols	1.535	1.677	-9.3	86	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	78	0.00
22	Acetophenone	0.514	0.526	-2.3	84	0.00
23 S	Nitrobenzene-d5	0.439	0.463	-5.5	88	0.00
24	Nitrobenzene	0.489	0.508	-3.9	85	0.00
25	Isophorone	0.817	0.813	0.5	81	0.00
26 C	2-Nitrophenol	0.181	0.199	-9.9	90	0.00
27	2,4-Dimethylphenol	0.318	0.327	-2.8	83	0.00
28	bis(2-Chloroethoxy)methane	0.423	0.424	-0.2	85	0.00
29 C	2,4-Dichlorophenol	0.333	0.343	-3.0	83	0.00
30	1,2,4-Trichlorobenzene	0.395	0.410	-3.8	87	0.00
31	Naphthalene	0.998	1.048	-5.0	86	0.00
32	Benzoic acid	0.264	0.217	17.8	68	0.00
33	4-Chloroaniline	0.433	0.464	-7.2	84	0.00
34 C	Hexachlorobutadiene	0.309	0.320	-3.6	89	0.00
35	Caprolactam	0.122	0.122	0.0	81	0.00
36 C	4-Chloro-3-methylphenol	0.402	0.439	-9.2	87	0.00
37	2-Methylnaphthalene	0.728	0.763	-4.8	87	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	76	0.00
39	1,2,4,5-Tetrachlorobenzene	0.675	0.689	-2.1	86	0.00
40 P	Hexachlorocyclopentadiene	0.380	0.354	6.8	76	0.00
41 S	2,4,6-Tribromophenol	0.299	0.330	-10.4	85	0.00
42 C	2,4,6-Trichlorophenol	0.405	0.434	-7.2	88	0.00
43	2,4,5-Trichlorophenol	0.438	0.458	-4.6	84	0.00
44 S	2-Fluorobiphenyl	1.203	1.255	-4.3	86	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG110416\
 Data File : BG024685.D
 Acq On : 4 Nov 2016 21:18
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 99 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 00:38:55 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 26 11:14:13 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.388	1.421	-2.4	84	0.00
46	2-Chloronaphthalene	1.057	1.090	-3.1	86	0.00
47	2-Nitroaniline	0.433	0.466	-7.6	81	0.00
48	Acenaphthylene	1.621	1.693	-4.4	83	0.00
49	Dimethylphthalate	1.420	1.483	-4.4	82	0.00
50	2,6-Dinitrotoluene	0.303	0.336	-10.9	84	0.00
51 C	Acenaphthene	1.208	1.128	6.6	75	0.00
52	3-Nitroaniline	0.314	0.348	-10.8	86	0.00
53 P	2,4-Dinitrophenol	0.220	0.192	12.7	65	0.00
54	Dibenzofuran	1.670	1.757	-5.2	84	0.00
55 P	4-Nitrophenol	0.273	0.267	2.2	76	0.00
56	2,4-Dinitrotoluene	0.440	0.489	-11.1	82	0.00
57	Fluorene	1.385	1.449	-4.6	82	0.00
58	2,3,4,6-Tetrachlorophenol	0.454	0.472	-4.0	80	0.00
59	Diethylphthalate	1.489	1.506	-1.1	79	0.00
60	4-Chlorophenyl-phenylether	0.777	0.805	-3.6	84	0.00
61	4-Nitroaniline	0.343	0.375	-9.3	85	0.00
62	Azobenzene	1.485	1.503	-1.2	80	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	75	0.00
64	4,6-Dinitro-2-methylphenol	0.138	0.134	2.9	75	0.00
65 c	n-Nitrosodiphenylamine	0.547	0.563	-2.9	82	0.00
66	4-Bromophenyl-phenylether	0.243	0.254	-4.5	84	0.00
67	Hexachlorobenzene	0.268	0.286	-6.7	84	0.00
68	Atrazine	0.235	0.238	-1.3	79	0.00
69 C	Pentachlorophenol	0.177	0.154	13.0	66	0.00
70	Phenanthrene	1.019	1.073	-5.3	85	0.00
71	Anthracene	1.028	1.090	-6.0	84	0.00
72	Carbazole	0.938	0.996	-6.2	86	0.00
73	Di-n-butylphthalate	1.081	1.137	-5.2	84	0.00
74 C	Fluoranthene	1.289	1.416	-9.9	86	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
76	Benzidine	0.471	0.424	10.0	75	0.00
77	Pyrene	0.978	1.015	-3.8	85	0.00
78 S	Terphenyl-d14	0.669	0.689	-3.0	87	0.00
79	Butylbenzylphthalate	0.408	0.413	-1.2	85	0.00
80	Benzo(a)anthracene	1.099	1.107	-0.7	86	0.00
81	3,3'-Dichlorobenzidine	0.443	0.436	1.6	83	0.00
82	Chrysene	1.046	1.074	-2.7	87	0.00
83	Bis(2-ethylhexyl)phthalate	0.583	0.587	-0.7	85	0.00
84 c	Di-n-octyl phthalate	0.968	0.977	-0.9	84	-0.02
85	Indeno(1,2,3-cd)pyrene	1.444	1.376	4.7	84	-0.03
86 I	Perylene-d12	1.000	1.000	0.0	77	-0.02
87	Benzo(b)fluoranthene	1.081	1.118	-3.4	86	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG110416\
 Data File : BG024685.D
 Acq On : 4 Nov 2016 21:18
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 99 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 00:38:55 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 26 11:14:13 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.044	1.096	-5.0	86	-0.02
89 C	Benzo(a)pyrene	1.056	1.085	-2.7	84	0.00
90	Dibenzo(a,h)anthracene	1.159	1.131	2.4	83	-0.03
91	Benzo(a,h,i)perylene	1.158	1.119	3.4	83	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0