

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG110419\
 Data File : BG043484.D
 Acq On : 4 Nov 2019 17:12
 Operator : JU
 Sample : K5612-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DRUM-COMP

Manual Integrations
 APPROVED

mohammad
 11/5/2019 10:41:08 AM

Quant Time: Nov 05 10:31:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG102119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 04 15:22:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.01	152	42713	20.00	ng	0.00
21) Naphthalene-d8	10.82	136	167461	20.00	ng	0.00
39) Acenaphthene-d10	14.64	164	662907m	20.00	ng	0.00
64) Phenanthrene-d10	17.39	188	274508	20.00	ng	0.00
76) Chrysene-d12	21.65	240	312638	20.00	ng	0.00
87) Perylene-d12	24.85	264	372691	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	171659	73.64	ng	0.00
7) Phenol-d6	7.19	99	168788	50.33	ng	0.00
23) Nitrobenzene-d5	9.17	82	216095	66.53	ng	0.00
42) 2,4,6-Tribromophenol	16.13	330	228417	18.11	ng	0.00
45) 2-Fluorobiphenyl	13.26	172	569771	11.88	ng	0.00
79) Terphenyl-d14	20.00	244	1154852	69.30	ng	0.00
Target Compounds						
6) Aniline	7.34	93	10481	2.713	ng	94
10) Phenol	7.22	94	29141	9.509	ng	95
32) Benzoic acid	10.22	122	106918	62.118	ng	94
50) Dimethylphthalate	14.09	163	175055	3.409	ng	98
74) Di-n-butylphthalate	18.34	149	145518	9.417	ng	100
84) Bis(2-ethylhexyl)phthalate	21.54	149	42370	4.068	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG110419\
Data File : BG043484.D
Acq On : 4 Nov 2019 17:12
Operator : JU
Sample : K5612-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DRUM-COMP

Manual Integrations
APPROVED

mohammad
11/5/2019 10:41:08 AM

Quant Time: Nov 05 10:31:53 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG102119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 04 15:22:15 2019
Response via : Initial Calibration

