

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110420\
 Data File : BG047214.D
 Acq On : 4 Nov 2020 16:56
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG110420

Quant Time: Nov 04 17:37:57 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 04 16:19:40 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	117	0.00
2	1,4-Dioxane	40.000	39.263	1.8	116	0.00
3	Pyridine	40.000	41.011	-2.5	112	0.00
4	n-Nitrosodimethylamine	40.000	40.351	-0.9	111	0.00
5 S	2-Fluorophenol	80.000	79.491	0.6	112	0.00
6	Aniline	40.000	40.136	-0.3	110	0.00
7 S	Phenol-d6	80.000	80.434	-0.5	111	0.00
8	2-Chlorophenol	40.000	39.632	0.9	112	0.00
9	Benzaldehyde	40.000	41.157	-2.9	116	0.00
10 C	Phenol	40.000	39.551	1.1	110	0.00
11	bis(2-Chloroethyl)ether	40.000	39.128	2.2	109	0.00
12	1,3-Dichlorobenzene	40.000	39.521	1.2	113	0.00
13 C	1,4-Dichlorobenzene	40.000	39.625	0.9	111	0.00
14	1,2-Dichlorobenzene	40.000	39.758	0.6	115	0.00
15	Benzyl Alcohol	40.000	41.707	-4.3	111	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.633	0.9	112	0.00
17	2-Methylphenol	40.000	39.631	0.9	111	0.00
18	Hexachloroethane	40.000	38.875	2.8	111	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.253	-0.6	109	0.00
20	3+4-Methylphenols	40.000	41.098	-2.7	113	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	111	0.00
22	Acetophenone	40.000	41.029	-2.6	111	0.00
23 S	Nitrobenzene-d5	80.000	82.938	-3.7	110	0.00
24	Nitrobenzene	40.000	40.477	-1.2	109	0.00
25	Isophorone	40.000	41.615	-4.0	109	0.00
26 C	2-Nitrophenol	40.000	41.340	-3.4	108	0.00
27	2,4-Dimethylphenol	40.000	41.553	-3.9	109	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.482	-1.2	108	0.00
29 C	2,4-Dichlorophenol	40.000	41.398	-3.5	108	0.00
30	1,2,4-Trichlorobenzene	40.000	40.303	-0.8	111	0.00
31	Naphthalene	40.000	40.619	-1.5	111	0.00
32	Benzoic acid	40.000	39.894	0.3	106	0.00
33	4-Chloroaniline	40.000	40.895	-2.2	106	0.00
34 C	Hexachlorobutadiene	40.000	39.842	0.4	112	0.00
35	Caprolactam	40.000	43.167	-7.9	105	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.648	-4.1	107	0.00
37	2-Methylnaphthalene	40.000	41.061	-2.7	109	0.00
38	1-Methylnaphthalene	40.000	40.348	-0.9	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.183	-5.5	111	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.014	-7.5	108	0.00
42 S	2,4,6-Tribromophenol	80.000	87.453	-9.3	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.449	-6.1	108	0.00
44	2,4,5-Trichlorophenol	40.000	42.673	-6.7	106	0.00
45 S	2-Fluorobiphenyl	80.000	82.770	-3.5	108	0.00
46	1,1'-Biphenyl	40.000	40.853	-2.1	106	0.00
47	2-Chloronaphthalene	40.000	41.972	-4.9	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110420\
 Data File : BG047214.D
 Acq On : 4 Nov 2020 16:56
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG110420

Quant Time: Nov 04 17:37:57 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 04 16:19:40 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.553	-11.4	109	0.00
49	Acenaphthylene	40.000	41.643	-4.1	105	0.00
50	Dimethylphthalate	40.000	41.906	-4.8	106	0.00
51	2,6-Dinitrotoluene	40.000	44.472	-11.2	111	0.00
52 C	Acenaphthene	40.000	42.717	-6.8	109	0.00
53	3-Nitroaniline	40.000	45.130	-12.8	111	0.00
54 P	2,4-Dinitrophenol	40.000	45.032	-12.6	111	0.00
55	Dibenzofuran	40.000	41.985	-5.0	106	0.00
56 P	4-Nitrophenol	40.000	46.870	-17.2	109	0.00
57	2,4-Dinitrotoluene	40.000	43.815	-9.5	106	0.00
58	Fluorene	40.000	42.068	-5.2	105	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.789	-7.0	105	0.00
60	Diethylphthalate	40.000	42.539	-6.3	106	0.00
61	4-Chlorophenyl-phenylether	40.000	42.258	-5.6	107	0.00
62	4-Nitroaniline	40.000	45.867	-14.7	112	0.00
63	Azobenzene	40.000	41.562	-3.9	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.353	-5.9	110	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.643	-1.6	107	0.00
67	4-Bromophenyl-phenylether	40.000	41.174	-2.9	108	0.00
68	Hexachlorobenzene	40.000	41.069	-2.7	107	0.00
69	Atrazine	40.000	41.938	-4.8	110	0.00
70 C	Pentachlorophenol	40.000	44.365	-10.9	111	0.00
71	Phenanthrene	40.000	41.169	-2.9	107	0.00
72	Anthracene	40.000	40.841	-2.1	107	0.00
73	Carbazole	40.000	41.661	-4.2	109	0.00
74	Di-n-butylphthalate	40.000	42.000	-5.0	108	0.00
75 C	Fluoranthene	40.000	41.412	-3.5	110	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	113	0.00
77	Benzydine	40.000	42.396	-6.0	113	0.00
78	Pyrene	40.000	41.347	-3.4	110	0.00
79 S	Terphenyl-d14	80.000	82.487	-3.1	110	0.00
80	Butylbenzylphthalate	40.000	42.147	-5.4	111	0.00
81	Benzo(a)anthracene	40.000	41.501	-3.8	114	0.00
82	3,3'-Dichlorobenzidine	40.000	43.014	-7.5	116	0.00
83	Chrysene	40.000	40.862	-2.2	113	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.724	-6.8	114	0.00
85 c	Di-n-octyl phthalate	40.000	42.630	-6.6	114	0.00
86 I	Perylene-d12	20.000	20.000	0.0	112	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.671	-4.2	113	0.00
88	Benzo(b)fluoranthene	40.000	41.468	-3.7	113	0.00
89	Benzo(k)fluoranthene	40.000	41.905	-4.8	114	0.00
90 C	Benzo(a)pyrene	40.000	41.940	-4.8	114	0.00
91	Dibenzo(a,h)anthracene	40.000	41.063	-2.7	112	0.00
92	Benzo(g,h,i)perylene	40.000	41.097	-2.7	112	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110420\
Data File : BG047214.D
Acq On : 4 Nov 2020 16:56
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG110420

Quant Time: Nov 04 17:37:57 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG110420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 04 16:19:40 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0