

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
 Data File : BG059660.D
 Acq On : 5 Nov 2023 18:24
 Operator : MA/JU
 Sample : 05013-17
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 A4992

Quant Time: Nov 06 02:26:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Nov 04 02:46:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.353	152	21454	20.000	ng/u1	# 0.00
20) Naphthalene-d8	11.196	136	103121	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.980	164	77806	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.730	188	173512	20.000	ng/u1	# 0.00
79) Chrysene-d12	22.066	240	199312	20.000	ng/u1	0.00
88) Perylene-d12	25.668	264	231210	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.652	96	2200	3.843	ng/uL	0.00
4) Pyridine-d5	0.000	84	0d	0.000	ng/u1	
7) Phenol-d5	7.460	99	17511	7.983	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.665	67	33423	28.500	ng/u1	0.00
11) 2-Chlorophenol-d4	7.865	132	42297	28.393	ng/u1	0.00
15) 4-Methylphenol-d8	9.034	113	32352	18.022	ng/u1	0.00
21) Nitrobenzene-d5	9.540	128	29739	39.434	ng/u1	-0.01
24) 2-Nitrophenol-d4	10.268	143	33124	36.396	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.791	165	59252	36.645	ng/u1	0.00
31) 4-Chloroaniline-d4	11.337	131	9393	3.692	ng/u1	0.00
46) Dimethylphthalate-d6	14.375	166	263870	42.257	ng/u1	0.00
49) Acenaphthylene-d8	14.681	160	270041	36.267	ng/u1	0.00
54) 4-Nitrophenol-d4	15.145	143	10887	9.875	ng/u1	-0.01
60) Fluorene-d10	15.967	176	222103	37.780	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.067	200	43306	39.432	ng/u1	0.00
73) Anthracene-d10	17.830	188	341552	37.611	ng/u1	0.00
81) Pyrene-d10	20.098	212	506160	39.684	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.409	264	537124	39.684	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059660.D
Acq On : 5 Nov 2023 18:24
Operator : MA/JU
Sample : 05013-17
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
A4992

Quant Time: Nov 06 02:26:51 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration

