




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:14:58 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(34) Caprolactam




12.141min (-0.006) 12.21 ng/ul

response 14981

Ion	Exp%	Act%
113.00	100.00	100.00
55.00	59.30	78.61#
56.00	83.20	79.27
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:14:58 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(34) Caprolactam




12.141min (-0.006) 21.72 ng/ul m

response 26657

Ion	Exp%	Act%
113.00	100.00	100.00
55.00	59.30	78.61#
56.00	83.20	79.27
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:14:58 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(64) Phenanthrene-d10 (I)




17.829min (+ 0.094) 20.00 ng/u1

response 296874

Ion	Exp%	Act%
188.00	100.00	100.00
94.00	6.20	5.55
80.00	6.40	5.24
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:14:58 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(64) Phenanthrene-d10 (I)




17.729min (-0.006) 20.00 ng/ul m

response 295482

Ion	Exp%	Act%
188.00	100.00	100.00
94.00	6.20	3.37#
80.00	6.40	5.08#
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:16:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(66) 4,6-Dinitro-2-methylphenol




16.125min (+ 0.035) 0.05 ng/ul

response 86

Ion	Exp%	Act%
198.00	100.00	100.00
182.00	5.40	76.23#
77.00	36.50	0.00#
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:16:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(66) 4,6-Dinitro-2-methylphenol




16.084min (-0.006) 20.77 ng/ul m

response 37708

Ion	Exp%	Act%
198.00	100.00	100.00
182.00	5.40	3.29#
77.00	36.50	49.93#
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:16:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(71) Pentachlorophenol (C)




17.447min (+ 0.094) 0.03 ng/u1

response 69

Ion	Exp%	Act%
265.70	100.00	100.00
264.00	58.30	0.00#
268.00	62.90	0.00#
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:16:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(71) Pentachlorophenol (C)




17.347min (-0.006) 22.34 ng/ul m

response 45562

Ion	Exp%	Act%
265.70	100.00	100.00
264.00	58.30	83.32#
268.00	62.90	75.98#
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:16:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(72) Phenanthrene




17.864min (+ 0.088) 19.96 ng/u1

response 345416

Ion	Exp%	Act%
178.00	100.00	100.00
179.00	14.70	16.44
176.00	20.50	20.15
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:16:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(72) Phenanthrene




17.776min (-0.000) 19.50 ng/ul m

response 337437

Ion	Exp%	Act%
178.00	100.00	100.00
179.00	14.70	13.95
176.00	20.50	17.91
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:16:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(85) Benzo(a)anthracene




22.112min (+ 0.065) 17.99 ng/u1

response 478329

Ion	Exp%	Act%
228.00	100.00	100.00
229.00	22.10	21.27
226.00	26.40	28.34
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:16:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(85) Benzo(a)anthracene




22.041min (-0.006) 19.64 ng/ul m

response 522123

Ion	Exp%	Act%
228.00	100.00	100.00
229.00	22.10	17.40#
226.00	26.40	25.00
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:16:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(94) Indeno(1,2,3-cd)pyrene




29.838min (-0.018) 17.69 ng/u1

response 592020

Ion	Exp%	Act%
276.00	100.00	100.00
138.00	6.80	4.79#
277.00	23.10	25.73
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:16:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(94) Indeno(1,2,3-cd)pyrene




29.838min (-0.018) 19.83 ng/ul m

response 663824

Ion	Exp%	Act%
276.00	100.00	100.00
138.00	6.80	4.79#
277.00	23.10	25.73
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:16:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(96) Benzo(g,h,i)perylene




31.148min (-0.047) 9.71 ng/u1

response 255219

Ion	Exp%	Act%
276.00	100.00	100.00
138.00	5.70	8.24#
277.00	21.90	22.21
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:16:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration



TIC: BG059689.D\data.ms

(96) Benzo(g,h,i)perylene

31.172min (-0.024) 19.80 ng/ul m

response 520368

Ion	Exp%	Act%
276.00	100.00	100.00
138.00	5.70	6.27
277.00	21.90	20.29
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
 Data File : BG059689.D
 Acq On : 6 Nov 2023 14:30
 Operator : MA/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:18:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Nov 04 02:46:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.352	152	36661	20.000	ng/ul	# 0.00
20) Naphthalene-d8	11.195	136	232203	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.979	164	123014	20.000	ng/ul	# 0.00
64) Phenanthrene-d10	17.729	188	295482m	20.000	ng/ul	0.00
79) Chrysene-d12	22.065	240	339137	20.000	ng/ul	# 0.00
88) Perylene-d12	25.667	264	397534	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.645	96	8422	8.610	ng/uL	0.00
4) Pyridine-d5	4.074	84	69848	24.283	ng/ul	0.00
7) Phenol-d5	7.459	99	73368	19.574	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.664	67	30664	15.301	ng/ul	0.00
11) 2-Chlorophenol-d4	7.864	132	53391	20.974	ng/ul	0.00
15) 4-Methylphenol-d8	9.033	113	83189	27.119	ng/ul	0.00
21) Nitrobenzene-d5	9.544	128	36128	21.275	ng/ul	0.00
24) 2-Nitrophenol-d4	10.267	143	40373	19.701	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.790	165	84789	23.288	ng/ul	0.00
31) 4-Chloroaniline-d4	11.342	131	115588	20.174	ng/ul	0.00
46) Dimethylphthalate-d6	14.368	166	213884	21.665	ng/ul	0.00
49) Acenaphthylene-d8	14.680	160	236227	20.066	ng/ul	0.00
54) 4-Nitrophenol-d4	15.155	143	33527	19.234	ng/ul	0.00
60) Fluorene-d10	15.960	176	192507	20.712	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	16.072	200	42049	22.483	ng/ul	0.00
73) Anthracene-d10	17.829	188	296874	19.197	ng/ul	0.00
81) Pyrene-d10	20.097	212	312243	14.387	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.402	264	453642	19.493	ng/ul	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.681	88	8693	7.778	ng/ul#	79
5) Pyridine	4.098	79	71448	23.520	ng/ul#	91
6) Benzaldehyde	7.494	77	48510	20.471	ng/ul	97
8) Phenol	7.488	94	72585	19.449	ng/ul#	85
10) Bis(2-Chloroethyl)ether	7.764	93	46211	17.922	ng/ul	98
12) 2-Chlorophenol	7.893	128	51535	19.406	ng/ul	90
13) 2-Methylphenol	8.769	108	57189	18.522	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.863	45	17343	15.039	ng/ul#	90
16) Acetophenone	9.192	105	130183	23.921	ng/ul#	86
17) N-Nitroso-di-n-propyla...	9.151	70	70674	24.580	ng/ul#	84
18) 4-Methylphenol	9.098	108	83904	25.383	ng/ul	94
19) Hexachloroethane	9.439	117	31528	27.082	ng/ul#	79
22) Nitrobenzene	9.591	77	111364	22.064	ng/ul#	96
23) Isophorone	10.097	82	206798	17.791	ng/ul	98
25) 2-Nitrophenol	10.296	139	38667	17.527	ng/ul#	23
26) 2,4-Dimethylphenol	10.320	107	117719	21.048	ng/ul	94
27) Bis(2-Chloroethoxy)met...	10.584	93	85652	17.081	ng/ul	92
29) 2,4-Dichlorophenol	10.825	162	78024	22.170	ng/ul#	92
30) Naphthalene	11.248	128	257186	18.917	ng/ul#	93
32) 4-Chloroaniline	11.360	127	106681	19.572	ng/ul	96
33) Hexachlorobutadiene	11.477	225	55312	22.345	ng/ul#	86
34) Caprolactam	12.141	113	26657m	21.717	ng/ul	
35) 4-Chloro-3-methylphenol	12.429	107	111300	23.351	ng/ul#	80

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
 Data File : BG059689.D
 Acq On : 6 Nov 2023 14:30
 Operator : MA/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:18:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Nov 04 02:46:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 2-Methylnaphthalene	12.823	142	151150	15.258	ng/ul	89
37) 1-Methylnaphthalene	13.040	142	148730	14.700	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	13.170	216	72486	21.092	ng/ul#	96
40) Hexachlorocyclopentadiene	13.123	237	53131	21.472	ng/ul#	81
41) 2,4,6-Trichlorophenol	13.405	196	53166	23.802	ng/ul#	93
42) 2,4,5-Trichlorophenol	13.475	196	58675	23.786	ng/ul#	94
43) 1,1'-Biphenyl	13.816	154	211400	20.103	ng/ul#	90
44) 2-Chloronaphthalene	13.869	162	160144	20.444	ng/ul	94
45) 2-Nitroaniline	14.080	65	51324	26.142	ng/ul#	83
47) Dimethylphthalate	14.415	163	217041	21.599	ng/ul#	96
48) 2,6-Dinitrotoluene	14.562	165	39194	21.645	ng/ul#	86
50) Acenaphthylene	14.709	152	272178	20.491	ng/ul	99
51) 3-Nitroaniline	14.903	138	43856	22.009	ng/ul#	84
52) Acenaphthene	15.044	153	171909	19.631	ng/ul	95
53) 2,4-Dinitrophenol	15.091	184	26143	20.506	ng/ul#	64
55) 4-Nitrophenol	15.167	109	79346	25.606	ng/ul#	82
56) Dibenzofuran	15.373	168	254465	21.412	ng/ul	96
57) 2,4-Dinitrotoluene	15.338	165	63854	23.595	ng/ul#	58
58) 2,3,4,6-Tetrachlorophenol	15.578	232	51725	24.225	ng/ul#	89
59) Diethylphthalate	15.766	149	234492	22.185	ng/ul	97
61) Fluorene	16.025	166	212445	20.827	ng/ul	99
62) 4-Chlorophenyl-phenyle...	16.001	204	108714	21.906	ng/ul	97
63) 4-Nitroaniline	16.060	138	41857	21.725	ng/ul#	61
66) 4,6-Dinitro-2-methylph...	16.084	198	37708m	20.774	ng/ul	
67) N-Nitrosodiphenylamine	16.219	169	188419	19.450	ng/ul	97
68) 4-Bromophenyl-phenylether	16.900	248	69271	22.180	ng/ul	92
69) Hexachlorobenzene	16.994	284	71323	24.390	ng/ul	95
70) Atrazine	17.153	200	76539	23.071	ng/ul#	88
71) Pentachlorophenol	17.347	266	45562m	22.335	ng/ul	
72) Phenanthrene	17.776	178	337437m	19.495	ng/ul	
74) Anthracene	17.864	178	345416	18.891	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.775	216	73973	20.472	ng/uL#	77
76) Pentachlorobenzene	15.267	250	76303	20.889	ng/uL#	82
77) Carbazole	18.140	167	296829	18.861	ng/ul	99
78) Di-n-butylphthalate	18.645	149	365951	20.542	ng/ul#	96
80) Fluoranthene	19.762	202	394536	14.826	ng/ul#	92
82) Pyrene	20.126	202	384942	14.374	ng/ul#	93
83) Butylbenzylphthalate	20.990	149	161653	16.103	ng/ul	85
84) 3,3'-Dichlorobenzidine	21.947	252	192628	22.971	ng/ul	98
85) Benzo(a)anthracene	22.041	228	522123m	19.639	ng/ul	
86) Bis(2-ethylhexyl)phtha...	21.871	149	303511	20.248	ng/ul#	94
87) Chrysene	22.112	228	478030	19.804	ng/ul	97
89) Di-n-octyl phthalate	23.217	149	512817	23.393	ng/ul	100
90) Benzo(b)fluoranthene	24.486	252	546916	20.202	ng/ul#	95
91) Benzo(k)fluoranthene	24.568	252	536601	19.432	ng/ul#	93
93) Benzo(a)pyrene	25.485	252	500630	19.127	ng/ul#	94
94) Indeno(1,2,3-cd)pyrene	29.838	276	663824m	19.833	ng/ul	
95) Dibenzo(a,h)anthracene	29.920	278	551403	19.855	ng/ul	98
96) Benzo(g,h,i)perylene	31.172	276	520368m	19.803	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110423\
Data File : BG059689.D
Acq On : 6 Nov 2023 14:30
Operator : MA/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Quant Time: Nov 06 16:18:39 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG110323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 04 02:46:50 2023
Response via : Initial Calibration

