

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110424\
 Data File : BG063150.D
 Acq On : 4 Nov 2024 10:44
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 04 11:09:27 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	122	0.00
2	1,4-Dioxane	40.000	37.411	6.5	113	-0.02
3	Pyridine	40.000	38.590	3.5	114	-0.02
4	n-Nitrosodimethylamine	40.000	39.975	0.1	110	-0.02
5 S	2-Fluorophenol	80.000	71.058	11.2	114	0.00
6	Aniline	40.000	38.405	4.0	117	0.00
7 S	Phenol-d6	80.000	76.389	4.5	117	0.00
8	2-Chlorophenol	40.000	38.431	3.9	117	-0.01
9	Benzaldehyde	40.000	38.710	3.2	117	0.00
10 C	Phenol	40.000	38.513	3.7	119	0.00
11	bis(2-Chloroethyl)ether	40.000	38.070	4.8	118	0.00
12	1,3-Dichlorobenzene	40.000	36.274	9.3	112	0.00
13 C	1,4-Dichlorobenzene	40.000	36.906	7.7	115	0.00
14	1,2-Dichlorobenzene	40.000	36.796	8.0	114	0.00
15	Benzyl Alcohol	40.000	39.912	0.2	119	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.943	2.6	121	0.00
17	2-Methylphenol	40.000	40.020	-0.1	125	0.00
18	Hexachloroethane	40.000	38.091	4.8	115	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.279	1.8	120	0.00
20	3+4-Methylphenols	40.000	39.477	1.3	120	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	122	0.00
22	Acetophenone	40.000	38.945	2.6	117	0.00
23 S	Nitrobenzene-d5	80.000	77.018	3.7	119	0.00
24	Nitrobenzene	40.000	39.386	1.5	122	0.00
25	Isophorone	40.000	40.757	-1.9	123	0.00
26 C	2-Nitrophenol	40.000	41.012	-2.5	127	0.00
27	2,4-Dimethylphenol	40.000	40.486	-1.2	125	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.278	-0.7	123	0.00
29 C	2,4-Dichlorophenol	40.000	38.762	3.1	117	-0.01
30	1,2,4-Trichlorobenzene	40.000	37.805	5.5	115	0.00
31	Naphthalene	40.000	39.535	1.2	120	0.00
32	Benzoic acid	40.000	37.865	5.3	112	0.00
33	4-Chloroaniline	40.000	41.954	-4.9	125	0.00
34 C	Hexachlorobutadiene	40.000	36.727	8.2	111	0.00
35	Caprolactam	40.000	41.511	-3.8	124	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.634	3.4	120	0.00
37	2-Methylnaphthalene	40.000	38.395	4.0	121	0.00
38	1-Methylnaphthalene	40.000	38.970	2.6	119	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	125	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.468	8.8	115	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.697	8.3	112	0.00
42 S	2,4,6-Tribromophenol	80.000	78.844	1.4	125	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.919	7.7	112	0.00
44	2,4,5-Trichlorophenol	40.000	38.049	4.9	116	0.00
45 S	2-Fluorobiphenyl	80.000	74.253	7.2	116	0.00
46	1,1'-Biphenyl	40.000	37.766	5.6	119	0.00
47	2-Chloronaphthalene	40.000	38.093	4.8	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110424\
 Data File : BG063150.D
 Acq On : 4 Nov 2024 10:44
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 04 11:09:27 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.272	-0.7	121	0.00
49	Acenaphthylene	40.000	39.071	2.3	125	0.00
50	Dimethylphthalate	40.000	39.686	0.8	128	0.00
51	2,6-Dinitrotoluene	40.000	40.402	-1.0	127	0.00
52 C	Acenaphthene	40.000	39.531	1.2	126	0.00
53	3-Nitroaniline	40.000	43.082	-7.7	133	0.00
54 P	2,4-Dinitrophenol	40.000	39.294	1.8	119	0.00
55	Dibenzofuran	40.000	39.694	0.8	127	0.00
56 P	4-Nitrophenol	40.000	44.919	-12.3	138	0.00
57	2,4-Dinitrotoluene	40.000	43.018	-7.5	135	0.00
58	Fluorene	40.000	40.502	-1.3	128	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.948	0.1	125	0.00
60	Diethylphthalate	40.000	41.300	-3.2	129	0.00
61	4-Chlorophenyl-phenylether	40.000	39.823	0.4	126	0.00
62	4-Nitroaniline	40.000	44.422	-11.1	142	0.00
63	Azobenzene	40.000	41.485	-3.7	131	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	130	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.276	-0.7	132	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.425	3.9	129	0.00
67	4-Bromophenyl-phenylether	40.000	38.237	4.4	127	0.00
68	Hexachlorobenzene	40.000	38.055	4.9	126	0.00
69	Atrazine	40.000	32.466	18.8	132	0.00
70 C	Pentachlorophenol	40.000	39.970	0.1	123	0.00
71	Phenanthrene	40.000	39.231	1.9	132	0.00
72	Anthracene	40.000	39.538	1.2	131	0.00
73	Carbazole	40.000	40.243	-0.6	138	0.00
74	Di-n-butylphthalate	40.000	39.899	0.3	137	0.00
75 C	Fluoranthene	40.000	37.985	5.0	133	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	138	0.00
77	Benzydine	40.000	24.383	39.0#	99	0.00
78	Pyrene	40.000	39.530	1.2	133	0.00
79 S	Terphenyl-d14	80.000	74.561	6.8	126	0.00
80	Butylbenzylphthalate	40.000	40.343	-0.9	138	0.00
81	Benzo(a)anthracene	40.000	38.309	4.2	132	0.00
82	3,3'-Dichlorobenzidine	40.000	38.156	4.6	133	-0.01
83	Chrysene	40.000	38.519	3.7	132	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.000	0.0	137	0.00
85 c	Di-n-octyl phthalate	40.000	40.430	-1.1	139	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	137	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	39.392	1.5	136	-0.01
88	Benzo(b)fluoranthene	40.000	39.303	1.7	137	0.00
89	Benzo(k)fluoranthene	40.000	38.888	2.8	132	-0.01
90 C	Benzo(a)pyrene	40.000	38.767	3.1	136	0.00
91	Dibenzo(a,h)anthracene	40.000	39.522	1.2	137	-0.02
92	Benzo(g,h,i)perylene	40.000	39.207	2.0	136	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110424\
Data File : BG063150.D
Acq On : 4 Nov 2024 10:44
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 11:09:27 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 01 05:13:40 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0