

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110424\
 Data File : BG063163.D
 Acq On : 4 Nov 2024 20:05
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 01:47:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.507	0.496	2.2	98	-0.01
3	Pyridine	1.223	1.375	-12.4	110	-0.01
4	n-Nitrosodimethylamine	0.760	0.744	2.1	89	-0.01
5 S	2-Fluorophenol	1.290	1.205	6.6	99	0.00
6	Aniline	1.520	1.466	3.6	97	0.00
7 S	Phenol-d6	1.690	1.655	2.1	99	0.00
8	2-Chlorophenol	1.198	1.204	-0.5	101	0.00
9	Benzaldehyde	1.050	1.132	-7.8	108	0.00
10 C	Phenol	1.811	1.747	3.5	99	0.00
11	bis(2-Chloroethyl)ether	1.221	1.199	1.8	100	0.00
12	1,3-Dichlorobenzene	1.453	1.349	7.2	95	0.00
13 C	1,4-Dichlorobenzene	1.481	1.431	3.4	99	0.00
14	1,2-Dichlorobenzene	1.428	1.344	5.9	97	0.00
15	Benzyl Alcohol	1.358	1.334	1.8	97	0.00
16	2,2'-oxybis(1-Chloropropane	2.215	2.190	1.1	101	0.00
17	2-Methylphenol	1.152	1.110	3.6	99	0.00
18	Hexachloroethane	0.530	0.529	0.2	99	0.00
19 P	n-Nitroso-di-n-propylamine	1.177	1.154	2.0	99	0.00
20	3+4-Methylphenols	1.536	1.533	0.2	100	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.536	0.533	0.6	101	0.00
23 S	Nitrobenzene-d5	0.460	0.446	3.0	101	0.00
24	Nitrobenzene	0.497	0.478	3.8	101	0.00
25	Isophorone	0.795	0.779	2.0	100	0.00
26 C	2-Nitrophenol	0.183	0.175	4.4	100	0.00
27	2,4-Dimethylphenol	0.225	0.231	-2.7	107	0.00
28	bis(2-Chloroethoxy)methane	0.417	0.412	1.2	102	0.00
29 C	2,4-Dichlorophenol	0.340	0.317	6.8	96	0.00
30	1,2,4-Trichlorobenzene	0.418	0.394	5.7	97	0.00
31	Naphthalene	1.055	1.037	1.7	101	0.00
32	Benzoic acid	0.196	0.151	23.0	77	0.00
33	4-Chloroaniline	0.356	0.361	-1.4	102	0.00
34 C	Hexachlorobutadiene	0.344	0.325	5.5	97	0.00
35	Caprolactam	0.108	0.101	6.5	94	0.00
36 C	4-Chloro-3-methylphenol	0.391	0.372	4.9	100	0.00
37	2-Methylnaphthalene	0.748	0.714	4.5	102	0.00
38	1-Methylnaphthalene	0.737	0.701	4.9	98	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	0.719	0.686	4.6	97	0.00
41 P	Hexachlorocyclopentadiene	0.234	0.225	3.8	94	0.00
42 S	2,4,6-Tribromophenol	0.340	0.329	3.2	99	0.00
43 C	2,4,6-Trichlorophenol	0.430	0.413	4.0	93	0.00
44	2,4,5-Trichlorophenol	0.473	0.468	1.1	97	0.00
45 S	2-Fluorobiphenyl	1.350	1.300	3.7	97	0.00
46	1,1'-Biphenyl	1.357	1.292	4.8	97	0.00
47	2-Chloronaphthalene	1.137	1.096	3.6	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110424\
 Data File : BG063163.D
 Acq On : 4 Nov 2024 20:05
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 01:47:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.399	0.396	0.8	96	0.00
49	Acenaphthylene	1.605	1.565	2.5	100	0.00
50	Dimethylphthalate	1.425	1.363	4.4	99	0.00
51	2,6-Dinitrotoluene	0.304	0.297	2.3	99	0.00
52 C	Acenaphthene	0.994	0.964	3.0	99	0.00
53	3-Nitroaniline	0.289	0.299	-3.5	103	0.00
54 P	2,4-Dinitrophenol	0.186	0.162	12.9	85	0.00
55	Dibenzofuran	1.752	1.721	1.8	101	0.00
56 P	4-Nitrophenol	0.219	0.241	-10.0	109	0.00
57	2,4-Dinitrotoluene	0.435	0.441	-1.4	102	0.00
58	Fluorene	1.388	1.335	3.8	97	0.00
59	2,3,4,6-Tetrachlorophenol	0.454	0.424	6.6	94	0.00
60	Diethylphthalate	1.464	1.430	2.3	99	0.00
61	4-Chlorophenyl-phenylether	0.863	0.841	2.5	99	0.00
62	4-Nitroaniline	0.319	0.330	-3.4	106	0.00
63	Azobenzene	1.514	1.494	1.3	101	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	0.129	0.126	2.3	99	0.00
66 c	n-Nitrosodiphenylamine	0.499	0.496	0.6	103	0.00
67	4-Bromophenyl-phenylether	0.245	0.244	0.4	101	0.00
68	Hexachlorobenzene	0.284	0.280	1.4	100	0.00
69	Atrazine	0.229	0.186	18.8	101	0.00
70 C	Pentachlorophenol	0.141	0.137	2.8	92	0.00
71	Phenanthrene	0.970	0.967	0.3	103	0.00
72	Anthracene	0.952	0.945	0.7	101	0.00
73	Carbazole	0.896	0.921	-2.8	108	0.00
74	Di-n-butylphthalate	1.017	1.048	-3.0	108	0.00
75 C	Fluoranthene	1.372	1.440	-5.0	113	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
77	Benzidine	0.308	0.471	-52.9#	215#	0.00
78	Pyrene	1.186	1.161	2.1	114	0.00
79 S	Terphenyl-d14	0.972	0.937	3.6	112	0.00
80	Butylbenzylphthalate	0.382	0.393	-2.9	122	0.00
81	Benzo(a)anthracene	1.259	1.225	2.7	116	0.00
82	3,3'-Dichlorobenzidine	0.467	0.479	-2.6	124	-0.01
83	Chrysene	1.183	1.154	2.5	116	0.00
84	Bis(2-ethylhexyl)phthalate	0.554	0.550	0.7	118	0.00
85 c	Di-n-octyl phthalate	0.927	0.937	-1.1	120	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	123	-0.01
87	Indeno(1,2,3-cd)pyrene	1.342	1.295	3.5	120	-0.02
88	Benzo(b)fluoranthene	1.166	1.100	5.7	118	-0.01
89	Benzo(k)fluoranthene	1.112	1.097	1.3	120	-0.01
90 C	Benzo(a)pyrene	1.018	0.971	4.6	120	-0.01
91	Dibenzo(a,h)anthracene	1.120	1.090	2.7	121	-0.02
92	Benzo(g,h,i)perylene	1.108	1.060	4.3	119	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110424\
Data File : BG063163.D
Acq On : 4 Nov 2024 20:05
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 05 01:47:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 01 05:13:40 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0