

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110424\
 Data File : BG063163.D
 Acq On : 4 Nov 2024 20:05
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 01:47:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	39.176	2.1	98	-0.01
3	Pyridine	40.000	44.953	-12.4	110	-0.01
4	n-Nitrosodimethylamine	40.000	39.159	2.1	89	-0.01
5 S	2-Fluorophenol	80.000	74.746	6.6	99	0.00
6	Aniline	40.000	38.581	3.5	97	0.00
7 S	Phenol-d6	80.000	78.322	2.1	99	0.00
8	2-Chlorophenol	40.000	40.212	-0.5	101	0.00
9	Benzaldehyde	40.000	43.128	-7.8	108	0.00
10 C	Phenol	40.000	38.594	3.5	99	0.00
11	bis(2-Chloroethyl)ether	40.000	39.260	1.9	100	0.00
12	1,3-Dichlorobenzene	40.000	37.117	7.2	95	0.00
13 C	1,4-Dichlorobenzene	40.000	38.637	3.4	99	0.00
14	1,2-Dichlorobenzene	40.000	37.636	5.9	97	0.00
15	Benzyl Alcohol	40.000	39.275	1.8	97	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.540	1.2	101	0.00
17	2-Methylphenol	40.000	38.536	3.7	99	0.00
18	Hexachloroethane	40.000	39.893	0.3	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.219	2.0	99	0.00
20	3+4-Methylphenols	40.000	39.906	0.2	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	39.816	0.5	101	0.00
23 S	Nitrobenzene-d5	80.000	77.511	3.1	101	0.00
24	Nitrobenzene	40.000	38.490	3.8	101	0.00
25	Isophorone	40.000	39.153	2.1	100	0.00
26 C	2-Nitrophenol	40.000	38.234	4.4	100	0.00
27	2,4-Dimethylphenol	40.000	41.020	-2.6	107	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.531	1.2	102	0.00
29 C	2,4-Dichlorophenol	40.000	37.354	6.6	96	0.00
30	1,2,4-Trichlorobenzene	40.000	37.722	5.7	97	0.00
31	Naphthalene	40.000	39.298	1.8	101	0.00
32	Benzoic acid	40.000	30.826	22.9	77	0.00
33	4-Chloroaniline	40.000	40.557	-1.4	102	0.00
34 C	Hexachlorobutadiene	40.000	37.771	5.6	97	0.00
35	Caprolactam	40.000	37.400	6.5	94	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.021	4.9	100	0.00
37	2-Methylnaphthalene	40.000	38.144	4.6	102	0.00
38	1-Methylnaphthalene	40.000	38.090	4.8	98	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.160	4.6	97	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.592	3.5	94	0.00
42 S	2,4,6-Tribromophenol	80.000	77.518	3.1	99	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.384	4.0	93	0.00
44	2,4,5-Trichlorophenol	40.000	39.554	1.1	97	0.00
45 S	2-Fluorobiphenyl	80.000	77.043	3.7	97	0.00
46	1,1'-Biphenyl	40.000	38.067	4.8	97	0.00
47	2-Chloronaphthalene	40.000	38.577	3.6	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110424\
 Data File : BG063163.D
 Acq On : 4 Nov 2024 20:05
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 01:47:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 01 05:13:40 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.717	0.7	96	0.00
49	Acenaphthylene	40.000	39.012	2.5	100	0.00
50	Dimethylphthalate	40.000	38.248	4.4	99	0.00
51	2,6-Dinitrotoluene	40.000	38.976	2.6	99	0.00
52 C	Acenaphthene	40.000	38.826	2.9	99	0.00
53	3-Nitroaniline	40.000	41.365	-3.4	103	0.00
54 P	2,4-Dinitrophenol	40.000	34.852	12.9	85	0.00
55	Dibenzofuran	40.000	39.288	1.8	101	0.00
56 P	4-Nitrophenol	40.000	44.014	-10.0	109	0.00
57	2,4-Dinitrotoluene	40.000	40.485	-1.2	102	0.00
58	Fluorene	40.000	38.459	3.9	97	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.318	6.7	94	0.00
60	Diethylphthalate	40.000	39.072	2.3	99	0.00
61	4-Chlorophenyl-phenylether	40.000	38.985	2.5	99	0.00
62	4-Nitroaniline	40.000	41.408	-3.5	106	0.00
63	Azobenzene	40.000	39.468	1.3	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.301	1.7	99	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.760	0.6	103	0.00
67	4-Bromophenyl-phenylether	40.000	39.761	0.6	101	0.00
68	Hexachlorobenzene	40.000	39.385	1.5	100	0.00
69	Atrazine	40.000	32.356	19.1	101	0.00
70 C	Pentachlorophenol	40.000	38.847	2.9	92	0.00
71	Phenanthrene	40.000	39.874	0.3	103	0.00
72	Anthracene	40.000	39.701	0.7	101	0.00
73	Carbazole	40.000	41.115	-2.8	108	0.00
74	Di-n-butylphthalate	40.000	41.202	-3.0	108	0.00
75 C	Fluoranthene	40.000	41.957	-4.9	113	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	120	0.00
77	Benzydine	40.000	61.108	-52.8#	215	0.00
78	Pyrene	40.000	39.166	2.1	114	0.00
79 S	Terphenyl-d14	80.000	77.133	3.6	112	0.00
80	Butylbenzylphthalate	40.000	41.203	-3.0	122	0.00
81	Benzo(a)anthracene	40.000	38.918	2.7	116	0.00
82	3,3'-Dichlorobenzidine	40.000	41.045	-2.6	124	-0.01
83	Chrysene	40.000	39.015	2.5	116	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.676	0.8	118	0.00
85 c	Di-n-octyl phthalate	40.000	40.445	-1.1	120	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	123	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	38.586	3.5	120	-0.02
88	Benzo(b)fluoranthene	40.000	37.725	5.7	118	-0.01
89	Benzo(k)fluoranthene	40.000	39.433	1.4	120	-0.01
90 C	Benzo(a)pyrene	40.000	38.142	4.6	120	-0.01
91	Dibenzo(a,h)anthracene	40.000	38.930	2.7	121	-0.02
92	Benzo(g,h,i)perylene	40.000	38.271	4.3	119	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110424\
Data File : BG063163.D
Acq On : 4 Nov 2024 20:05
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 05 01:47:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 01 05:13:40 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0