


Data Path : Z:\HPCHEM1\BNA G\Data\BG110515\  
Data File : BG019483.D  
Acq On : 5 Nov 2015 8:34  
Operator : UM/NP  
Sample : DFTPP16  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP16

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_G\METHODS\SOM02.2-EPA-BG102815.M  
Title : SVOA CALIBRATION  
Last Update : Fri Nov 06 02:11:58 2015



AutoFind: Scans 2573, 2574, 2575; Background Corrected with Scan 2567

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 29.1         | 21245      | PASS                |
| 68             | 69              | 0.00            | 2               | 0.7          | 224        | PASS                |
| 69             | 198             | 0.00            | 100             | 41.3         | 30066      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.4          | 114        | PASS                |
| 127            | 198             | 10              | 80              | 39.2         | 28584      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.4          | 268        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 72885      | PASS                |
| 199            | 198             | 5               | 9               | 6.8          | 4961       | PASS                |
| 275            | 198             | 10              | 60              | 29.8         | 21712      | PASS                |
| 365            | 198             | 1               | 100             | 4.4          | 3216       | PASS                |
| 441            | 442             | 0.01            | 24              | 17.0         | 7692       | PASS                |
| 442            | 198             | 50              | 100             | 61.9         | 45146      | PASS                |
| 443            | 442             | 15              | 24              | 19.9         | 8967       | PASS                |

Scan 2567 (17.921 min): BG019483.D  
DFTPP16

| m/z    | Abundance |
|--------|-----------|
| 43.90  | 526.0     |
| 61.30  | 150.0     |
| 69.90  | 177.0     |
| 192.90 | 198.0     |
| 206.90 | 299.0     |
| 376.90 | 242.0     |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP16

| m/z    | Abundance |
|--------|-----------|
| 37.90  | 247.0     |
| 39.00  | 1563.0    |
| 39.90  | 299.0     |
| 41.00  | 186.0     |
| 43.90  | 296.0     |
| 49.00  | 151.0     |
| 50.00  | 5962.0    |
| 51.00  | 21328.0   |
| 52.10  | 906.0     |
| 54.80  | 212.0     |
| 55.90  | 723.0     |
| 57.00  | 1581.0    |
| 61.00  | 670.0     |
| 61.90  | 716.0     |
| 63.00  | 1097.0    |
| 63.90  | 273.0     |
| 65.00  | 552.0     |
| 67.90  | 674.0     |
| 68.90  | 28328.0   |
| 70.10  | 250.0     |
| 72.90  | 401.0     |
| 74.00  | 3111.0    |
| 74.90  | 5752.0    |
| 76.00  | 1724.0    |
| 77.00  | 34408.0   |
| 78.00  | 2432.0    |
| 79.00  | 2533.0    |
| 79.90  | 1987.0    |
| 81.00  | 2281.0    |
| 81.90  | 733.0     |
| 82.90  | 372.0     |
| 84.10  | 167.0     |
| 85.00  | 341.0     |
| 85.90  | 821.0     |
| 87.00  | 349.0     |
| 90.90  | 802.0     |
| 92.00  | 592.0     |
| 92.90  | 4217.0    |
| 93.90  | 180.0     |
| 95.70  | 193.0     |
| 97.90  | 3119.0    |
| 98.90  | 2183.0    |
| 100.00 | 196.0     |
| 101.00 | 1427.0    |
| 103.00 | 728.0     |
| 103.90 | 891.0     |
| 105.00 | 1015.0    |
| 107.00 | 8987.0    |
| 108.00 | 1948.0    |
| 108.80 | 410.0     |
| 109.90 | 14276.0   |
| 110.90 | 2348.0    |
| 115.90 | 635.0     |
| 116.90 | 9584.0    |
| 117.90 | 779.0     |
| 121.90 | 456.0     |
| 122.90 | 1210.0    |
| 123.80 | 510.0     |
| 124.90 | 470.0     |
| 126.90 | 26424.0   |
| 127.90 | 3116.0    |
| 129.00 | 10512.0   |
| 129.90 | 796.0     |
| 131.10 | 178.0     |
| 134.00 | 459.0     |
| 134.90 | 570.0     |
| 135.90 | 413.0     |
| 137.00 | 569.0     |
| 140.00 | 151.0     |
| 140.90 | 1928.0    |
| 141.90 | 310.0     |
| 143.00 | 366.0     |
| 145.90 | 445.0     |
| 146.90 | 1073.0    |
| 147.90 | 2200.0    |
| 148.90 | 309.0     |
| 150.00 | 201.0     |
| 151.10 | 184.0     |
| 151.30 | 190.0     |
| 151.80 | 159.0     |
| 153.00 | 677.0     |
| 153.90 | 468.0     |
| 155.00 | 1294.0    |
| 155.90 | 1183.0    |
| 156.90 | 273.0     |
| 158.00 | 367.0     |
| 159.00 | 296.0     |
| 159.90 | 454.0     |
| 161.10 | 626.0     |
| 162.00 | 312.0     |
| 162.90 | 318.0     |
| 164.90 | 966.0     |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP16

|        |         |
|--------|---------|
| 165.90 | 419.0   |
| 166.90 | 2717.0  |
| 168.00 | 2061.0  |
| 168.90 | 281.0   |
| 171.80 | 465.0   |
| 172.80 | 577.0   |
| 173.90 | 536.0   |
| 174.90 | 1589.0  |
| 175.60 | 238.0   |
| 176.90 | 813.0   |
| 177.90 | 376.0   |
| 178.90 | 2869.0  |
| 179.90 | 1701.0  |
| 180.90 | 502.0   |
| 181.90 | 317.0   |
| 183.90 | 194.0   |
| 184.80 | 1182.0  |
| 186.00 | 10006.0 |
| 187.00 | 2612.0  |
| 188.00 | 256.0   |
| 189.00 | 775.0   |
| 191.00 | 719.0   |
| 191.80 | 911.0   |
| 193.00 | 711.0   |
| 194.00 | 295.0   |
| 195.90 | 1808.0  |
| 197.90 | 65208.0 |
| 198.90 | 4312.0  |
| 199.90 | 408.0   |
| 200.90 | 277.0   |
| 203.00 | 842.0   |
| 203.90 | 3369.0  |
| 204.90 | 5163.0  |
| 206.00 | 21624.0 |
| 207.00 | 3466.0  |
| 208.00 | 1014.0  |
| 209.00 | 252.0   |
| 209.80 | 428.0   |
| 210.90 | 1005.0  |
| 211.80 | 256.0   |
| 214.90 | 303.0   |
| 216.90 | 6168.0  |
| 217.90 | 530.0   |
| 220.90 | 1973.0  |
| 221.90 | 1024.0  |
| 222.90 | 1404.0  |
| 223.90 | 11388.0 |
| 225.00 | 2675.0  |
| 226.00 | 498.0   |
| 226.90 | 6442.0  |
| 227.90 | 791.0   |
| 228.90 | 1081.0  |
| 229.90 | 161.0   |
| 230.90 | 416.0   |
| 234.00 | 561.0   |
| 235.00 | 354.0   |
| 235.80 | 378.0   |
| 236.60 | 250.0   |
| 238.90 | 322.0   |
| 239.90 | 360.0   |
| 241.00 | 304.0   |
| 241.90 | 613.0   |
| 242.90 | 515.0   |
| 244.00 | 7758.0  |
| 245.00 | 1179.0  |
| 245.90 | 2555.0  |
| 246.90 | 356.0   |
| 248.90 | 379.0   |
| 252.50 | 174.0   |
| 253.10 | 365.0   |
| 254.90 | 44736.0 |
| 255.90 | 6837.0  |
| 257.00 | 360.0   |
| 257.90 | 3048.0  |
| 259.00 | 495.0   |
| 260.20 | 268.0   |
| 264.90 | 1186.0  |
| 266.00 | 339.0   |
| 272.90 | 1646.0  |
| 274.00 | 2936.0  |
| 275.00 | 17880.0 |
| 276.00 | 2040.0  |
| 276.90 | 1494.0  |
| 277.90 | 369.0   |
| 285.10 | 406.0   |
| 291.60 | 172.0   |
| 292.90 | 524.0   |
| 295.90 | 5827.0  |
| 296.90 | 799.0   |
| 302.90 | 548.0   |
| 303.80 | 303.0   |
| 309.90 | 156.0   |
| 314.10 | 299.0   |
| 315.10 | 727.0   |
| 315.90 | 248.0   |
| 321.10 | 377.0   |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP16

|        |         |
|--------|---------|
| 323.00 | 1625.0  |
| 324.00 | 471.0   |
| 327.00 | 527.0   |
| 332.00 | 220.0   |
| 333.00 | 519.0   |
| 333.90 | 1081.0  |
| 335.00 | 383.0   |
| 352.00 | 263.0   |
| 353.00 | 347.0   |
| 354.00 | 325.0   |
| 355.00 | 190.0   |
| 364.90 | 2337.0  |
| 365.60 | 276.0   |
| 371.90 | 896.0   |
| 401.90 | 153.0   |
| 403.00 | 245.0   |
| 404.00 | 198.0   |
| 420.90 | 378.0   |
| 421.90 | 243.0   |
| 423.00 | 1673.0  |
| 424.00 | 471.0   |
| 441.00 | 5637.0  |
| 442.00 | 31520.0 |
| 443.00 | 6305.0  |
| 444.00 | 408.0   |
| 481.10 | 178.0   |

**Instrument :**  
BNA\_G  
**ClientSampleId :**  
DFTPP16

| m/z    | Abundance |
|--------|-----------|
| 38.00  | 325.0     |
| 39.00  | 1652.0    |
| 40.00  | 219.0     |
| 43.90  | 529.0     |
| 50.00  | 6900.0    |
| 51.00  | 24696.0   |
| 52.00  | 1587.0    |
| 56.00  | 721.0     |
| 57.00  | 2147.0    |
| 58.10  | 186.0     |
| 60.80  | 385.0     |
| 61.90  | 472.0     |
| 63.00  | 1385.0    |
| 64.90  | 734.0     |
| 65.50  | 200.0     |
| 68.90  | 35464.0   |
| 70.00  | 294.0     |
| 72.90  | 452.0     |
| 74.00  | 4257.0    |
| 75.00  | 6926.0    |
| 76.00  | 1895.0    |
| 77.00  | 38832.0   |
| 77.90  | 2997.0    |
| 78.90  | 3399.0    |
| 79.90  | 2222.0    |
| 80.90  | 3355.0    |
| 81.90  | 743.0     |
| 82.90  | 901.0     |
| 83.90  | 214.0     |
| 84.90  | 504.0     |
| 86.00  | 1123.0    |
| 87.10  | 538.0     |
| 90.90  | 583.0     |
| 92.00  | 774.0     |
| 92.90  | 5035.0    |
| 93.90  | 411.0     |
| 95.10  | 197.0     |
| 96.00  | 289.0     |
| 97.10  | 161.0     |
| 98.00  | 3372.0    |
| 98.90  | 3273.0    |
| 100.10 | 248.0     |
| 100.90 | 1681.0    |
| 102.90 | 477.0     |
| 104.00 | 847.0     |
| 104.90 | 852.0     |
| 105.90 | 310.0     |
| 106.90 | 11829.0   |
| 108.00 | 2104.0    |
| 108.80 | 421.0     |
| 109.90 | 19480.0   |
| 111.00 | 2421.0    |
| 111.80 | 295.0     |
| 115.90 | 1045.0    |
| 116.90 | 10829.0   |
| 117.90 | 1042.0    |
| 121.00 | 233.0     |
| 121.80 | 705.0     |
| 122.90 | 1378.0    |
| 123.80 | 392.0     |
| 125.00 | 482.0     |
| 127.00 | 32976.0   |
| 127.90 | 2310.0    |
| 128.90 | 14019.0   |
| 129.90 | 1297.0    |
| 131.00 | 266.0     |
| 131.90 | 165.0     |
| 133.90 | 376.0     |
| 135.00 | 1019.0    |
| 136.00 | 607.0     |
| 137.00 | 602.0     |
| 139.90 | 424.0     |
| 141.00 | 2110.0    |
| 141.90 | 756.0     |
| 143.00 | 712.0     |
| 144.10 | 180.0     |
| 146.00 | 271.0     |
| 146.90 | 1365.0    |
| 147.90 | 2013.0    |
| 148.90 | 836.0     |
| 151.00 | 584.0     |
| 151.90 | 406.0     |
| 153.00 | 711.0     |
| 154.00 | 609.0     |
| 155.00 | 996.0     |
| 155.90 | 2164.0    |
| 157.00 | 389.0     |
| 157.70 | 391.0     |
| 158.80 | 190.0     |
| 159.90 | 1182.0    |
| 160.90 | 1344.0    |
| 161.90 | 248.0     |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP16

|        |         |
|--------|---------|
| 164.90 | 934.0   |
| 165.90 | 989.0   |
| 166.90 | 4152.0  |
| 167.90 | 2233.0  |
| 168.90 | 322.0   |
| 170.00 | 169.0   |
| 170.90 | 251.0   |
| 172.00 | 640.0   |
| 173.00 | 413.0   |
| 173.90 | 1050.0  |
| 174.90 | 1714.0  |
| 175.90 | 368.0   |
| 176.70 | 549.0   |
| 177.90 | 288.0   |
| 179.00 | 3440.0  |
| 180.00 | 3015.0  |
| 180.90 | 1187.0  |
| 182.90 | 213.0   |
| 183.90 | 345.0   |
| 184.90 | 1603.0  |
| 186.00 | 13309.0 |
| 187.00 | 4462.0  |
| 187.90 | 379.0   |
| 188.90 | 859.0   |
| 190.00 | 163.0   |
| 190.80 | 438.0   |
| 192.00 | 1703.0  |
| 192.90 | 1774.0  |
| 194.00 | 516.0   |
| 194.90 | 208.0   |
| 195.90 | 2743.0  |
| 197.90 | 84184.0 |
| 198.80 | 6208.0  |
| 199.90 | 574.0   |
| 201.20 | 355.0   |
| 203.00 | 1615.0  |
| 204.00 | 4392.0  |
| 205.00 | 7148.0  |
| 206.00 | 28624.0 |
| 207.00 | 4460.0  |
| 207.90 | 776.0   |
| 208.90 | 460.0   |
| 209.90 | 279.0   |
| 210.90 | 1109.0  |
| 214.90 | 562.0   |
| 215.90 | 423.0   |
| 216.90 | 8841.0  |
| 217.90 | 963.0   |
| 218.80 | 193.0   |
| 220.90 | 2633.0  |
| 221.90 | 1030.0  |
| 222.90 | 1854.0  |
| 223.90 | 16448.0 |
| 224.90 | 3818.0  |
| 226.00 | 429.0   |
| 227.00 | 7604.0  |
| 227.90 | 933.0   |
| 228.90 | 1733.0  |
| 229.90 | 193.0   |
| 230.90 | 440.0   |
| 232.90 | 310.0   |
| 234.00 | 419.0   |
| 234.80 | 630.0   |
| 235.80 | 234.0   |
| 236.90 | 416.0   |
| 237.90 | 182.0   |
| 238.90 | 316.0   |
| 239.90 | 263.0   |
| 240.90 | 604.0   |
| 241.90 | 783.0   |
| 243.00 | 1182.0  |
| 244.00 | 12085.0 |
| 245.00 | 1333.0  |
| 245.90 | 2713.0  |
| 246.90 | 295.0   |
| 247.90 | 319.0   |
| 249.00 | 456.0   |
| 251.10 | 160.0   |
| 252.90 | 638.0   |
| 254.90 | 67920.0 |
| 256.00 | 10595.0 |
| 256.90 | 833.0   |
| 257.90 | 4866.0  |
| 258.90 | 689.0   |
| 260.90 | 195.0   |
| 264.90 | 2003.0  |
| 265.90 | 286.0   |
| 269.60 | 203.0   |
| 270.90 | 313.0   |
| 272.00 | 300.0   |
| 272.90 | 1723.0  |
| 273.90 | 5167.0  |
| 275.00 | 23792.0 |
| 275.90 | 2699.0  |
| 276.90 | 2563.0  |
| 277.90 | 397.0   |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP16

|        |         |
|--------|---------|
| 280.90 | 196.0   |
| 282.80 | 249.0   |
| 283.80 | 194.0   |
| 284.90 | 462.0   |
| 289.00 | 230.0   |
| 291.00 | 329.0   |
| 292.00 | 200.0   |
| 293.00 | 468.0   |
| 293.80 | 221.0   |
| 294.80 | 244.0   |
| 296.00 | 9030.0  |
| 296.90 | 959.0   |
| 298.10 | 228.0   |
| 301.00 | 276.0   |
| 301.80 | 438.0   |
| 302.90 | 1321.0  |
| 304.00 | 249.0   |
| 308.70 | 196.0   |
| 313.90 | 387.0   |
| 314.90 | 1094.0  |
| 315.90 | 609.0   |
| 321.00 | 324.0   |
| 321.90 | 307.0   |
| 323.00 | 1927.0  |
| 324.00 | 239.0   |
| 326.90 | 515.0   |
| 327.90 | 207.0   |
| 328.90 | 210.0   |
| 331.80 | 347.0   |
| 332.90 | 343.0   |
| 333.90 | 1249.0  |
| 335.10 | 239.0   |
| 341.00 | 273.0   |
| 345.90 | 514.0   |
| 352.00 | 355.0   |
| 352.90 | 588.0   |
| 354.00 | 827.0   |
| 364.90 | 3680.0  |
| 366.00 | 502.0   |
| 372.00 | 919.0   |
| 382.70 | 268.0   |
| 383.10 | 291.0   |
| 390.00 | 338.0   |
| 402.00 | 746.0   |
| 403.00 | 839.0   |
| 403.90 | 292.0   |
| 420.90 | 578.0   |
| 422.00 | 589.0   |
| 422.90 | 2973.0  |
| 423.90 | 844.0   |
| 441.00 | 8184.0  |
| 442.00 | 52536.0 |
| 443.00 | 9756.0  |
| 444.00 | 923.0   |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP16



| m/z    | Abundance |
|--------|-----------|
| 37.90  | 219.0     |
| 39.00  | 1486.0    |
| 40.00  | 371.0     |
| 43.90  | 514.0     |
| 48.90  | 220.0     |
| 50.00  | 4432.0    |
| 51.00  | 17712.0   |
| 52.00  | 850.0     |
| 55.90  | 831.0     |
| 57.00  | 1078.0    |
| 61.00  | 287.0     |
| 61.90  | 427.0     |
| 63.00  | 1034.0    |
| 64.00  | 211.0     |
| 64.90  | 408.0     |
| 68.10  | 644.0     |
| 68.90  | 26408.0   |
| 69.90  | 331.0     |
| 72.90  | 285.0     |
| 74.00  | 3036.0    |
| 74.90  | 5089.0    |
| 76.10  | 1587.0    |
| 77.00  | 26904.0   |
| 78.00  | 1710.0    |
| 78.90  | 2665.0    |
| 79.90  | 1089.0    |
| 80.90  | 2317.0    |
| 81.90  | 1213.0    |
| 83.00  | 223.0     |
| 85.00  | 471.0     |
| 86.00  | 591.0     |
| 87.00  | 184.0     |
| 87.80  | 264.0     |
| 90.90  | 575.0     |
| 92.00  | 499.0     |
| 93.00  | 3738.0    |
| 94.00  | 249.0     |
| 95.00  | 169.0     |
| 96.00  | 190.0     |
| 97.10  | 170.0     |
| 97.90  | 2947.0    |
| 98.90  | 1926.0    |
| 100.90 | 1119.0    |
| 101.80 | 179.0     |
| 102.80 | 310.0     |
| 103.90 | 1029.0    |
| 104.80 | 776.0     |
| 106.00 | 458.0     |
| 106.90 | 9337.0    |
| 108.00 | 1091.0    |
| 108.90 | 369.0     |
| 109.90 | 13881.0   |
| 110.90 | 2108.0    |
| 112.10 | 176.0     |
| 115.90 | 606.0     |
| 117.00 | 6910.0    |
| 117.90 | 534.0     |
| 122.00 | 577.0     |
| 123.00 | 888.0     |
| 123.70 | 481.0     |
| 125.00 | 504.0     |
| 126.90 | 26352.0   |
| 127.90 | 2482.0    |
| 128.90 | 11922.0   |
| 130.00 | 912.0     |
| 130.90 | 401.0     |
| 132.10 | 187.0     |
| 133.90 | 166.0     |
| 135.00 | 1214.0    |
| 135.90 | 453.0     |
| 137.00 | 815.0     |
| 140.90 | 1831.0    |
| 141.90 | 581.0     |
| 142.90 | 363.0     |
| 146.10 | 298.0     |
| 146.80 | 1032.0    |
| 147.90 | 2332.0    |
| 149.00 | 426.0     |
| 151.90 | 338.0     |
| 153.00 | 777.0     |
| 154.00 | 686.0     |
| 155.00 | 1521.0    |
| 155.90 | 1219.0    |
| 156.90 | 298.0     |
| 157.80 | 646.0     |
| 158.80 | 404.0     |
| 159.80 | 594.0     |
| 160.90 | 1102.0    |
| 161.70 | 404.0     |
| 163.00 | 229.0     |
| 165.00 | 998.0     |
| 165.90 | 567.0     |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP16

|        |         |
|--------|---------|
| 166.90 | 3238.0  |
| 167.90 | 2300.0  |
| 168.90 | 491.0   |
| 172.00 | 472.0   |
| 172.90 | 574.0   |
| 173.90 | 721.0   |
| 175.00 | 1150.0  |
| 176.00 | 297.0   |
| 176.90 | 698.0   |
| 178.00 | 212.0   |
| 178.90 | 3461.0  |
| 180.00 | 2623.0  |
| 180.90 | 829.0   |
| 183.80 | 611.0   |
| 184.90 | 832.0   |
| 186.00 | 10603.0 |
| 186.90 | 2951.0  |
| 188.10 | 311.0   |
| 188.90 | 1146.0  |
| 190.90 | 240.0   |
| 192.00 | 1058.0  |
| 192.90 | 1152.0  |
| 194.10 | 183.0   |
| 195.90 | 1947.0  |
| 196.90 | 805.0   |
| 197.90 | 69264.0 |
| 198.90 | 4364.0  |
| 199.90 | 437.0   |
| 201.30 | 293.0   |
| 202.90 | 783.0   |
| 203.90 | 3140.0  |
| 205.00 | 5434.0  |
| 206.00 | 22624.0 |
| 207.00 | 3123.0  |
| 207.90 | 890.0   |
| 209.00 | 192.0   |
| 209.90 | 289.0   |
| 210.90 | 1222.0  |
| 211.60 | 206.0   |
| 215.00 | 250.0   |
| 215.90 | 312.0   |
| 216.90 | 8012.0  |
| 218.00 | 815.0   |
| 219.30 | 239.0   |
| 220.00 | 288.0   |
| 220.90 | 1781.0  |
| 221.60 | 1206.0  |
| 222.90 | 1458.0  |
| 224.00 | 14181.0 |
| 224.90 | 3799.0  |
| 226.90 | 6510.0  |
| 227.90 | 992.0   |
| 228.90 | 1393.0  |
| 229.70 | 182.0   |
| 230.90 | 801.0   |
| 231.80 | 157.0   |
| 233.90 | 555.0   |
| 235.10 | 576.0   |
| 236.00 | 272.0   |
| 236.90 | 535.0   |
| 238.90 | 319.0   |
| 239.90 | 275.0   |
| 240.90 | 400.0   |
| 241.90 | 953.0   |
| 242.90 | 945.0   |
| 244.00 | 9210.0  |
| 245.00 | 1534.0  |
| 246.00 | 2754.0  |
| 246.90 | 305.0   |
| 249.10 | 215.0   |
| 250.90 | 174.0   |
| 252.10 | 159.0   |
| 252.90 | 247.0   |
| 254.90 | 55768.0 |
| 255.90 | 7759.0  |
| 257.00 | 744.0   |
| 257.90 | 3267.0  |
| 258.90 | 651.0   |
| 264.90 | 1613.0  |
| 265.90 | 194.0   |
| 270.90 | 229.0   |
| 272.90 | 1326.0  |
| 273.90 | 4192.0  |
| 274.90 | 23464.0 |
| 275.90 | 2687.0  |
| 277.00 | 2973.0  |
| 277.90 | 429.0   |
| 278.90 | 182.0   |
| 282.90 | 349.0   |
| 284.20 | 223.0   |
| 284.50 | 204.0   |
| 285.00 | 418.0   |
| 291.10 | 181.0   |
| 291.90 | 173.0   |
| 293.00 | 426.0   |
| 295.90 | 7449.0  |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP16

|        |         |
|--------|---------|
| 296.80 | 1129.0  |
| 297.90 | 172.0   |
| 302.00 | 272.0   |
| 302.90 | 818.0   |
| 303.90 | 273.0   |
| 308.00 | 255.0   |
| 310.00 | 181.0   |
| 313.70 | 184.0   |
| 315.00 | 911.0   |
| 315.90 | 377.0   |
| 320.90 | 399.0   |
| 323.00 | 2486.0  |
| 324.00 | 252.0   |
| 325.00 | 181.0   |
| 326.70 | 315.0   |
| 328.00 | 279.0   |
| 332.10 | 227.0   |
| 334.00 | 1826.0  |
| 334.90 | 358.0   |
| 341.00 | 398.0   |
| 345.90 | 611.0   |
| 352.00 | 526.0   |
| 352.90 | 423.0   |
| 353.90 | 508.0   |
| 364.90 | 3632.0  |
| 366.10 | 326.0   |
| 370.00 | 230.0   |
| 371.00 | 159.0   |
| 371.90 | 647.0   |
| 372.90 | 293.0   |
| 382.90 | 447.0   |
| 402.00 | 208.0   |
| 403.00 | 431.0   |
| 420.90 | 609.0   |
| 421.90 | 474.0   |
| 422.90 | 3033.0  |
| 423.90 | 323.0   |
| 441.00 | 9255.0  |
| 442.00 | 51384.0 |
| 443.00 | 10840.0 |
| 443.90 | 975.0   |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP16