

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110520\
 Data File : BG047217.D
 Acq On : 5 Nov 2020 9:23
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 10:19:45 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 04 16:19:40 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.611	0.602	1.5	100	0.00
3	Pyridine	1.634	1.676	-2.6	96	0.00
4	n-Nitrosodimethylamine	0.820	0.860	-4.9	99	0.00
5 S	2-Fluorophenol	1.234	1.222	1.0	96	0.00
6	Aniline	2.108	2.183	-3.6	98	0.00
7 S	Phenol-d6	1.786	1.883	-5.4	100	0.00
8	2-Chlorophenol	1.310	1.337	-2.1	99	0.00
9	Benzaldehyde	1.202	1.240	-3.2	100	0.00
10 C	Phenol	1.694	1.729	-2.1	97	0.00
11	bis(2-Chloroethyl)ether	1.344	1.383	-2.9	99	0.00
12	1,3-Dichlorobenzene	1.705	1.733	-1.6	99	0.00
13 C	1,4-Dichlorobenzene	1.712	1.743	-1.8	97	0.00
14	1,2-Dichlorobenzene	1.624	1.629	-0.3	100	0.00
15	Benzyl Alcohol	1.373	1.502	-9.4	100	0.00
16	2,2'-oxybis(1-Chloropropane	2.074	2.122	-2.3	99	0.00
17	2-Methylphenol	1.147	1.156	-0.8	96	0.00
18	Hexachloroethane	0.627	0.636	-1.4	99	0.00
19 P	n-Nitroso-di-n-propylamine	1.174	1.236	-5.3	98	0.00
20	3+4-Methylphenols	1.536	1.642	-6.9	100	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
22	Acetophenone	0.577	0.599	-3.8	98	0.00
23 S	Nitrobenzene-d5	0.466	0.494	-6.0	98	0.00
24	Nitrobenzene	0.474	0.491	-3.6	98	0.00
25	Isophorone	0.799	0.843	-5.5	97	0.00
26 C	2-Nitrophenol	0.205	0.215	-4.9	96	0.00
27	2,4-Dimethylphenol	0.275	0.292	-6.2	98	0.00
28	bis(2-Chloroethoxy)methane	0.483	0.517	-7.0	100	0.00
29 C	2,4-Dichlorophenol	0.379	0.407	-7.4	98	0.00
30	1,2,4-Trichlorobenzene	0.468	0.491	-4.9	101	0.00
31	Naphthalene	1.145	1.194	-4.3	100	0.00
32	Benzoic acid	0.191	0.244	-27.7#	96	0.00
33	4-Chloroaniline	0.482	0.514	-6.6	97	0.00
34 C	Hexachlorobutadiene	0.354	0.363	-2.5	101	0.00
35	Caprolactam	0.124	0.137	-10.5	94	0.00
36 C	4-Chloro-3-methylphenol	0.387	0.419	-8.3	97	0.00
37	2-Methylnaphthalene	0.861	0.914	-6.2	99	0.00
38	1-Methylnaphthalene	0.818	0.861	-5.3	98	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	0.790	0.815	-3.2	100	0.00
41 P	Hexachlorocyclopentadiene	0.409	0.435	-6.4	98	0.00
42 S	2,4,6-Tribromophenol	0.335	0.357	-6.6	96	0.00
43 C	2,4,6-Trichlorophenol	0.513	0.548	-6.8	100	0.00
44	2,4,5-Trichlorophenol	0.521	0.538	-3.3	94	0.00
45 S	2-Fluorobiphenyl	1.559	1.610	-3.3	99	0.00
46	1,1'-Biphenyl	1.637	1.683	-2.8	98	0.00
47	2-Chloronaphthalene	1.336	1.373	-2.8	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110520\
 Data File : BG047217.D
 Acq On : 5 Nov 2020 9:23
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 10:19:45 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 04 16:19:40 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.443	0.484	-9.3	99	0.00
49	Acenaphthylene	1.954	2.035	-4.1	97	0.00
50	Dimethylphthalate	1.766	1.826	-3.4	96	0.00
51	2,6-Dinitrotoluene	0.366	0.390	-6.6	98	0.00
52 C	Acenaphthene	1.249	1.319	-5.6	99	0.00
53	3-Nitroaniline	0.365	0.402	-10.1	100	0.00
54 P	2,4-Dinitrophenol	0.235	0.256	-8.9	99	0.00
55	Dibenzofuran	2.053	2.128	-3.7	97	0.00
56 P	4-Nitrophenol	0.282	0.312	-10.6	95	0.00
57	2,4-Dinitrotoluene	0.535	0.580	-8.4	97	0.00
58	Fluorene	1.650	1.736	-5.2	97	0.00
59	2,3,4,6-Tetrachlorophenol	0.547	0.593	-8.4	98	0.00
60	Diethylphthalate	1.745	1.845	-5.7	97	0.00
61	4-Chlorophenyl-phenylether	0.977	1.014	-3.8	96	0.00
62	4-Nitroaniline	0.392	0.422	-7.7	97	0.00
63	Azobenzene	1.564	1.636	-4.6	97	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	0.139	0.142	-2.2	96	0.00
66 c	n-Nitrosodiphenylamine	0.615	0.633	-2.9	98	0.00
67	4-Bromophenyl-phenylether	0.279	0.287	-2.9	98	0.00
68	Hexachlorobenzene	0.304	0.308	-1.3	96	0.00
69	Atrazine	0.263	0.271	-3.0	98	0.00
70 C	Pentachlorophenol	0.165	0.180	-9.1	98	0.00
71	Phenanthrene	1.161	1.201	-3.4	98	0.00
72	Anthracene	1.127	1.164	-3.3	98	0.00
73	Carbazole	0.995	1.036	-4.1	99	0.00
74	Di-n-butylphthalate	1.217	1.271	-4.4	98	0.00
75 C	Fluoranthene	1.478	1.513	-2.4	98	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
77	Benzydine	0.671	0.711	-6.0	101	0.00
78	Pyrene	1.393	1.428	-2.5	98	0.00
79 S	Terphenyl-d14	1.137	1.158	-1.8	98	0.00
80	Butylbenzylphthalate	0.525	0.545	-3.8	98	0.00
81	Benzo(a)anthracene	1.418	1.446	-2.0	100	0.00
82	3,3'-Dichlorobenzidine	0.491	0.511	-4.1	101	0.00
83	Chrysene	1.365	1.390	-1.8	101	0.00
84	Bis(2-ethylhexyl)phthalate	0.735	0.767	-4.4	100	0.00
85 c	Di-n-octyl phthalate	1.235	1.294	-4.8	100	0.00
86 I	Perylene-d12	1.000	1.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	1.518	1.567	-3.2	101	0.00
88	Benzo(b)fluoranthene	1.365	1.398	-2.4	100	0.00
89	Benzo(k)fluoranthene	1.334	1.370	-2.7	101	0.00
90 C	Benzo(a)pyrene	1.274	1.314	-3.1	101	0.00
91	Dibenzo(a,h)anthracene	1.236	1.269	-2.7	100	0.00
92	Benzo(g,h,i)perylene	1.225	1.250	-2.0	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110520\
Data File : BG047217.D
Acq On : 5 Nov 2020 9:23
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 05 10:19:45 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG110420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 04 16:19:40 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0