


Data Path : Z:\HPCHEM1\BNA G\DATA\BG110615\
Data File : BG019501.D
Acq On : 6 Nov 2015 7:47
Operator : UM/NP
Sample : DFTPP17
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP17

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG102815.M
Title : SVOA CALIBRATION
Last Update : Fri Nov 06 23:58:14 2015



AutoFind: Scans 2573, 2574, 2575; Background Corrected with Scan 2567

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.8	21990	PASS
68	69	0.00	2	1.7	540	PASS
69	198	0.00	100	40.7	31072	PASS
70	69	0.00	2	0.7	216	PASS
127	198	10	80	38.5	29389	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	76298	PASS
199	198	5	9	6.9	5286	PASS
275	198	10	60	29.3	22338	PASS
365	198	1	100	4.3	3302	PASS
441	442	0.01	24	17.1	8241	PASS
442	198	50	100	63.1	48160	PASS
443	442	15	24	20.1	9702	PASS

Scan 2567 (17.922 min): BG019501.D
DFTPP17

m/z	Abundance
40.00	250.0
43.90	539.0
50.10	155.0
89.10	158.0
206.80	263.0
207.80	169.0
328.00	195.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

m/z	Abundance
38.00	287.0
39.00	1371.0
39.90	330.0
41.00	212.0
44.00	637.0
45.00	210.0
50.00	7252.0
51.00	25608.0
52.00	1432.0
55.00	249.0
55.90	1037.0
57.00	2008.0
57.90	183.0
61.10	306.0
61.90	373.0
62.90	1040.0
63.80	250.0
64.90	710.0
68.00	878.0
68.90	35184.0
69.90	384.0
72.90	383.0
74.00	5082.0
75.00	6175.0
76.00	2389.0
77.00	39544.0
78.00	2845.0
78.90	4378.0
79.90	2134.0
81.00	3501.0
81.90	641.0
83.00	801.0
83.90	155.0
85.00	732.0
85.90	746.0
86.80	482.0
87.90	175.0
88.80	193.0
90.90	646.0
91.90	886.0
92.90	5105.0
93.70	233.0
94.00	154.0
97.90	3439.0
98.90	2694.0
100.00	302.0
100.90	1262.0
103.00	485.0
104.00	894.0
105.00	799.0
106.00	301.0
107.00	11336.0
107.90	1931.0
108.90	427.0
110.00	17424.0
110.90	2905.0
111.90	315.0
112.90	397.0
115.90	615.0
117.00	9190.0
118.00	991.0
119.80	160.0
122.00	1031.0
123.00	1081.0
123.80	417.0
124.90	580.0
126.90	31416.0
127.90	2492.0
128.90	14272.0
129.90	1201.0
133.90	718.0
135.00	1098.0
136.00	516.0
136.90	373.0
137.50	244.0
139.70	214.0
141.00	2334.0
141.80	612.0
142.90	279.0
145.90	326.0
147.00	948.0
147.90	2849.0
148.90	619.0
151.00	360.0
152.90	416.0
154.00	335.0
155.00	1072.0
155.90	1452.0
157.00	353.0
157.90	521.0
159.00	301.0
159.70	711.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

160.90	714.0
162.00	418.0
164.90	660.0
165.90	605.0
166.90	4201.0
167.90	2101.0
168.90	406.0
169.70	267.0
170.80	211.0
172.00	610.0
172.80	494.0
174.00	681.0
175.00	2185.0
175.90	296.0
176.90	933.0
177.90	441.0
178.90	3685.0
180.00	2516.0
180.90	1596.0
181.90	188.0
183.80	386.0
184.20	283.0
184.90	1307.0
185.90	13206.0
186.90	3423.0
187.90	383.0
188.90	846.0
189.90	331.0
191.00	479.0
192.00	1173.0
193.00	1294.0
193.80	290.0
194.90	266.0
196.00	3533.0
197.90	81816.0
198.90	5025.0
199.90	417.0
201.10	256.0
202.00	212.0
202.90	1045.0
204.00	4311.0
205.00	6797.0
205.90	26728.0
207.00	4024.0
207.90	1322.0
209.00	544.0
209.90	276.0
210.90	1214.0
212.00	243.0
214.80	280.0
215.10	241.0
216.00	547.0
216.90	7675.0
217.90	854.0
218.80	175.0
219.90	182.0
220.90	2909.0
221.60	981.0
222.90	1605.0
223.90	15904.0
225.00	3685.0
226.00	496.0
227.00	7205.0
227.90	1138.0
228.80	1781.0
229.90	236.0
230.90	664.0
232.90	226.0
233.80	239.0
234.80	363.0
235.90	511.0
236.90	490.0
238.90	447.0
239.80	288.0
241.00	515.0
241.90	540.0
243.00	977.0
244.00	10599.0
244.90	1351.0
245.90	2463.0
246.80	494.0
247.90	217.0
248.80	513.0
250.90	153.0
251.90	373.0
253.00	306.0
254.90	56432.0
255.90	10654.0
256.80	473.0
257.90	3791.0
258.90	358.0
264.90	1510.0
265.60	189.0
266.00	252.0
269.70	331.0
272.90	1993.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

273.90	4779.0
274.90	20920.0
275.90	3313.0
277.00	2446.0
278.00	545.0
283.00	269.0
284.90	305.0
288.40	176.0
289.80	226.0
293.00	488.0
295.90	7306.0
296.90	1043.0
297.90	177.0
301.90	199.0
303.00	881.0
304.00	373.0
314.10	487.0
314.80	781.0
315.80	330.0
320.90	252.0
323.00	1573.0
324.00	467.0
326.80	242.0
327.90	264.0
332.80	359.0
334.00	1576.0
334.90	575.0
341.00	188.0
342.10	166.0
352.00	556.0
353.00	536.0
353.90	762.0
365.00	2576.0
365.90	642.0
372.00	559.0
372.90	322.0
402.00	318.0
403.00	823.0
403.90	305.0
421.00	459.0
422.10	374.0
423.00	2239.0
424.00	693.0
441.00	7083.0
442.00	40760.0
443.00	8783.0
443.90	643.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

m/z	Abundance
37.00	253.0
37.90	271.0
39.00	1557.0
39.90	272.0
43.90	431.0
48.90	367.0
50.00	6844.0
51.00	25048.0
52.00	1076.0
55.10	224.0
55.90	910.0
57.00	2808.0
60.90	400.0
61.90	567.0
63.00	1624.0
63.90	281.0
64.90	693.0
67.90	513.0
68.90	35488.0
72.90	593.0
74.00	3949.0
75.00	6518.0
76.00	1936.0
77.00	40336.0
77.90	2326.0
78.90	2238.0
79.90	2379.0
80.90	3754.0
81.90	736.0
83.00	498.0
84.90	547.0
85.90	869.0
86.90	609.0
89.00	160.0
91.00	517.0
91.90	724.0
92.90	4135.0
93.90	294.0
95.80	161.0
98.00	3170.0
98.90	2004.0
100.00	517.0
100.90	1834.0
102.80	465.0
103.90	1367.0
105.00	625.0
106.00	290.0
106.90	12899.0
108.00	2509.0
109.00	524.0
109.90	20216.0
110.90	3514.0
112.90	183.0
113.80	255.0
115.90	812.0
116.90	11454.0
117.80	1082.0
118.80	235.0
121.00	172.0
122.00	906.0
122.80	1531.0
123.90	630.0
124.90	795.0
127.00	34312.0
128.00	2124.0
128.90	15256.0
129.90	1224.0
131.00	218.0
133.90	617.0
134.90	1018.0
136.00	495.0
136.90	607.0
137.60	151.0
138.00	226.0
140.90	2877.0
142.00	751.0
143.00	422.0
144.10	177.0
144.80	151.0
145.90	406.0
146.90	1310.0
147.90	2387.0
149.00	566.0
150.90	394.0
152.90	1208.0
153.90	536.0
154.90	1354.0
155.90	1985.0
156.90	338.0
158.00	508.0
158.90	438.0
159.90	883.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

160.90	959.0
161.80	310.0
162.90	213.0
163.90	164.0
165.00	1181.0
165.90	1199.0
166.90	5279.0
168.00	2041.0
168.90	656.0
171.10	269.0
171.80	893.0
172.90	1196.0
173.90	1002.0
175.00	2212.0
175.90	663.0
176.90	645.0
177.90	585.0
178.90	3948.0
179.90	2368.0
180.90	1013.0
181.90	169.0
183.90	350.0
184.90	2087.0
186.00	14277.0
186.90	4149.0
187.90	551.0
188.90	948.0
190.90	576.0
191.90	1292.0
192.90	1759.0
194.00	422.0
194.90	253.0
195.90	2719.0
197.90	91664.0
198.90	6765.0
199.90	665.0
201.00	178.0
202.20	241.0
202.90	1189.0
204.00	5270.0
204.90	5963.0
205.90	30936.0
207.00	5701.0
207.90	1790.0
208.70	368.0
210.00	291.0
210.20	305.0
210.90	1288.0
211.60	268.0
212.80	217.0
213.90	228.0
214.90	705.0
215.90	519.0
216.90	9508.0
217.90	1350.0
220.90	3686.0
221.90	1717.0
222.90	1975.0
224.00	17000.0
224.90	3924.0
225.90	392.0
226.90	8592.0
227.90	1197.0
228.90	1340.0
230.10	269.0
230.90	543.0
232.10	170.0
234.00	523.0
235.00	633.0
235.90	626.0
236.90	470.0
238.90	409.0
239.80	203.0
240.80	295.0
241.10	223.0
241.90	1311.0
243.00	1124.0
244.00	12332.0
245.00	1345.0
245.90	3120.0
246.90	628.0
248.90	443.0
249.80	266.0
252.90	362.0
253.40	286.0
254.90	69992.0
256.00	10326.0
256.90	512.0
257.90	3874.0
259.00	515.0
260.10	179.0
260.90	213.0
263.90	191.0
265.00	1792.0
265.90	304.0
270.80	327.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

271.90	232.0
272.90	2182.0
273.90	5625.0
274.90	26896.0
275.90	3260.0
276.90	2757.0
277.90	431.0
278.90	391.0
281.80	165.0
283.00	201.0
283.80	517.0
285.00	437.0
291.80	428.0
292.90	420.0
293.70	354.0
293.90	343.0
295.00	287.0
295.90	9816.0
296.80	1342.0
298.00	156.0
300.80	184.0
302.00	221.0
302.90	755.0
304.00	328.0
308.90	226.0
313.90	463.0
314.90	1092.0
315.90	545.0
317.00	178.0
321.20	276.0
322.10	189.0
323.00	2512.0
324.00	490.0
324.70	340.0
326.80	506.0
328.00	250.0
332.00	223.0
333.00	170.0
333.90	2258.0
334.80	375.0
340.90	269.0
342.20	171.0
345.90	527.0
346.80	194.0
352.00	1102.0
353.00	608.0
354.00	960.0
354.90	222.0
365.00	3823.0
365.80	532.0
371.90	1651.0
373.00	211.0
376.70	187.0
383.00	365.0
384.10	161.0
390.00	246.0
390.80	155.0
401.90	474.0
402.90	536.0
404.00	299.0
420.90	774.0
422.00	378.0
422.90	3195.0
423.90	562.0
424.80	213.0
441.00	9531.0
442.00	54072.0
443.00	11315.0
444.00	1431.0
488.70	309.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

m/z	Abundance
39.00	1089.0
39.90	312.0
43.90	738.0
49.10	239.0
50.00	3861.0
51.00	15315.0
52.10	625.0
56.00	485.0
56.90	979.0
61.00	376.0
62.00	271.0
62.90	588.0
64.00	229.0
64.90	190.0
67.90	230.0
68.90	22544.0
69.90	265.0
73.00	296.0
73.90	2685.0
74.90	3801.0
76.00	1416.0
77.00	24440.0
78.00	1984.0
79.00	1317.0
79.90	1267.0
80.90	1205.0
82.00	595.0
83.00	291.0
83.90	237.0
84.90	215.0
85.10	212.0
85.90	497.0
86.80	289.0
88.00	186.0
90.80	668.0
91.90	314.0
92.90	3186.0
94.00	304.0
94.80	225.0
97.00	227.0
98.00	2524.0
99.00	1673.0
100.90	1117.0
102.80	300.0
104.00	616.0
104.90	658.0
105.90	252.0
106.90	7751.0
108.00	1285.0
109.90	12282.0
111.00	2446.0
111.80	268.0
115.90	344.0
117.00	6252.0
118.00	328.0
120.70	202.0
121.90	649.0
122.90	503.0
124.00	301.0
124.80	280.0
126.90	22440.0
127.90	1522.0
128.90	8618.0
129.90	800.0
130.80	286.0
133.00	152.0
133.90	415.0
134.90	1223.0
135.90	546.0
136.90	355.0
139.90	207.0
140.90	1419.0
141.90	260.0
142.90	212.0
146.90	756.0
148.00	1811.0
148.90	248.0
151.00	317.0
151.50	218.0
152.80	367.0
154.00	369.0
154.90	1324.0
156.00	1647.0
156.90	360.0
157.60	236.0
158.10	229.0
158.90	529.0
159.80	633.0
160.80	410.0
161.80	329.0
164.90	1089.0
166.00	367.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

167.00	3092.0
168.00	1253.0
169.00	227.0
169.90	167.0
172.00	433.0
172.90	485.0
173.90	958.0
175.00	1322.0
175.90	330.0
176.50	390.0
177.10	411.0
177.90	245.0
178.90	2123.0
179.90	1391.0
180.90	625.0
181.80	158.0
183.90	374.0
185.00	1088.0
186.00	8739.0
186.90	3942.0
187.80	372.0
188.90	473.0
190.00	228.0
190.80	217.0
191.90	892.0
193.00	844.0
193.80	463.0
195.90	2075.0
196.60	892.0
197.90	55416.0
198.90	4069.0
200.00	312.0
201.80	410.0
202.20	360.0
202.80	596.0
204.00	3190.0
204.90	4984.0
206.00	18520.0
207.00	2889.0
207.90	1146.0
209.00	247.0
209.90	397.0
210.90	726.0
212.00	172.0
215.00	395.0
216.00	430.0
216.90	5775.0
217.90	661.0
220.90	1409.0
221.20	1336.0
221.90	805.0
223.00	1638.0
223.90	12833.0
225.00	2934.0
226.90	5564.0
227.90	1160.0
228.90	1443.0
229.80	170.0
231.00	536.0
231.90	158.0
233.00	176.0
234.00	695.0
234.90	385.0
235.90	266.0
236.80	393.0
238.80	209.0
240.00	237.0
240.90	227.0
241.80	686.0
243.00	535.0
243.90	6684.0
245.00	975.0
245.90	1669.0
246.90	467.0
248.90	326.0
253.00	228.0
254.90	50184.0
255.90	7312.0
256.90	681.0
257.90	3394.0
258.90	424.0
260.90	197.0
264.90	1359.0
265.60	447.0
271.00	197.0
272.00	154.0
272.90	1770.0
273.90	3002.0
274.90	19200.0
275.90	1969.0
276.90	1624.0
278.00	316.0
281.70	157.0
282.80	324.0
284.00	194.0
285.00	307.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

293.00	546.0
293.90	228.0
295.90	7568.0
296.90	773.0
297.80	169.0
303.00	842.0
303.90	209.0
309.90	154.0
314.80	599.0
316.10	344.0
319.80	156.0
320.90	270.0
322.90	1917.0
323.90	660.0
326.80	554.0
327.80	202.0
331.90	163.0
333.10	368.0
334.00	1219.0
334.90	309.0
345.90	491.0
346.70	305.0
352.00	792.0
352.90	554.0
354.10	475.0
355.10	188.0
364.90	3508.0
366.20	317.0
370.80	211.0
371.90	1081.0
372.90	206.0
402.00	223.0
403.20	407.0
421.00	619.0
422.00	281.0
422.90	2203.0
424.00	319.0
441.00	8111.0
442.00	49648.0
443.00	9010.0
443.90	676.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17