


Data Path : Z:\HPCHEM1\BNA G\DATA\BG110716\
 Data File : BG024749.D
 Acq On : 8 Nov 2016 5:35
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG102516.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Oct 26 11:14:13 2016



AutoFind: Scans 2582, 2583, 2584; Background Corrected with Scan 2575

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.5	27213	PASS
68	69	0.00	2	0.9	284	PASS
69	198	0.00	100	34.5	30840	PASS
70	69	0.00	2	0.7	208	PASS
127	198	10	80	38.0	33914	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	89298	PASS
199	198	5	9	7.7	6849	PASS
275	198	10	60	33.3	29714	PASS
365	198	1	100	4.4	3891	PASS
441	442	0.01	24	17.8	15294	PASS
442	198	50	100	96.4	86120	PASS
443	442	15	24	20.0	17200	PASS