

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG110717\
 Data File : BG030827.D
 Acq On : 8 Nov 2017 3:16
 Operator : SJ/JU
 Sample : PB103871BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK71

Quant Time: Nov 08 04:45:39 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG101417.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 08 04:22:00 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.15	152	68250	20.00	ng/ul	0.00
18) Naphthalene-d8	10.97	136	327194	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.77	164	221068	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.51	188	535568	20.00	ng/ul	0.00
75) Chrysene-d12	21.78	240	559191	20.00	ng/ul	0.00
83) Perylene-d12	25.10	264	573500	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.53	96	10213	6.08	ng/uL	0.00
5) Phenol-d5	7.31	99	191296	29.20	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.46	67	115064	28.07	ng/ul	0.00
9) 2-Chlorophenol-d4	7.68	132	154530	33.44	ng/ul	0.00
13) 4-Methylphenol-d8	8.86	113	168450	31.84	ng/ul	0.00
19) Nitrobenzene-d5	9.32	128	82396	35.08	ng/ul	0.00
22) 2-Nitrophenol-d4	10.04	143	102022	42.39	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.59	165	174350	31.79	ng/ul	0.00
29) 4-Chloroaniline-d4	11.10	131	246526	38.69	ng/ul	0.00
43) Dimethylphthalate-d6	14.16	166	608325	32.19	ng/ul	0.00
46) Acenaphthylene-d8	14.46	160	740938	32.24	ng/ul	0.00
51) 4-Nitrophenol-d4	14.97	143	101857	29.19	ng/ul	0.00
57) Fluorene-d10	15.75	176	548416	35.43	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.88	200	94959	36.32	ng/ul	0.00
70) Anthracene-d10	17.61	188	859477	33.93	ng/ul	0.00
76) Pyrene-d10	19.89	212	958818	33.12	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.87	264	960098	35.35	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA G\DATA\BG110717\
Data File : BG030827.D
Acq On : 8 Nov 2017 3:16
Operator : SJ/JU
Sample : PB103871BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK71

Quant Time: Nov 08 04:45:39 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG101417.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 08 04:22:00 2017
Response via : Initial Calibration

