


Data Path : Z:\HPCHEM1\BNA_G\Data\BG110717\
 Data File : BG030809.D
 Acq On : 7 Nov 2017 15:44
 Operator : SJ/JU
 Sample : DFTPP67
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP67

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG101417.M
 Title : SVOA CALIBRATION
 Last Update : Wed Nov 08 04:22:00 2017



AutoFind: Scans 2490, 2491, 2492; Background Corrected with Scan 2483

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.1	33168	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	28.7	31662	PASS
70	69	0.00	2	0.8	260	PASS
127	198	10	80	39.2	43269	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	110256	PASS
199	198	5	9	6.6	7249	PASS
275	198	10	60	29.6	32624	PASS
365	198	1	100	3.8	4140	PASS
441	442	0.01	100	15.6	16681	PASS
442	198	50	100	97.0	106912	PASS
443	442	15	24	19.1	20373	PASS

m/z	Abundance
37.00	169.0
38.00	579.0
39.00	2627.0
40.00	189.0
41.00	235.0
43.10	360.0
44.00	758.0
49.00	518.0
50.00	9041.0
51.00	34720.0
52.00	1590.0
54.90	369.0
56.00	924.0
57.00	2579.0
59.80	158.0
61.00	509.0
62.10	472.0
63.10	1307.0
64.00	294.0
65.00	639.0
65.90	165.0
69.00	32512.0
70.10	299.0
71.10	234.0
73.00	399.0
74.00	4077.0
75.00	5299.0
76.00	1926.0
77.00	39840.0
78.00	2629.0
79.00	2250.0
79.90	2167.0
81.00	2813.0
82.00	586.0
83.00	831.0
83.90	215.0
85.00	535.0
85.90	728.0
86.90	344.0
90.90	1010.0
92.00	884.0
93.00	4323.0
93.90	433.0
95.00	180.0
95.90	226.0
97.10	307.0
98.00	3624.0
99.00	2878.0
100.00	300.0
100.90	1709.0
103.00	515.0
103.90	849.0
105.00	1093.0
106.20	511.0
107.00	11829.0
107.90	1782.0
109.00	536.0
110.00	22792.0
111.00	2790.0
112.00	391.0
113.00	229.0
116.00	856.0
117.00	8579.0
118.00	688.0
118.70	194.0
120.00	253.0
122.00	724.0
122.90	958.0
123.90	470.0
125.00	534.0
127.00	41888.0
128.00	3547.0
129.00	15938.0
130.00	1402.0
131.00	401.0
132.00	171.0
133.10	161.0
133.90	507.0
135.00	1661.0
136.00	672.0
137.00	727.0
138.00	156.0
138.90	209.0
139.90	279.0
141.00	2263.0
141.90	703.0
142.90	690.0
143.90	159.0
144.80	161.0
146.00	440.0
147.00	1230.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

148.00	2671.0
148.90	625.0
150.20	162.0
151.00	358.0
151.90	340.0
153.00	749.0
154.00	593.0
155.00	1699.0
156.00	2273.0
156.90	451.0
158.00	430.0
159.00	421.0
159.90	575.0
160.90	1126.0
162.00	298.0
164.00	201.0
164.90	629.0
166.00	1026.0
167.00	4584.0
168.00	2301.0
168.90	392.0
169.90	154.0
170.90	235.0
172.00	415.0
172.90	552.0
174.00	897.0
175.00	2164.0
176.00	586.0
177.00	1045.0
178.00	322.0
179.00	3003.0
179.90	2343.0
181.00	1515.0
181.90	218.0
182.70	332.0
183.90	284.0
185.00	2011.0
186.00	13576.0
187.00	4717.0
188.00	591.0
189.00	1165.0
190.00	195.0
191.00	465.0
192.00	1223.0
193.00	1413.0
194.00	158.0
194.90	245.0
196.00	2855.0
197.90	99448.0
198.90	7052.0
200.00	556.0
200.70	175.0
201.40	393.0
203.00	891.0
204.00	4079.0
205.00	6376.0
206.00	26896.0
207.00	4032.0
208.00	1097.0
208.90	187.0
210.20	425.0
211.00	1178.0
212.00	211.0
215.00	254.0
216.00	679.0
216.90	6496.0
218.00	1001.0
221.00	4345.0
222.00	949.0
223.00	1958.0
224.00	14475.0
225.00	3937.0
226.00	340.0
227.00	5834.0
227.90	907.0
228.90	1311.0
230.00	190.0
231.00	573.0
232.00	188.0
234.00	596.0
235.00	423.0
236.00	283.0
236.90	345.0
239.00	193.0
240.00	276.0
240.90	310.0
241.90	946.0
243.10	989.0
244.00	12101.0
245.00	1574.0
246.00	2100.0
246.90	339.0
249.00	315.0
252.10	270.0
252.80	239.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

255.00	59664.0
256.00	8604.0
256.90	611.0
258.00	3552.0
258.90	641.0
265.00	1118.0
266.00	298.0
270.90	157.0
272.00	207.0
272.90	1944.0
274.00	5122.0
275.00	28048.0
276.00	3992.0
277.00	2262.0
278.00	288.0
281.10	229.0
283.00	215.0
284.10	301.0
285.00	349.0
293.00	424.0
295.00	293.0
296.00	7480.0
297.00	1182.0
303.00	1012.0
304.00	185.0
309.00	172.0
314.00	253.0
315.00	850.0
316.00	458.0
321.00	179.0
322.00	160.0
323.00	2522.0
324.00	640.0
327.00	318.0
328.00	201.0
331.90	170.0
332.80	182.0
334.00	1204.0
335.00	340.0
341.40	203.0
346.00	435.0
352.00	780.0
352.90	485.0
353.90	540.0
355.10	153.0
365.00	3464.0
366.00	498.0
370.90	355.0
372.00	1293.0
373.00	310.0
383.00	282.0
383.90	330.0
390.00	256.0
402.00	461.0
402.90	445.0
404.00	257.0
420.90	704.0
422.00	548.0
423.00	3873.0
424.10	858.0
441.00	12156.0
442.00	76648.0
443.00	14479.0
444.00	1640.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

m/z	Abundance
38.10	620.0
39.00	2703.0
39.90	213.0
41.10	348.0
42.00	186.0
43.00	334.0
44.00	547.0
49.20	445.0
50.00	9915.0
51.00	37400.0
52.00	2238.0
55.00	433.0
56.00	1305.0
57.00	2748.0
58.00	297.0
61.00	449.0
62.00	680.0
63.00	1607.0
63.90	160.0
65.00	832.0
67.10	191.0
69.00	37312.0
70.00	192.0
71.00	193.0
73.00	570.0
74.00	4450.0
75.00	5204.0
76.00	2549.0
77.00	45712.0
78.00	2954.0
79.00	2982.0
80.00	2145.0
81.00	3425.0
82.00	948.0
83.00	888.0
84.00	156.0
85.00	623.0
86.00	908.0
86.90	477.0
91.00	683.0
91.90	872.0
93.00	4830.0
94.00	422.0
94.90	230.0
96.00	273.0
97.10	281.0
98.00	3814.0
98.90	2555.0
100.00	331.0
101.00	2139.0
101.90	163.0
103.00	602.0
104.00	1165.0
105.00	822.0
106.00	532.0
107.00	14011.0
108.00	2177.0
110.00	25616.0
111.00	3766.0
112.00	540.0
113.10	193.0
115.90	831.0
117.00	10968.0
117.90	865.0
118.90	215.0
120.10	164.0
121.00	172.0
122.00	1202.0
123.00	1479.0
123.90	843.0
125.00	921.0
127.00	49544.0
128.00	4275.0
129.00	19536.0
130.00	1878.0
130.90	385.0
133.20	182.0
133.90	508.0
135.00	1924.0
136.00	577.0
137.00	1013.0
138.00	270.0
139.00	173.0
140.00	304.0
141.00	2758.0
142.00	1016.0
143.00	736.0
143.90	357.0
144.90	162.0
145.90	512.0
146.90	1627.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

148.00	3098.0
149.00	720.0
150.00	286.0
151.00	383.0
152.00	302.0
152.90	849.0
154.00	829.0
155.00	1806.0
156.00	1993.0
157.00	457.0
158.00	429.0
159.00	386.0
160.00	686.0
161.00	1404.0
161.90	307.0
164.00	264.0
164.90	1349.0
166.00	813.0
167.00	6140.0
168.00	2605.0
169.00	383.0
170.10	194.0
171.00	200.0
171.90	830.0
173.10	528.0
174.00	1415.0
175.00	2459.0
176.00	813.0
176.90	1013.0
178.00	508.0
179.00	4210.0
180.00	3304.0
181.00	1569.0
182.00	421.0
184.10	534.0
185.00	2230.0
186.00	17264.0
187.00	4975.0
188.00	400.0
188.90	1313.0
189.60	272.0
190.00	232.0
191.00	797.0
192.00	1885.0
193.00	1866.0
194.00	414.0
194.90	231.0
195.90	3772.0
197.90	128984.0
198.90	8998.0
199.90	940.0
201.40	462.0
203.00	1167.0
204.00	4284.0
205.00	8043.0
206.00	36280.0
207.00	5306.0
208.00	1003.0
208.90	344.0
210.00	561.0
211.00	1367.0
211.90	288.0
215.00	557.0
216.00	900.0
216.90	9060.0
218.00	1033.0
221.00	5131.0
221.90	1289.0
223.00	2456.0
224.00	19984.0
225.00	5139.0
226.00	720.0
227.00	9346.0
228.00	1308.0
229.00	1630.0
230.00	219.0
231.00	867.0
233.90	519.0
234.90	416.0
236.00	612.0
236.90	635.0
238.90	411.0
239.90	229.0
241.00	556.0
242.00	1275.0
243.10	1265.0
244.00	14969.0
245.00	2290.0
246.00	2952.0
246.90	493.0
247.80	180.0
248.90	538.0
250.80	193.0
252.10	315.0
252.90	500.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

253.30	405.0
255.00	81496.0
256.00	11365.0
256.90	937.0
258.00	4243.0
259.00	769.0
265.00	1745.0
266.10	199.0
267.10	152.0
269.90	168.0
271.00	160.0
272.00	547.0
273.00	2912.0
274.00	6819.0
275.00	36544.0
276.00	4439.0
277.00	3038.0
277.90	444.0
282.90	375.0
283.90	368.0
284.90	794.0
285.90	216.0
290.90	176.0
291.90	159.0
292.90	814.0
294.90	442.0
296.00	10266.0
297.00	1232.0
302.00	256.0
303.00	1226.0
304.00	193.0
309.90	174.0
314.00	404.0
315.00	1776.0
316.00	544.0
317.00	179.0
319.10	183.0
320.90	438.0
321.90	312.0
323.00	3663.0
324.00	699.0
325.10	184.0
326.90	474.0
327.80	178.0
328.90	221.0
332.00	253.0
333.00	322.0
334.10	2238.0
335.00	634.0
340.90	290.0
342.00	229.0
346.00	427.0
346.90	247.0
350.10	157.0
352.00	997.0
353.00	699.0
354.00	1136.0
354.90	212.0
365.00	5056.0
366.00	749.0
371.00	166.0
372.00	1832.0
373.00	470.0
383.00	316.0
390.10	341.0
402.00	445.0
403.00	1038.0
404.00	488.0
421.00	1041.0
422.00	771.0
423.00	6527.0
424.00	1331.0
425.00	183.0
441.00	18560.0
442.00	123040.0
443.00	23816.0
444.00	1947.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

m/z	Abundance
36.90	181.0
38.00	225.0
39.00	2388.0
40.00	227.0
41.00	433.0
43.10	463.0
44.00	458.0
48.90	287.0
50.00	7313.0
51.00	27384.0
52.00	1414.0
55.00	529.0
56.00	1056.0
57.00	2232.0
61.00	568.0
62.00	306.0
63.00	1266.0
65.00	602.0
66.00	255.0
66.90	233.0
69.00	25848.0
70.00	291.0
71.00	194.0
73.10	375.0
74.00	3046.0
75.00	4309.0
76.00	1634.0
77.00	31504.0
78.00	2024.0
79.00	2112.0
80.00	1999.0
81.00	2202.0
82.00	569.0
83.00	640.0
84.00	247.0
85.00	597.0
85.90	614.0
87.00	230.0
91.00	616.0
92.00	464.0
93.00	3681.0
94.00	421.0
95.10	169.0
97.00	184.0
98.00	2862.0
99.00	2276.0
99.90	259.0
101.00	1354.0
102.00	161.0
103.00	484.0
104.00	946.0
104.90	633.0
107.00	11640.0
107.90	1902.0
110.00	17528.0
111.00	2720.0
112.00	327.0
116.00	991.0
117.00	8722.0
118.00	520.0
120.00	248.0
122.00	758.0
122.90	1040.0
123.90	495.0
125.00	523.0
127.00	38376.0
128.00	3131.0
129.00	13958.0
130.00	1406.0
130.90	407.0
133.90	448.0
135.00	1659.0
135.90	442.0
137.00	593.0
138.00	192.0
139.80	250.0
140.90	1790.0
142.00	613.0
143.00	636.0
144.00	259.0
145.90	306.0
147.00	1222.0
148.00	2406.0
149.00	638.0
150.00	264.0
151.20	285.0
151.90	274.0
153.00	562.0
153.90	600.0
155.00	1217.0
156.00	2449.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

157.00	305.0
157.90	463.0
158.90	246.0
160.00	959.0
161.00	937.0
161.80	356.0
165.00	1076.0
165.90	720.0
167.00	4236.0
168.00	2076.0
169.00	283.0
170.20	174.0
171.00	192.0
171.90	340.0
172.90	501.0
174.00	1006.0
175.00	2311.0
175.90	494.0
176.90	816.0
178.00	506.0
178.90	3314.0
180.00	2423.0
181.00	933.0
183.00	175.0
184.00	282.0
185.00	1648.0
186.00	12743.0
187.00	3883.0
188.10	559.0
188.90	897.0
189.90	189.0
191.00	551.0
192.00	1154.0
193.00	1505.0
193.80	350.0
194.90	383.0
196.00	2817.0
197.90	102336.0
198.90	5697.0
199.90	543.0
201.40	442.0
201.90	214.0
202.80	541.0
204.00	4144.0
205.00	6487.0
206.00	28288.0
207.00	4146.0
207.90	840.0
208.90	418.0
210.10	505.0
211.00	1422.0
212.90	160.0
215.00	376.0
215.90	684.0
217.00	7411.0
218.00	806.0
221.00	4414.0
223.00	1872.0
224.00	15958.0
225.00	4482.0
226.00	672.0
227.00	6246.0
228.00	1039.0
229.00	1321.0
231.00	773.0
233.90	484.0
235.00	551.0
235.90	467.0
237.00	338.0
238.90	203.0
240.10	210.0
240.90	273.0
241.90	777.0
243.00	1217.0
244.00	13618.0
245.00	2322.0
246.00	2773.0
246.90	409.0
248.10	173.0
248.90	587.0
252.10	252.0
253.00	375.0
255.00	67840.0
256.00	10008.0
257.00	1001.0
258.00	3786.0
258.90	446.0
263.90	161.0
265.00	1910.0
265.90	319.0
270.90	289.0
272.00	277.0
273.00	2329.0
274.00	6047.0
275.00	33280.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

276.00	4502.0
277.00	2710.0
278.10	268.0
282.00	163.0
283.00	406.0
283.90	240.0
285.00	466.0
289.00	173.0
290.00	176.0
293.00	689.0
293.90	191.0
294.30	152.0
296.00	9208.0
297.00	1133.0
298.00	168.0
302.00	151.0
303.00	1156.0
304.00	284.0
313.90	374.0
315.00	848.0
316.00	471.0
317.00	239.0
321.00	366.0
322.20	202.0
323.00	2973.0
324.00	552.0
326.90	797.0
327.90	331.0
332.00	226.0
333.00	281.0
334.00	2160.0
335.00	451.0
341.00	310.0
346.00	648.0
347.00	257.0
352.00	1005.0
353.00	504.0
354.00	926.0
354.90	334.0
365.00	3901.0
365.90	767.0
370.80	210.0
372.00	1404.0
373.00	398.0
382.90	398.0
389.90	373.0
402.00	785.0
403.00	1044.0
404.00	498.0
421.00	857.0
422.00	765.0
423.00	6014.0
424.00	1096.0
441.00	19328.0
442.00	121048.0
443.00	22824.0
444.00	2546.0
445.10	201.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

m/z	Abundance
41.10	313.0
43.10	425.0
44.00	322.0
54.90	446.0
57.00	200.0
67.10	159.0
69.00	228.0
84.10	242.0
91.00	222.0
147.20	186.0
173.00	154.0
207.10	391.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67