

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110922\
 Data File : BG055525.D
 Acq On : 9 Nov 2022 15:52
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG110922

Quant Time: Nov 10 03:49:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 10 03:39:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.533	0.500	6.2	95	0.00
3	Pyridine	1.483	1.561	-5.3	101	0.00
4	n-Nitrosodimethylamine	0.807	0.824	-2.1	102	0.00
5 S	2-Fluorophenol	1.050	1.109	-5.6	103	0.00
6	Aniline	1.927	1.950	-1.2	100	0.00
7 S	Phenol-d6	1.450	1.520	-4.8	101	0.00
8	2-Chlorophenol	1.176	1.240	-5.4	103	0.00
9	Benzaldehyde	1.144	1.142	0.2	103	0.00
10 C	Phenol	1.670	1.730	-3.6	102	0.00
11	bis(2-Chloroethyl)ether	1.344	1.331	1.0	100	0.00
12	1,3-Dichlorobenzene	1.468	1.463	0.3	102	0.00
13 C	1,4-Dichlorobenzene	1.486	1.489	-0.2	101	0.00
14	1,2-Dichlorobenzene	1.411	1.384	1.9	99	0.00
15	Benzyl Alcohol	1.265	1.341	-6.0	101	0.00
16	2,2'-oxybis(1-Chloropropane	2.583	2.591	-0.3	100	0.00
17	2-Methylphenol	1.175	1.207	-2.7	101	0.00
18	Hexachloroethane	0.493	0.500	-1.4	101	0.00
19 P	n-Nitroso-di-n-propylamine	1.199	1.225	-2.2	100	0.00
20	3+4-Methylphenols	1.637	1.727	-5.5	101	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
22	Acetophenone	0.532	0.524	1.5	101	0.00
23 S	Nitrobenzene-d5	0.295	0.350	-18.6	109	0.00
24	Nitrobenzene	0.339	0.382	-12.7	106	0.00
25	Isophorone	0.754	0.753	0.1	99	0.00
26 C	2-Nitrophenol	0.091	0.125	-37.4#	118	0.00
27	2,4-Dimethylphenol	0.269	0.286	-6.3	103	0.00
28	bis(2-Chloroethoxy)methane	0.406	0.401	1.2	99	0.00
29 C	2,4-Dichlorophenol	0.296	0.324	-9.5	104	0.00
30	1,2,4-Trichlorobenzene	0.371	0.372	-0.3	101	0.00
31	Naphthalene	1.081	1.069	1.1	100	0.00
32	Benzoic acid	0.150	0.170	-13.3	117	0.00
33	4-Chloroaniline	0.452	0.449	0.7	100	0.00
34 C	Hexachlorobutadiene	0.247	0.243	1.6	99	0.00
35	Caprolactam	0.101	0.112	-10.9	104	0.00
36 C	4-Chloro-3-methylphenol	0.349	0.370	-6.0	102	0.00
37	2-Methylnaphthalene	0.812	0.809	0.4	101	0.00
38	1-Methylnaphthalene	0.795	0.789	0.8	101	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	0.597	0.601	-0.7	100	0.00
41 P	Hexachlorocyclopentadiene	0.256	0.282	-10.2	103	0.00
42 S	2,4,6-Tribromophenol	0.195	0.208	-6.7	96	0.00
43 C	2,4,6-Trichlorophenol	0.351	0.402	-14.5	106	0.00
44	2,4,5-Trichlorophenol	0.402	0.445	-10.7	102	0.00
45 S	2-Fluorobiphenyl	1.326	1.298	2.1	100	0.00
46	1,1'-Biphenyl	1.463	1.446	1.2	99	0.00
47	2-Chloronaphthalene	1.130	1.119	1.0	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110922\
 Data File : BG055525.D
 Acq On : 9 Nov 2022 15:52
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG110922

Quant Time: Nov 10 03:49:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 10 03:39:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.232	0.317	-36.6#	111	0.00
49	Acenaphthylene	1.870	1.861	0.5	99	0.00
50	Dimethylphthalate	1.520	1.577	-3.7	100	0.00
51	2,6-Dinitrotoluene	0.219	0.281	-28.3#	104	0.00
52 C	Acenaphthene	1.167	1.165	0.2	99	0.00
53	3-Nitroaniline	0.281	0.326	-16.0	105	0.00
54 P	2,4-Dinitrophenol	0.094	0.098	-4.3	105	0.00
55	Dibenzofuran	1.866	1.849	0.9	100	0.00
56 P	4-Nitrophenol	0.225	0.248	-10.2	106	0.00
57	2,4-Dinitrotoluene	0.283	0.379	-33.9#	108	0.00
58	Fluorene	1.485	1.466	1.3	99	0.00
59	2,3,4,6-Tetrachlorophenol	0.373	0.417	-11.8	102	0.00
60	Diethylphthalate	1.546	1.578	-2.1	99	0.00
61	4-Chlorophenyl-phenylether	0.790	0.774	2.0	98	0.00
62	4-Nitroaniline	0.301	0.332	-10.3	102	0.00
63	Azobenzene	1.499	1.465	2.3	98	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	0.061	0.065	-6.6	108	0.00
66 c	n-Nitrosodiphenylamine	0.560	0.558	0.4	98	0.00
67	4-Bromophenyl-phenylether	0.204	0.194	4.9	93	0.00
68	Hexachlorobenzene	0.243	0.240	1.2	98	0.00
69	Atrazine	0.203	0.212	-4.4	98	0.00
70 C	Pentachlorophenol	0.133	0.142	-6.8	102	0.00
71	Phenanthrene	1.075	1.058	1.6	98	0.00
72	Anthracene	1.053	1.028	2.4	97	0.00
73	Carbazole	0.952	0.933	2.0	97	0.00
74	Di-n-butylphthalate	1.092	1.146	-4.9	98	0.00
75 C	Fluoranthene	1.282	1.259	1.8	98	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
77	Benzydine	0.541	0.541	0.0	94	0.00
78	Pyrene	1.354	1.338	1.2	98	0.00
79 S	Terphenyl-d14	0.996	0.952	4.4	97	0.00
80	Butylbenzylphthalate	0.453	0.533	-17.7	106	0.00
81	Benzo(a)anthracene	1.370	1.360	0.7	98	0.00
82	3,3'-Dichlorobenzidine	0.447	0.458	-2.5	97	0.00
83	Chrysene	1.303	1.297	0.5	99	0.00
84	Bis(2-ethylhexyl)phthalate	0.683	0.769	-12.6	103	0.00
85 c	Di-n-octyl phthalate	1.181	1.306	-10.6	105	0.00
86 I	Perylene-d12	1.000	1.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	1.523	1.523	0.0	100	0.00
88	Benzo(b)fluoranthene	1.267	1.281	-1.1	101	0.00
89	Benzo(k)fluoranthene	1.296	1.253	3.3	97	0.00
90 C	Benzo(a)pyrene	1.092	1.084	0.7	99	0.00
91	Dibenzo(a,h)anthracene	1.261	1.245	1.3	99	0.00
92	Benzo(g,h,i)perylene	1.233	1.233	0.0	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG110922\
Data File : BG055525.D
Acq On : 9 Nov 2022 15:52
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG110922

Quant Time: Nov 10 03:49:42 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 10 03:39:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1