

Data Path : Z:\HPCHEM1\BNA G\DATA\BG111015\
 Data File : BG019548.D
 Acq On : 10 Nov 2015 17:01
 Operator : UM/NP
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02010

Quant Time: Nov 13 08:48:31 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 13 08:28:14 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	150#	0.00
2	1,4-Dioxane	0.476	0.463	2.7	179#	0.00
3 S	1,4-Dioxane-d8	0.418	0.428	-2.4	177#	0.00
4	Benzaldehyde	1.194	1.425	-19.3	156#	0.00
5 S	Phenol-d5	1.836	1.846	-0.5	159#	0.00
6	Phenol	1.956	1.977	-1.1	155#	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.164	1.238	-6.4	165#	0.00
8	Bis(2-Chloroethyl)ether	1.344	1.485	-10.5	168#	0.00
9 S	2-Chlorophenol-d4	1.146	1.142	0.3	156#	0.00
10	2-Chlorophenol	1.180	1.171	0.8	154#	0.00
11	2-Methylphenol	1.446	1.484	-2.6	161#	0.00
12	2,2'-oxybis(1-Chloropropane	1.077	1.271	-18.0	186#	0.00
13 S	4-Methylphenol-d8	1.462	1.526	-4.4	160#	0.00
14	Acetophenone	2.455	2.575	-4.9	158#	0.00
15 P	N-Nitroso-di-n-propylamine	1.609	1.702	-5.8	170#	0.00
16	4-Methylphenol	1.580	1.637	-3.6	159#	0.00
17	Hexachloroethane	0.616	0.553	10.2	154#	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	156#	0.00
19 S	Nitrobenzene-d5	0.147	0.147	0.0	161#	0.00
20	Nitrobenzene	0.575	0.584	-1.6	161#	0.00
21	Isophorone	1.041	1.136	-9.1	176#	0.00
22 S	2-Nitrophenol-d4	0.174	0.172	1.1	151#	0.00
23 C	2-Nitrophenol	0.198	0.186	6.1	151#	0.00
24	2,4-Dimethylphenol	0.518	0.513	1.0	158#	0.00
25	Bis(2-Chloroethoxy)methane	0.493	0.535	-8.5	172#	0.00
26 S	2,4-Dichlorophenol-d3	0.379	0.372	1.8	154#	0.00
27 C	2,4-Dichlorophenol	0.362	0.360	0.6	151#	0.00
28	Naphthalene	1.001	0.974	2.7	153#	0.00
29 S	4-Chloroaniline-d4	0.383	0.436	-13.8	155#	0.00
30	4-Chloroaniline	0.401	0.450	-12.2	156#	0.00
31 C	Hexachlorobutadiene	0.370	0.329	11.1	138	0.00
32	Caprolactam	0.130	0.147	-13.1	176#	0.00
33 C	4-Chloro-3-methylphenol	0.486	0.489	-0.6	156#	0.00
34	2-Methylnaphthalene	0.804	0.779	3.1	151#	0.00
35	1-Methylnaphthalene	0.761	0.782	-2.8	158#	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	164#	0.00
37	1,2,4,5-Tetrachlorobenzene	0.824	0.730	11.4	145	0.00
38	Hexachlorocyclopentadiene	0.612	0.459	25.0#	123	0.00
39 C	2,4,6-Trichlorophenol	0.487	0.446	8.4	149	0.00
40	2,4,5-Trichlorophenol	0.535	0.503	6.0	158#	0.00
41	1,1'-Biphenyl	1.414	1.345	4.9	153#	0.00
42	2-Chloronaphthalene	1.103	1.032	6.4	154#	0.00
43	2-Nitroaniline	0.472	0.460	2.5	157#	0.00
44 S	Dimethylphthalate-d6	1.649	1.666	-1.0	166#	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG111015\
 Data File : BG019548.D
 Acq On : 10 Nov 2015 17:01
 Operator : UM/NP
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02010

Quant Time: Nov 13 08:48:31 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 13 08:28:14 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.686	1.639	2.8	162#	0.00
46	2,6-Dinitrotoluene	0.341	0.324	5.0	156#	0.00
47 S	Acenaphthylene-d8	1.826	1.755	3.9	156#	0.00
48	Acenaphthylene	1.871	1.788	4.4	157#	0.00
49	3-Nitroaniline	0.294	0.290	1.4	155#	0.00
50 C	Acenaphthene	1.263	1.194	5.5	155#	0.00
51	2,4-Dinitrophenol	0.257	0.149	42.0#	100	0.00
52 S	4-Nitrophenol-d4	0.262	0.244	6.9	151#	0.00
53	4-Nitrophenol	0.409	0.332	18.8	131	0.00
54	Dibenzofuran	1.861	1.763	5.3	155#	0.00
55	2,4-Dinitrotoluene	0.539	0.494	8.3	152#	0.00
56	2,3,4,6-Tetrachlorophenol	0.559	0.497	11.1	144	0.00
57	Diethylphthalate	1.665	1.600	3.9	159#	0.00
58 S	Fluorene-d10	1.525	1.418	7.0	149	0.00
59	Fluorene	1.623	1.530	5.7	154#	0.00
60	4-Chlorophenyl-phenylether	0.965	0.888	8.0	151#	0.00
61	4-Nitroaniline	0.338	0.316	6.5	150	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	156#	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.129	0.112	13.2	137	0.00
64	4,6-Dinitro-2-methylphenol	0.140	0.115	17.9	130	0.00
65	N-Nitrosodiphenylamine	0.517	0.512	1.0	154#	0.00
66	4-Bromophenyl-phenylether	0.238	0.227	4.6	147	0.00
67	Hexachlorobenzene	0.252	0.235	6.7	145	0.00
68	Atrazine	0.256	0.239	6.6	144	0.00
69 C	Pentachlorophenol	0.149	0.118	20.8	130	0.00
70	Phenanthrene	1.009	0.978	3.1	152#	0.00
71 S	Anthracene-d10	0.912	0.890	2.4	154#	0.00
72	Anthracene	1.029	0.988	4.0	151#	0.00
73	1,2,3,4-Tetrachlorobenzene	0.290	0.276	4.8	145	0.00
74	Pentachlorobenzene	0.292	0.278	4.8	147	0.00
75	Carbazole	0.821	0.808	1.6	148	0.00
76	Di-n-butylphthalate	1.018	0.949	6.8	147	0.00
77 C	Fluoranthene	1.307	1.260	3.6	142	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	158#	0.00
79 S	Pyrene-d10	0.875	0.810	7.4	146	0.00
80	Pyrene	1.122	1.032	8.0	145	0.00
81	Butylbenzylphthalate	0.365	0.359	1.6	157#	0.00
82	3,3'-Dichlorobenzidine	0.392	0.374	4.6	141	0.00
83	Benzo(a)anthracene	1.165	1.118	4.0	152#	0.00
84	Bis(2-ethylhexyl)phthalate	0.518	0.526	-1.5	162#	0.00
85	Chrysene	1.093	1.016	7.0	148	0.00
86 I	Perylene-d12	1.000	1.000	0.0	157#	-0.01
87	Di-n-octyl phthalate	0.906	0.950	-4.9	160#	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG111015\
Data File : BG019548.D
Acq On : 10 Nov 2015 17:01
Operator : UM/NP
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02010

Quant Time: Nov 13 08:48:31 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 13 08:28:14 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.180	1.143	3.1	154#	0.00
89	Benzo(k)fluoranthene	1.135	1.088	4.1	152#	0.00
90 S	Benzo(a)pyrene-d12	1.004	0.963	4.1	153#	0.00
91 C	Benzo(a)pyrene	1.133	1.085	4.2	152#	0.00
92	Indeno(1,2,3-cd)pyrene	1.277	1.225	4.1	153#	-0.02
93	Dibenzo(a,h)anthracene	1.066	1.000	6.2	151#	-0.01
94	Benzo(a,h,i)perylene	0.969	1.014	-4.6	154#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0