

Data Path : Z:\HPCHEM1\BNA G\DATA\BG111015\
 Data File : BG019557.D
 Acq On : 10 Nov 2015 22:47
 Operator : UM/NP
 Sample : G4308-13
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 D9KR1

Quant Time: Nov 13 08:38:42 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 13 08:28:14 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	45006	20.00	ng/ul	0.00
18) Naphthalene-d8	11.01	136	189945	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.81	164	155955	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.55	188	442293	20.00	ng/ul	0.00
78) Chrysene-d12	21.83	240	564101	20.00	ng/ul	0.00
86) Perylene-d12	25.18	264	541612	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.47	96	905	0.96	ng/uL	-0.01
5) Phenol-d5	7.32	99	27527	6.66	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.49	67	81921	31.28	ng/ul	0.00
9) 2-Chlorophenol-d4	7.70	132	58074	22.52	ng/ul	0.00
13) 4-Methylphenol-d8	8.88	113	51575	15.67	ng/ul	0.00
19) Nitrobenzene-d5	9.34	128	39330	28.17	ng/ul	0.00
22) 2-Nitrophenol-d4	10.07	143	46837	28.26	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.62	165	90715	25.23	ng/ul	0.00
29) 4-Chloroaniline-d4	11.14	131	3494	0.96	ng/ul	0.00
44) Dimethylphthalate-d6	14.20	166	382117	29.72	ng/ul	0.00
47) Acenaphthylene-d8	14.50	160	405682	28.49	ng/ul	0.00
52) 4-Nitrophenol-d4	14.99	143	11686	5.72	ng/ul	0.00
58) Fluorene-d10	15.79	176	334555	28.12	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.91	200	67405	23.55	ng/ul	0.00
71) Anthracene-d10	17.65	188	564822	28.01	ng/ul	0.00
79) Pyrene-d10	19.92	212	661420	26.79	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.95	264	756820	27.85	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG111015\
Data File : BG019557.D
Acq On : 10 Nov 2015 22:47
Operator : UM/NP
Sample : G4308-13
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
D9KR1

Quant Time: Nov 13 08:38:42 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 13 08:28:14 2015
Response via : Initial Calibration

