


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG111017\
 Data File : BG030901.D
 Acq On : 10 Nov 2017 10:32
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG111017.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Nov 10 15:38:25 2017



AutoFind: Scans 2492, 2493, 2494; Background Corrected with Scan 2485

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.6	19195	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	27.9	19388	PASS
70	69	0.00	2	1.6	311	PASS
127	198	10	80	37.7	26221	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	69592	PASS
199	198	5	9	6.5	4517	PASS
275	198	10	60	30.5	21229	PASS
365	198	1	100	4.2	2905	PASS
441	442	0.01	24	15.1	9762	PASS
442	198	50	100	93.1	64778	PASS
443	442	15	24	19.8	12847	PASS