

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111022\
 Data File : BG055544.D
 Acq On : 10 Nov 2022 18:11
 Operator : CG/JU
 Sample : N5491-11
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-4

Quant Time: Nov 11 02:01:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 10 03:39:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.270	152	48866	20.000	ng	0.00
21) Naphthalene-d8	11.097	136	185932	20.000	ng	0.00
39) Acenaphthene-d10	14.886	164	94920	20.000	ng	0.00
64) Phenanthrene-d10	17.624	188	285006	20.000	ng	0.00
76) Chrysene-d12	21.919	240	201282	20.000	ng	#-0.01
86) Perylene-d12	25.333	264	10971	20.000	ng	#-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.797	112	421322	164.225	ng	0.00
7) Phenol-d6	7.407	99	460259	129.876	ng	0.00
23) Nitrobenzene-d5	9.440	82	349572	127.456	ng	0.00
42) 2,4,6-Tribromophenol	16.367	330	236854	255.744	ng	0.00
45) 2-Fluorobiphenyl	13.517	172	803329	127.659	ng	0.00
79) Terphenyl-d14	20.215	244	1435278	143.224	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111022\
Data File : BG055544.D
Acq On : 10 Nov 2022 18:11
Operator : CG/JU
Sample : N5491-11
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-4

Quant Time: Nov 11 02:01:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 10 03:39:46 2022
Response via : Initial Calibration

