

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111022\
 Data File : BG055534.D
 Acq On : 10 Nov 2022 11:02
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 11 01:57:47 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 10 03:39:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	113	0.00
2	1,4-Dioxane	40.000	38.749	3.1	109	0.00
3	Pyridine	40.000	42.366	-5.9	113	0.00
4	n-Nitrosodimethylamine	40.000	39.618	1.0	110	0.00
5 S	2-Fluorophenol	80.000	80.562	-0.7	110	0.00
6	Aniline	40.000	39.653	0.9	109	0.00
7 S	Phenol-d6	80.000	80.949	-1.2	108	0.00
8	2-Chlorophenol	40.000	39.636	0.9	108	0.00
9	Benzaldehyde	40.000	38.530	3.7	111	0.00
10 C	Phenol	40.000	39.464	1.3	108	0.00
11	bis(2-Chloroethyl)ether	40.000	39.601	1.0	112	0.00
12	1,3-Dichlorobenzene	40.000	39.148	2.1	111	0.00
13 C	1,4-Dichlorobenzene	40.000	39.293	1.8	110	0.00
14	1,2-Dichlorobenzene	40.000	39.215	2.0	110	0.00
15	Benzyl Alcohol	40.000	39.920	0.2	106	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.548	1.1	110	0.00
17	2-Methylphenol	40.000	40.221	-0.6	109	0.00
18	Hexachloroethane	40.000	40.084	-0.2	111	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.976	0.1	109	0.00
20	3+4-Methylphenols	40.000	40.017	-0.0	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	109	0.00
22	Acetophenone	40.000	38.976	2.6	109	0.00
23 S	Nitrobenzene-d5	80.000	87.241	-9.1	109	0.00
24	Nitrobenzene	40.000	41.933	-4.8	108	0.00
25	Isophorone	40.000	40.414	-1.0	110	0.00
26 C	2-Nitrophenol	40.000	34.371	14.1	96	0.00
27	2,4-Dimethylphenol	40.000	40.793	-2.0	108	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.793	-2.0	111	0.00
29 C	2,4-Dichlorophenol	40.000	40.526	-1.3	105	0.00
30	1,2,4-Trichlorobenzene	40.000	39.442	1.4	108	0.00
31	Naphthalene	40.000	39.097	2.3	108	0.00
32	Benzoic acid	40.000	37.276	6.8	102	0.00
33	4-Chloroaniline	40.000	39.912	0.2	109	0.00
34 C	Hexachlorobutadiene	40.000	38.923	2.7	107	0.00
35	Caprolactam	40.000	42.587	-6.5	109	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.250	-3.1	109	0.00
37	2-Methylnaphthalene	40.000	39.226	1.9	108	0.00
38	1-Methylnaphthalene	40.000	38.878	2.8	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	108	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.504	3.7	105	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.717	0.7	103	0.00
42 S	2,4,6-Tribromophenol	80.000	87.985	-10.0	109	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.135	-2.8	105	0.00
44	2,4,5-Trichlorophenol	40.000	41.652	-4.1	105	0.00
45 S	2-Fluorobiphenyl	80.000	76.452	4.4	107	0.00
46	1,1'-Biphenyl	40.000	38.887	2.8	107	0.00
47	2-Chloronaphthalene	40.000	39.346	1.6	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111022\
 Data File : BG055534.D
 Acq On : 10 Nov 2022 11:02
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 11 01:57:47 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 10 03:39:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.700	3.2	106	0.00
49	Acenaphthylene	40.000	39.416	1.5	108	0.00
50	Dimethylphthalate	40.000	40.173	-0.4	106	0.00
51	2,6-Dinitrotoluene	40.000	39.321	1.7	105	0.00
52 C	Acenaphthene	40.000	39.097	2.3	106	0.00
53	3-Nitroaniline	40.000	38.684	3.3	108	0.00
54 P	2,4-Dinitrophenol	40.000	37.851	5.4	105	0.00
55	Dibenzofuran	40.000	38.793	3.0	108	0.00
56 P	4-Nitrophenol	40.000	40.569	-1.4	108	0.00
57	2,4-Dinitrotoluene	40.000	39.693	0.8	108	0.00
58	Fluorene	40.000	38.413	4.0	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.990	-2.5	103	0.00
60	Diethylphthalate	40.000	40.199	-0.5	107	0.00
61	4-Chlorophenyl-phenylether	40.000	38.556	3.6	106	0.00
62	4-Nitroaniline	40.000	43.083	-7.7	110	0.00
63	Azobenzene	40.000	38.989	2.5	108	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.239	1.9	106	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.021	-0.1	107	0.00
67	4-Bromophenyl-phenylether	40.000	41.910	-4.8	111	0.00
68	Hexachlorobenzene	40.000	39.191	2.0	105	0.00
69	Atrazine	40.000	40.804	-2.0	104	0.00
70 C	Pentachlorophenol	40.000	39.691	0.8	103	0.00
71	Phenanthrene	40.000	39.111	2.2	106	0.00
72	Anthracene	40.000	38.800	3.0	104	0.00
73	Carbazole	40.000	39.248	1.9	106	0.00
74	Di-n-butylphthalate	40.000	41.281	-3.2	104	0.00
75 C	Fluoranthene	40.000	38.489	3.8	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzidine	40.000	41.678	-4.2	103	0.00
78	Pyrene	40.000	40.392	-1.0	106	0.00
79 S	Terphenyl-d14	80.000	79.122	1.1	106	0.00
80	Butylbenzylphthalate	40.000	39.376	1.6	105	0.00
81	Benzo(a)anthracene	40.000	39.902	0.2	103	0.00
82	3,3'-Dichlorobenzidine	40.000	42.146	-5.4	105	0.00
83	Chrysene	40.000	39.939	0.2	104	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.986	-10.0	105	0.00
85 c	Di-n-octyl phthalate	40.000	41.844	-4.6	104	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.325	1.7	104	0.00
88	Benzo(b)fluoranthene	40.000	39.619	1.0	105	0.00
89	Benzo(k)fluoranthene	40.000	38.979	2.6	103	0.00
90 C	Benzo(a)pyrene	40.000	39.238	1.9	103	0.00
91	Dibenzo(a,h)anthracene	40.000	39.363	1.6	104	0.00
92	Benzo(g,h,i)perylene	40.000	39.487	1.3	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111022\
Data File : BG055534.D
Acq On : 10 Nov 2022 11:02
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 11 01:57:47 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 10 03:39:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0