



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111022\
 Data File : BG055533.D
 Acq On : 10 Nov 2022 10:21
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110922.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Nov 10 03:39:46 2022






AutoFind: Scans 2511, 2512, 2513; Background Corrected with Scan 2504

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.8	53144	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	33.8	43955	PASS
70	69	0.00	2	0.6	277	PASS
127	198	10	80	43.0	56019	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	130200	PASS
199	198	5	9	7.0	9161	PASS
275	198	10	60	29.9	38885	PASS
365	198	1	100	4.5	5819	PASS
441	198	0.01	100	17.3	22477	PASS
442	442	50	100	100.0	143597	PASS
443	442	15	24	19.3	27741	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111022\
Data File : BG055533.D
Acq On : 10 Nov 2022 10:21
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Nov 11 04:58:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 10 03:39:46 2022
Response via : Initial Calibration



TIC: BG055533.D\data.ms

(70) Pentachlorophenol (C)

17.271min (+ 0.005) 17168.58 ng




response 89095

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.20	65.11
264.00	61.90	65.05
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111022\
Data File : BG055533.D
Acq On : 10 Nov 2022 10:21
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Nov 11 04:58:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG110922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 10 03:39:46 2022
Response via : Initial Calibration



TIC: BG055533.D\data.ms

(77) Benzidine

19.838min (+ 0.005) 0.00 ng

response 931044

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	14.38
183.00	12.70	12.55
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
11/10/2022	BNA_G	<u>BG055533.D</u>
Compound Name	Response	Retention Time
DDT	281184	21.078
DDD	13745	20.626
DDE	66	20.12
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
13811	294995	4.68

Instrument :
BNA_G
ClientSampleId :
DFTPP