

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG111117\
 Data File : BG030935.D
 Acq On : 11 Nov 2017 9:37
 Operator : SJ/JU
 Sample : I6413-01
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 FM728

Manual Integrations
 APPROVED

SOHIL
 11/13/2017 4:55:20 PM

Quant Time: Nov 13 03:48:46 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG111017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 10 15:38:25 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	41004	20.00	ng	-0.01
21) Naphthalene-d8	10.96	136	179342	20.00	ng	-0.01
38) Acenaphthene-d10	14.76	164	133625	20.00	ng	-0.01
63) Phenanthrene-d10	17.50	188	351138	20.00	ng	-0.01
75) Chrysene-d12	21.77	240	414555	20.00	ng	-0.01
86) Perylene-d12	25.08	264	418615	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.70	112	40723	17.03	ng	0.00
7) Phenol-d6	7.31	99	35008	10.87	ng	0.00
23) Nitrobenzene-d5	9.30	82	114828	37.94	ng	-0.02
41) 2,4,6-Tribromophenol	16.24	330	76542	35.41	ng	-0.01
44) 2-Fluorobiphenyl	13.38	172	364202	39.16	ng	-0.02
78) Terphenyl-d14	20.09	244	715689	38.12	ng	0.00
Target Compounds						
6) Aniline	7.46	93	25497	6.014	ng	99
10) Phenol	7.33	94	23794	7.112	ng	96
13) 1,4-Dichlorobenzene	8.18	146	13219	4.213	ng	89
14) 1,2-Dichlorobenzene	8.49	146	7252	2.408	ng	94
24) Nitrobenzene	9.35	77	98944	32.004	ng	92
30) 1,2,4-Trichlorobenzene	10.82	180	109963	32.060	ng	98
49) Dimethylphthalate	14.21	163	187614	16.238	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA G\DATA\BG111117\
Data File : BG030935.D
Acq On : 11 Nov 2017 9:37
Operator : SJ/JU
Sample : I6413-01
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
FM728

Manual Integrations
APPROVED

SOHIL
11/13/2017 4:55:20 PM

Quant Time: Nov 13 03:48:46 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG111017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 10 15:38:25 2017
Response via : Initial Calibration

