

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG111117\
 Data File : BG030957.D
 Acq On : 12 Nov 2017 00:00
 Operator : SJ/JU
 Sample : I6443-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 FADW4870L

Manual Integrations
 APPROVED

SOHIL
 11/13/2017 4:55:29 PM

Quant Time: Nov 13 06:06:54 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG111017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 10 15:38:25 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	43461	20.00	ng	0.00
21) Naphthalene-d8	10.96	136	197501	20.00	ng	0.00
38) Acenaphthene-d10	14.76	164	141258	20.00	ng	-0.02
63) Phenanthrene-d10	17.50	188	345874	20.00	ng	-0.02
75) Chrysene-d12	21.77	240	399518	20.00	ng	0.00
86) Perylene-d12	25.08	264	410787	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.70	112	138349	54.59	ng	0.00
7) Phenol-d6	7.30	99	116931	34.24	ng	0.00
23) Nitrobenzene-d5	9.31	82	285015	85.51	ng	-0.02
41) 2,4,6-Tribromophenol	16.25	330	310302	135.79	ng	0.00
44) 2-Fluorobiphenyl	13.38	172	904123	91.97	ng	-0.02
78) Terphenyl-d14	20.09	244	1464106	80.92	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.55	88	4518	4.393	ng	Qvalue # 78
10) Phenol	7.33	94	10552	2.976	ng	85
32) Benzoic acid	10.23	122	9866m	9.627	ng	
49) Dimethylphthalate	14.20	163	112828	9.238	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA G\DATA\BG111117\
Data File : BG030957.D
Acq On : 12 Nov 2017 00:00
Operator : SJ/JU
Sample : I6443-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
FADW4870L

Manual Integrations
APPROVED

SOHIL
11/13/2017 4:55:29 PM

Quant Time: Nov 13 06:06:54 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG111017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 10 15:38:25 2017
Response via : Initial Calibration

