




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111120\
Data File : BG047282.D
Acq On : 11 Nov 2020 13:14
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 12 05:20:00 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG110420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 04 16:19:40 2020
Response via : Initial Calibration



TIC: BG047282.D\data.ms

(70) Pentachlorophenol (C)




17.045min (+ 0.009) 140306.57 ng

response 88178

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.73
264.00	61.60	60.48
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111120\
Data File : BG047282.D
Acq On : 11 Nov 2020 13:14
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 12 05:20:00 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG110420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 04 16:19:40 2020
Response via : Initial Calibration



TIC: BG047282.D\data.ms

(77) Benzidine**19.624min (+ 0.009) 74015.92 ng****response 503925**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	15.26
183.00	13.20	13.25
0.00	0.00	0.00