

Data Path : Z:\HPCHEM1\BNA G\DATA\BG111215\
 Data File : BG019584.D
 Acq On : 12 Nov 2015 15:11
 Operator : UM/NP
 Sample : G4308-21
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 D9KS5

Quant Time: Nov 13 10:15:06 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 13 10:06:43 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.18	152	19611	20.00	ng/ul	0.00
18) Naphthalene-d8	11.00	136	84163	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.81	164	69684	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.55	188	195242	20.00	ng/ul	0.00
78) Chrysene-d12	21.82	240	242381	20.00	ng/ul	0.00
86) Perylene-d12	25.17	264	225978	20.00	ng/ul	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.48	96	679	1.66	ng/uL	0.00
5) Phenol-d5	7.32	99	18673	10.37	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.49	67	57846	50.69	ng/ul	0.00
9) 2-Chlorophenol-d4	7.70	132	41903	37.29	ng/ul	0.00
13) 4-Methylphenol-d8	8.87	113	35742	24.93	ng/ul	0.00
19) Nitrobenzene-d5	9.35	128	29324	47.41	ng/ul	0.00
22) 2-Nitrophenol-d4	10.07	143	33640	45.81	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.61	165	67063	42.10	ng/ul	0.00
29) 4-Chloroaniline-d4	0.00	131	0	0.00	ng/ul	
44) Dimethylphthalate-d6	14.20	166	274392	47.76	ng/ul	0.00
47) Acenaphthylene-d8	14.50	160	288787	45.39	ng/ul	0.00
52) 4-Nitrophenol-d4	14.99	143	7434	8.14	ng/ul	0.00
58) Fluorene-d10	15.80	176	238169	44.81	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.90	200	47785	37.81	ng/ul	0.00
71) Anthracene-d10	17.65	188	405270	45.54	ng/ul	0.00
79) Pyrene-d10	19.92	212	478629	45.12	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.94	264	517766	45.66	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG111215\
Data File : BG019584.D
Acq On : 12 Nov 2015 15:11
Operator : UM/NP
Sample : G4308-21
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
D9KS5

Quant Time: Nov 13 10:15:06 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102815.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 13 10:06:43 2015
Response via : Initial Calibration

