

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111321\
 Data File : BG051029.D
 Acq On : 13 Nov 2021 18:48
 Operator : CG/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC050

Quant Time: Nov 15 03:17:38 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 03:13:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.224	152	37653	20.00	ng	0.00
21) Naphthalene-d8	11.050	136	164560	20.00	ng	# 0.00
39) Acenaphthene-d10	14.851	164	109488	20.00	ng	0.00
64) Phenanthrene-d10	17.601	188	253431	20.00	ng	# 0.00
76) Chrysene-d12	21.896	240	210208	20.00	ng	# 0.00
86) Perylene-d12	25.298	264	217426	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.762	112	242348	96.26	ng	0.00
7) Phenol-d6	7.372	99	347890	95.44	ng	0.00
23) Nitrobenzene-d5	9.405	82	359205	90.36	ng	0.00
42) 2,4,6-Tribromophenol	16.338	330	111743	107.00	ng	0.00
45) 2-Fluorobiphenyl	13.476	172	682655	93.84	ng	0.00
79) Terphenyl-d14	20.186	244	1086648	100.74	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.611	88	53558	41.07	ng	91
3) Pyridine	4.029	79	166895	46.85	ng	# 91
4) n-Nitrosodimethylamine	3.940	42	72395	43.13	ng	# 36
6) Aniline	7.548	93	207628	49.03	ng	94
8) 2-Chlorophenol	7.789	128	124628	51.41	ng	88
9) Benzaldehyde	7.360	77	120863	44.93	ng	90
10) Phenol	7.401	94	176990	49.94	ng	83
11) bis(2-Chloroethyl)ether	7.636	93	133305	48.19	ng	88
12) 1,3-Dichlorobenzene	8.112	146	134926	48.22	ng	94
13) 1,4-Dichlorobenzene	8.259	146	137640	48.79	ng	97
14) 1,2-Dichlorobenzene	8.582	146	130862	48.56	ng	95
15) Benzyl Alcohol	8.465	79	142210	49.37	ng	91
16) 2,2'-oxybis(1-Chloropr...	8.747	45	208058	46.12	ng	87
17) 2-Methylphenol	8.664	107	122032	52.33	ng	# 91
18) Hexachloroethane	9.316	117	51172	47.93	ng	93
19) n-Nitroso-di-n-propyla...	9.034	70	128290	48.29	ng	# 87
20) 3+4-Methylphenols	8.993	107	171353	52.20	ng	93
22) Acetophenone	9.058	105	218143	46.63	ng	92
24) Nitrobenzene	9.446	77	175369	44.37	ng	92
25) Isophorone	9.969	82	326792	46.35	ng	# 95
26) 2-Nitrophenol	10.157	139	76825	52.93	ng	# 87
27) 2,4-Dimethylphenol	10.204	122	119211	47.54	ng	92
28) bis(2-Chloroethoxy)met...	10.439	93	189239	46.94	ng	93
29) 2,4-Dichlorophenol	10.697	162	125511	49.11	ng	97
30) 1,2,4-Trichlorobenzene	10.909	180	138538	47.72	ng	94
31) Naphthalene	11.103	128	418758	47.54	ng	98
32) Benzoic acid	10.356	122	99544	53.49	ng	# 82
33) 4-Chloroaniline	11.208	127	182998	49.52	ng	97
34) Hexachlorobutadiene	11.373	225	84071	46.13	ng	94
35) Caprolactam	12.007	113	35572	36.37	ng	# 71
36) 4-Chloro-3-methylphenol	12.313	107	155912	48.81	ng	# 92
37) 2-Methylnaphthalene	12.695	142	290942	47.87	ng	92
38) 1-Methylnaphthalene	12.912	142	285483	48.44	ng	91
40) 1,2,4,5-Tetrachloroben...	13.053	216	155836	48.20	ng	98
41) Hexachlorocyclopentadiene	13.024	237	61513	32.92	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111321\
 Data File : BG051029.D
 Acq On : 13 Nov 2021 18:48
 Operator : CG/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC050

Quant Time: Nov 15 03:17:38 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 03:13:38 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	13.294	196	115079	50.51	ng	94
44) 2,4,5-Trichlorophenol	13.365	196	122072	51.25	ng	93
46) 1,1'-Biphenyl	13.688	154	384861	48.19	ng	92
47) 2-Chloronaphthalene	13.735	162	300752	49.05	ng	98
48) 2-Nitroaniline	13.940	65	119837	49.13	ng	90
49) Acenaphthylene	14.581	152	489974	48.77	ng	97
50) Dimethylphthalate	14.299	163	416322	50.31	ng	99
51) 2,6-Dinitrotoluene	14.428	165	85655	51.67	ng	84
52) Acenaphthene	14.916	154	289311	44.81	ng	92
53) 3-Nitroaniline	14.763	138	105697	53.91	ng	98
54) 2,4-Dinitrophenol	14.969	184	55237	53.72	ng	84
55) Dibenzofuran	15.251	168	486592	48.73	ng	97
56) 4-Nitrophenol	15.057	139	84940	53.85	ng	# 70
57) 2,4-Dinitrotoluene	15.215	165	125658	53.78	ng	93
58) Fluorene	15.897	166	381201	49.70	ng	97
59) 2,3,4,6-Tetrachlorophenol	15.392	232	102778	49.53	ng	94
60) Diethylphthalate	15.650	149	439327	50.52	ng	98
61) 4-Chlorophenyl-phenyle...	15.879	204	209001	49.80	ng	98
62) 4-Nitroaniline	15.921	138	108616	53.24	ng	# 63
63) Azobenzene	16.173	77	464958	46.19	ng	81
65) 4,6-Dinitro-2-methylph...	15.973	198	83865	55.09	ng	90
66) n-Nitrosodiphenylamine	16.097	169	348021	48.33	ng	97
67) 4-Bromophenyl-phenylether	16.778	248	128426	50.29	ng	97
68) Hexachlorobenzene	16.902	284	126344	49.78	ng	94
69) Atrazine	17.037	200	84668	29.85	ng	92
70) Pentachlorophenol	17.243	266	79991	46.30	ng	96
71) Phenanthrene	17.642	178	641058	48.00	ng	99
72) Anthracene	17.736	178	637327	48.02	ng	99
73) Carbazole	18.000	167	632909	47.85	ng	98
74) Di-n-butylphthalate	18.529	149	747825	48.18	ng	# 97
75) Fluoranthene	19.640	202	758966	46.23	ng	95
77) Benzidine	19.816	184	165056	24.23	ng	98
78) Pyrene	20.004	202	755785	50.68	ng	97
80) Butylbenzylphthalate	20.868	149	334190	52.97	ng	97
81) Benzo(a)anthracene	21.878	228	698280	50.20	ng	98
82) 3,3'-Dichlorobenzidine	21.784	252	324849	70.47	ng	# 96
83) Chrysene	21.949	228	661993	50.60	ng	95
84) Bis(2-ethylhexyl)phtha...	21.743	149	465628	51.95	ng	# 98
85) Di-n-octyl phthalate	23.012	149	784912	52.50	ng	99
87) Indeno(1,2,3-cd)pyrene	29.223	276	751624	49.63	ng	# 88
88) Benzo(b)fluoranthene	24.211	252	671013	50.39	ng	# 92
89) Benzo(k)fluoranthene	24.287	252	641272	48.95	ng	# 95
90) Benzo(a)pyrene	25.139	252	568803	50.24	ng	# 96
91) Dibenzo(a,h)anthracene	29.293	278	617302	49.28	ng	# 94
92) Benzo(g,h,i)perylene	30.456	276	613683	50.02	ng	# 88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111321\
 Data File : BG051029.D
 Acq On : 13 Nov 2021 18:48
 Operator : CG/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC050

Quant Time: Nov 15 03:17:38 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 03:13:38 2021
 Response via : Initial Calibration





1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 8.224 min Scan# 1

Delta R.T. -0.005 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDIC050





Tgt Ion:152 Resp: 37653

Ion Ratio Lower Upper

152 100

150 152.3 119.4 179.0

115 62.9 60.2 90.4



1,4-Dioxane

Concen: 41.07 ng

RT: 3.611 min Scan# 21

Delta R.T. -0.005 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48




Tgt Ion: 88 Resp: 53558


Ion Ratio Lower Upper

88 100

58 77.2 55.7 83.5

43 31.0 29.3 43.9









Pyridine
Concen: 46.85 ng
RT: 4.029 min Scan# 93
Delta R.T. -0.005 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48




Instrument :
BNA_G
ClientSampleId :
SSTDICC050


Tgt Ion: 79 Resp: 166895
Ion Ratio Lower Upper
79 100
52 68.7 58.4 87.6
51 35.5 37.0 55.6#



n-Nitrosodimethylamine
Concen: 43.13 ng
RT: 3.940 min Scan# 77
Delta R.T. 0.001 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48

Tgt Ion: 42 Resp: 72395
Ion Ratio Lower Upper
42 100
74 138.7 64.0 96.0#
44 6.6 2.7 4.1#





2-Fluorophenol

Concen: 96.26 ng

RT: 5.762 min Scan#

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050





Tgt Ion: 112 Resp: 242348

Ion Ratio Lower Upper

112 100

64 66.4 46.7 70.1

63 35.7 35.3 52.9



Aniline

Concen: 49.03 ng

RT: 7.548 min Scan# 691

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48




Tgt Ion: 93 Resp: 207628


Ion Ratio Lower Upper

93 100

66 41.3 30.6 45.8

65 22.0 15.0 22.4









Phenol-d6
 Concen: 95.44 ng
 RT: 7.372 min Scan# 661
 Delta R.T. 0.001 min
 Lab File: BG051029.D
 Acq: 13 Nov 2021 18:48




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC050


Tgt Ion: 99 Resp: 347890
 Ion Ratio Lower Upper
 99 100
 42 19.3 20.5 30.7#
 71 38.9 41.5 62.3#



2-Chlorophenol
 Concen: 51.41 ng
 RT: 7.789 min Scan# 732
 Delta R.T. 0.001 min
 Lab File: BG051029.D
 Acq: 13 Nov 2021 18:48

Tgt Ion: 128 Resp: 124628
 Ion Ratio Lower Upper
 128 100
 130 34.0 10.0 50.0
 64 58.1 28.1 68.1





Benzaldehyde

Concen: 44.93 ng

RT: 7.360 min Scan# 659

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050





Tgt Ion: 77 Resp: 120863

Ion Ratio Lower Upper

77 100

105 88.6 68.8 108.8

106 85.6 47.9 87.9



Phenol

Concen: 49.94 ng

RT: 7.401 min Scan# 666

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48




Tgt Ion: 94 Resp: 176990


Ion Ratio Lower Upper

94 100

65 30.8 12.9 52.9

66 40.6 39.4 79.4









bis(2-Chloroethyl)ether
 Concen: 48.19 ng
 RT: 7.636 min Scan#
 Delta R.T. 0.001 min
 Lab File: BG051029.D
 Acq: 13 Nov 2021 18:48




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC050

Tgt Ion:	93	Resp:	133305
Ion	Ratio	Lower	Upper
93	100		
63	84.2	50.4	90.4
95	33.1	15.3	55.3



1,3-Dichlorobenzene
 Concen: 48.22 ng
 RT: 8.112 min Scan# 787
 Delta R.T. 0.001 min
 Lab File: BG051029.D
 Acq: 13 Nov 2021 18:48

Tgt Ion:	146	Resp:	134926
Ion	Ratio	Lower	Upper
146	100		
148	62.8	52.6	78.8
75	37.9	35.7	53.5



Abundance Scan 813 (8.264 min): BG051028.D\data.ms (-804) #13

1,4-Dichlorobenzene

Concen: 48.79 ng

RT: 8.259 min Scan# 813

Delta R.T. -0.005 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Tgt Ion:146 Resp: 137640



Ion Ratio Lower Upper

146 100


148 63.2 54.0 81.0

111 42.6 34.3 51.5

Abundance Scan 812 (8.259 min): BG051029.D\data.ms



Abundance Scan 812 (8.259 min): BG051029.D\data.ms (-723) (-)



Abundance Scan 867 (8.581 min): BG051028.D\data.ms (-859) #14

1,2-Dichlorobenzene

Concen: 48.56 ng

RT: 8.582 min Scan# 867

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Tgt Ion:146 Resp: 130862



Ion Ratio Lower Upper

146 100


148 62.7 47.5 71.3

111 45.0 39.1 58.7

Abundance Scan 867 (8.582 min): BG051029.D\data.ms



Abundance Scan 867 (8.582 min): BG051029.D\data.ms (-777) (-)



Abundance Scan 847 (8.464 min): BG051028.D\data.ms (-838) #15

Benzyl Alcohol

Concen: 49.37 ng

RT: 8.465 min Scan# 847

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Tgt Ion: 79 Resp: 142210



Ion Ratio Lower Upper

79 100


108 70.4 48.2 72.4

77 67.1 57.3 85.9

Abundance Scan 847 (8.465 min): BG051029.D\data.ms



Abundance Scan 847 (8.465 min): BG051029.D\data.ms (-757) (-)



Abundance Scan 895 (8.746 min): BG051028.D\data.ms (-886) #16

2,2'-oxybis(1-Chloropropane)

Concen: 46.12 ng

RT: 8.747 min Scan# 895

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Tgt Ion: 45 Resp: 208058



Ion Ratio Lower Upper

45 100


77 14.8 2.2 42.2

79 10.6 0.0 34.3

Abundance Scan 895 (8.747 min): BG051029.D\data.ms



Abundance Scan 895 (8.747 min): BG051029.D\data.ms (-805) (-)



Abundance Scan 881 (8.663 min): BG051028.D\data.ms (-872) #17

2-Methylphenol

Concen: 52.33 ng

RT: 8.664 min Scan# 881

Delta R.T. 0.001 min

Lab File: BG051029.D


Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050



Tgt Ion:107 Resp: 122032




Ion Ratio Lower Upper

107 100

108 106.6 84.6 127.0

77 52.2 50.7 76.1

79 48.4 49.4 74.0#



Abundance Scan 992 (9.315 min): BG051028.D\data.ms (-984) #18

Hexachloroethane

Concen: 47.93 ng

RT: 9.316 min Scan# 992

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48




Tgt Ion:117 Resp: 51172


Ion Ratio Lower Upper

117 100

119 95.1 74.4 111.6

201 99.4 88.6 133.0









n-Nitroso-di-n-propylamine
 Concen: 48.29 ng
 RT: 9.034 min Scan# 934
 Delta R.T. 0.001 min
 Lab File: BG051029.D
 Acq: 13 Nov 2021 18:48




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC050

Tgt Ion:	70	Resp:	128290
Ion	Ratio	Lower	Upper
70	100		
42	55.4	54.5	81.7
101	8.4	9.0	13.6#
130	17.2	17.0	25.6



3+4-Methylphenols
 Concen: 52.20 ng
 RT: 8.993 min Scan# 937
 Delta R.T. 0.001 min
 Lab File: BG051029.D
 Acq: 13 Nov 2021 18:48

Tgt Ion:	107	Resp:	171353
Ion	Ratio	Lower	Upper
107	100		
108	86.8	69.3	109.3
77	35.6	25.2	65.2
79	31.3	15.1	55.1



Abundance Scan 1288 (11.055 min): BG051028.D\data.ms (-1288) (-)

Naphthalene-d8

Concen: 20.00 ng

RT: 11.050 min Scan# 1288

Delta R.T. -0.005 min

Lab File: BG051029.D




Acq: 13 Nov 2021 18:48

Instrument :


BNA_G

ClientSampleId :

SSTDICC050



Tgt Ion:	136	Resp:	164560
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.2	12.2
54	9.1	5.5	8.3#
68	6.6	2.3	3.5#



Abundance Scan 948 (9.057 min): BG051028.D\data.ms (-940) (-)

Acetophenone




Concen: 46.63 ng

RT: 9.058 min Scan# 948


Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48



Tgt Ion:	105	Resp:	218143
Ion	Ratio	Lower	Upper
105	100		
71	0.7	0.6	0.8
51	26.8	26.6	40.0
120	21.2	16.2	24.4



Abundance Scan 1007 (9.404 min): BG051028.D\data.ms (-99723)

Nitrobenzene-d5

Concen: 90.36 ng

RT: 9.405 min Scan#

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Tgt Ion: 82 Resp: 359205


Ion Ratio Lower Upper

82 100

128 38.4 29.8 44.6


54 49.6 45.8 68.8

Abundance Scan 1007 (9.405 min): BG051029.D\data.ms




m/z-->

Abundance Scan 1007 (9.405 min): BG051029.D\data.ms (-917) (-)



m/z-->



Abundance Scan 1014 (9.445 min): BG051028.D\data.ms (-10624)

Nitrobenzene

Concen: 44.37 ng

RT: 9.446 min Scan# 1014

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Tgt Ion: 77 Resp: 175369


Ion Ratio Lower Upper

77 100

123 39.7 27.5 41.3


65 13.4 12.6 18.8

Abundance Scan 1014 (9.446 min): BG051029.D\data.ms




m/z-->


Abundance Scan 1014 (9.446 min): BG051029.D\data.ms (-924) (-)



m/z-->



Abundance Scan 1103 (9.968 min): BG051028.D\data.ms (-1025)



Isophorone

Concen: 46.35 ng

RT: 9.969 min Scan#

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050



Tgt Ion: 82 Resp: 326792

Ion Ratio Lower Upper


82 100

95 6.9 8.5 12.7#


138 15.0 11.8 17.6



Abundance Scan 1103 (9.969 min): BG051029.D\data.ms (-1013) (-)



Abundance Scan 1136 (10.162 min): BG051028.D\data.ms (-1126) (-)



2-Nitrophenol

Concen: 52.93 ng

RT: 10.157 min Scan# 1135

Delta R.T. -0.005 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48



Tgt Ion: 139 Resp: 76825

Ion Ratio Lower Upper


139 100

109 31.9 37.8 56.6#

65 60.2 51.8 77.8



Abundance Scan 1135 (10.157 min): BG051029.D\data.ms (-1046) (-)



Abundance Scan 1143 (10.203 min): BG051028.D\data.ms (-1#29) (-)

2,4-Dimethylphenol

Concen: 47.54 ng

RT: 10.204 min Scan# 1143

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Tgt Ion:122 Resp: 119211



Ion Ratio Lower Upper

122 100


107 124.7 110.1 165.1

121 58.6 49.0 73.6

Abundance Scan 1143 (10.204 min): BG051029.D\data.ms



Abundance Scan 1143 (10.204 min): BG051029.D\data.ms (-1053) (-)



Abundance Scan 1183 (10.438 min): BG051028.D\data.ms (-1#28) (-)

bis(2-Chloroethoxy)methane

Concen: 46.94 ng

RT: 10.439 min Scan# 1183

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Tgt Ion: 93 Resp: 189239



Ion Ratio Lower Upper

93 100


95 32.4 22.2 33.4

123 10.7 8.8 13.2

Abundance Scan 1183 (10.439 min): BG051029.D\data.ms



Abundance Scan 1183 (10.439 min): BG051029.D\data.ms (-1093) (-)



Abundance Scan 1227 (10.696 min): BG051028.D\data.ms (-1229) (-)

2,4-Dichlorophenol

Concen: 49.11 ng

RT: 10.697 min Scan# 1229

Delta R.T. 0.001 min

Lab File: BG051029.D




Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050




Tgt Ion:162 Resp: 125511

Ion Ratio Lower Upper

162 100

164 63.0 39.4 79.4

98 38.4 18.7 58.7



Abundance Scan 1264 (10.914 min): BG051028.D\data.ms (-1236) (-)

1,2,4-Trichlorobenzene




Concen: 47.72 ng

RT: 10.909 min Scan# 1263

Delta R.T. -0.005 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48




Tgt Ion:180 Resp: 138538


Ion Ratio Lower Upper

180 100

182 92.7 78.9 118.3

145 29.7 26.2 39.4









Scan 1296 (11.102 min): BG051028.D\data.ms (-1284)
 Naphthalene
 Concen: 47.54 ng
 RT: 11.103 min Scan# 1134
 Delta R.T. 0.001 min
 Lab File: BG051029.D
 Acq: 13 Nov 2021 18:48




Instrument : BNA_G
 ClientSampleId : BG051029.D
 SSTDICC050

Tgt Ion:	128	Resp:	418758
Ion	Ratio	Lower	Upper
128	100		
129	11.3	8.5	12.7
127	13.9	10.2	15.4



Scan 1168 (10.350 min): BG051028.D\data.ms (-1132)
 Benzoic acid
 Concen: 53.49 ng
 RT: 10.356 min Scan# 1169
 Delta R.T. 0.007 min
 Lab File: BG051029.D
 Acq: 13 Nov 2021 18:48

Tgt Ion:	122	Resp:	99544
Ion	Ratio	Lower	Upper
122	100		
105	127.6	92.3	132.3
77	108.2	66.5	106.5#



Abundance Scan 1315 (11.213 min): BG051028.D\data.ms (-1303)

4-Chloroaniline

Concen: 49.52 ng

RT: 11.208 min Scan# 1314

Delta R.T. -0.005 min

Lab File: BG051029.D


Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050



Tgt Ion:127 Resp: 182998




Ion Ratio Lower Upper

127 100

129 33.3 26.7 40.1

65 36.9 32.8 49.2

92 21.2 16.2 24.4



Abundance Scan 1342 (11.372 min): BG051028.D\data.ms (-1334)

Hexachlorobutadiene

Concen: 46.13 ng

RT: 11.373 min Scan# 1342

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48




Tgt Ion:225 Resp: 84071

Ion Ratio Lower Upper

225 100

223 65.1 49.8 74.8

227 64.7 46.4 69.6



Abundance Scan 1450 (12.006 min): BG051028.D\data.ms (-1435) (-)

Caprolactam

Concen: 36.37 ng

RT: 12.007 min Scan# 1450

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Tgt Ion:113 Resp: 35572



Ion Ratio Lower Upper

113 100


55 185.5 113.8 153.8#

56 142.0 105.6 145.6

Abundance Scan 1450 (12.007 min): BG051029.D\data.ms



Abundance Scan 1450 (12.007 min): BG051029.D\data.ms (-1360) (-)



Abundance Scan 1502 (12.312 min): BG051028.D\data.ms (-1436) (-)

4-Chloro-3-methylphenol

Concen: 48.81 ng

RT: 12.313 min Scan# 1502

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Tgt Ion:107 Resp: 155912



Ion Ratio Lower Upper

107 100


144 24.0 24.8 37.2#

142 72.8 62.2 93.4

Abundance Scan 1502 (12.313 min): BG051029.D\data.ms



Abundance Scan 1502 (12.313 min): BG051029.D\data.ms (-1412) (-)



Abundance Scan 1567 (12.694 min): BG051028.D\data.ms (-1537) (-)

2-Methylnaphthalene

Concen: 47.87 ng

RT: 12.695 min Scan# 1567

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Tgt Ion:142 Resp: 290942

Ion Ratio Lower Upper

142 100

141 94.1 68.8 103.2

115 40.4 36.4 54.6

Abundance Scan 1567 (12.695 min): BG051029.D\data.ms

Ref

Raw

Sub

m/z-->

Abundance Scan 1567 (12.695 min): BG051029.D\data.ms (-1477) (-)

Ref

Raw

Sub

m/z-->

Abundance

12.695

150000

100000

50000

0

Time-->

Abundance Scan 1604 (12.911 min): BG051028.D\data.ms (-1538) (-)

1-Methylnaphthalene

Concen: 48.44 ng

RT: 12.912 min Scan# 1604

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Tgt Ion:142 Resp: 285483

Ion Ratio Lower Upper

142 100

141 96.7 70.6 106.0

116 4.3 4.0 6.0

Abundance Scan 1604 (12.912 min): BG051029.D\data.ms

Ref

Raw

Sub

m/z-->

Abundance Scan 1604 (12.912 min): BG051029.D\data.ms (-1519) (-)

Ref

Raw

Sub

m/z-->

Abundance

12.912

150000

100000

50000

0

Time-->

Abundance Scan 1935 (14.856 min): BG051028.D\data.ms (-1629) (-)

Acenaphthene-d10

Concen: 20.00 ng

RT: 14.851 min Scan# 1629

Delta R.T. -0.005 min

Lab File: BG051029.D




Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050




Tgt Ion:164 Resp: 109488

Ion Ratio Lower Upper

164 100

162 104.3 81.4 122.0

160 44.9 34.9 52.3



Abundance Scan 1628 (13.052 min): BG051028.D\data.ms (-1620) (-)

1,2,4,5-Tetrachlorobenzene




Concen: 48.20 ng

RT: 13.053 min Scan# 1628

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48



Tgt Ion:216 Resp: 155836


Ion Ratio Lower Upper

216 100

214 78.6 61.4 92.2

179 20.9 17.3 25.9

108 18.9 16.3 24.5



Abundance Scan 1623 (13.023 min): BG051028.D\data.ms (-1614) (-)

Hexachlorocyclopentadiene

Concen: 32.92 ng

RT: 13.024 min Scan# 1614

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Tgt Ion:237 Resp: 61513

Ion Ratio Lower Upper

237 100

235 57.8 36.6 76.6

272 12.0 0.0 32.5

Abundance Scan 1623 (13.024 min): BG051029.D\data.ms

Raw

m/z-->

Abundance Scan 1623 (13.024 min): BG051029.D\data.ms (-1533) (-)

Sub

m/z-->

Abundance

13.024

40000

30000

20000

10000

0

Time-->

Abundance Scan 2188 (16.343 min): BG051028.D\data.ms (-2180) (-)

2,4,6-Tribromophenol

Concen: 107.00 ng

RT: 16.338 min Scan# 2187

Delta R.T. -0.005 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Tgt Ion:330 Resp: 111743

Ion Ratio Lower Upper

330 100

332 94.6 77.3 115.9

141 49.5 35.6 53.4

Abundance Scan 2187 (16.338 min): BG051029.D\data.ms

Raw

m/z-->

Abundance Scan 2187 (16.338 min): BG051029.D\data.ms (-2098) (-)

Sub

m/z-->

Abundance

16.338

60000


40000

20000

0

Time-->




Abundance Scan 1668 (13.287 min): BG051028.D\data.ms (-1668) (-)




2,4,6-Trichlorophenol
Concen: 50.51 ng
RT: 13.294 min Scan# 1668
Delta R.T. 0.007 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48

Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Tgt Ion:	196	Resp:	115079
Ion	Ratio	Lower	Upper
196	100		
198	95.6	82.1	123.1
200	30.5	23.8	35.8






Abundance Scan 1681 (13.364 min): BG051028.D\data.ms (-1681) (-)




2,4,5-Trichlorophenol
Concen: 51.25 ng
RT: 13.365 min Scan# 1681
Delta R.T. 0.001 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48

Tgt Ion:	196	Resp:	122072
Ion	Ratio	Lower	Upper
196	100		
198	96.8	87.4	131.0
97	52.8	43.3	64.9
132	27.3	21.4	32.2





Abundance Scan 1700 (13.475 min): BG051028.D\data.ms (-1695) (-)




2-Fluorobiphenyl
Concen: 93.84 ng
RT: 13.476 min Scan# 1700
Delta R.T. 0.001 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48

Instrument :
BNA_G
ClientSampleId :
SSTDICC050


Tgt Ion:	172	Resp:	682655
Ion	Ratio	Lower	Upper
172	100		
171	38.2	28.6	43.0
170	25.6	18.6	27.8



Abundance Scan 1700 (13.476 min): BG051029.D\data.ms (-1610) (-)





Abundance Scan 1736 (13.687 min): BG051028.D\data.ms (-1746) (-)




1,1'-Biphenyl
Concen: 48.19 ng
RT: 13.688 min Scan# 1736
Delta R.T. 0.001 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48

Tgt Ion:	154	Resp:	384861
Ion	Ratio	Lower	Upper
154	100		
153	43.5	20.5	60.5
76	16.6	0.0	29.3



Abundance Scan 1736 (13.688 min): BG051029.D\data.ms (-1646) (-)



Abundance Scan 1744 (13.734 min): BG051028.D\data.ms (-1744) (-)

2-Chloronaphthalene

Concen: 49.05 ng

RT: 13.735 min Scan# 1744

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Tgt Ion:162 Resp: 300752

Ion Ratio Lower Upper

162 100

127 41.5 32.7 49.1

164 32.0 27.0 40.4

Abundance Scan 1744 (13.735 min): BG051029.D\data.ms

Ref

50

162.0

127.0

81.0

63.0

39.0

101.0

m/z-->

Abundance Scan 1744 (13.735 min): BG051029.D\data.ms (-1654) (-)

Sub

50

162.0

127.0

81.0

63.0

39.0

101.0

m/z-->

Abundance

150000

100000

50000

0

13.735

13.70

13.80

Time-->

Abundance Scan 1779 (13.939 min): BG051028.D\data.ms (-1779) (-)

2-Nitroaniline

Concen: 49.13 ng

RT: 13.940 min Scan# 1779

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Tgt Ion: 65 Resp: 119837

Ion Ratio Lower Upper

65 100

92 67.2 47.2 70.8

138 94.3 68.0 102.0

Abundance Scan 1779 (13.940 min): BG051029.D\data.ms

Ref

50

138.0

92.0

65.0

39.0

108.0

123.1

m/z-->

Abundance Scan 1779 (13.940 min): BG051029.D\data.ms (-1689) (-)

Sub

50

138.0

92.0

65.0

39.0

108.0

123.1

m/z-->

Abundance

60000

40000

20000

0

13.940

13.90

14.00

Time-->

Abundance Scan 1888 (14.580 min): BG051028.D\data.ms (-1749) (-)

Acenaphthylene

Concen: 48.77 ng

RT: 14.581 min Scan# 1888

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Tgt Ion:152 Resp: 489974

Ion Ratio Lower Upper

152 100

151 21.2 18.2 27.4

153 13.6 11.7 17.5

Abundance Scan 1888 (14.581 min): BG051029.D\data.ms

Ref 50

152.0

Raw 50

50.0 76.0 98.0 126.0

m/z-->

Abundance Scan 1888 (14.581 min): BG051029.D\data.ms (-1798) (-)

Sub 50

152.0

50.0 76.0 98.0 126.0

m/z-->

Abundance

14.581

200000

100000

0

Time-->

Abundance Scan 1840 (14.298 min): BG051028.D\data.ms (-1750) (-)

Dimethylphthalate

Concen: 50.31 ng

RT: 14.299 min Scan# 1840

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Tgt Ion:163 Resp: 416322

Ion Ratio Lower Upper

163 100

194 3.7 3.0 4.6

164 10.1 8.6 12.8

Abundance Scan 1840 (14.299 min): BG051029.D\data.ms

Ref 50

163.0

Raw 50

50.0 77.0 104.0 133.0 194.0

m/z-->

Abundance Scan 1840 (14.299 min): BG051029.D\data.ms (-1750) (-)

Sub 50

163.0

50.0 77.0 104.0 133.0 194.0

m/z-->

Abundance

14.299

250000

200000

150000

100000

50000

0

Time-->

Abundance Scan 1862 (14.427 min): BG051028.D\data.ms (-1854) (-)

2,6-Dinitrotoluene

Concen: 51.67 ng

RT: 14.428 min Scan# 1862

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Tgt Ion:165 Resp: 85655

Ion Ratio Lower Upper

165 100

63 64.3 63.3 94.9

89 66.7 63.4 95.0

Abundance Scan 1862 (14.428 min): BG051029.D\data.ms

Ref

63.0 89.0 165.0

39.0 121.0 207.0

m/z-->

Abundance Scan 1862 (14.428 min): BG051029.D\data.ms (-1772) (-)

Ref

63.0 89.0 165.0

39.0 121.0 207.0

m/z-->

Abundance Scan 1862 (14.428 min): BG051029.D\data.ms (-1772) (-)

Ref

63.0 89.0 165.0

39.0 121.0 207.0

m/z-->

Abundance Scan 1862 (14.428 min): BG051029.D\data.ms (-1772) (-)

Ref

63.0 89.0 165.0

39.0 121.0 207.0

m/z-->

Abundance

50000

40000

30000

20000

10000

0

14.40

Time-->

Abundance Scan 1945 (14.915 min): BG051028.D\data.ms (-1950) (-)

Acenaphthene

Concen: 44.81 ng

RT: 14.916 min Scan# 1945

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Tgt Ion:154 Resp: 289311

Ion Ratio Lower Upper

154 100

153 115.7 87.0 130.4

152 57.2 40.1 60.1

Abundance Scan 1945 (14.916 min): BG051029.D\data.ms

Ref

76.0 153.0

51.0 98.0 126.0

m/z-->

Abundance Scan 1945 (14.916 min): BG051029.D\data.ms

Ref

76.0 153.0

51.0 98.0 126.0

m/z-->

Abundance Scan 1945 (14.916 min): BG051029.D\data.ms (-1855) (-)

Ref

76.0 153.0

51.0 98.0 126.0

m/z-->

Abundance

200000

150000

100000

50000

0

14.85 14.90 14.95

Time-->

Abundance Scan 1918 (14.756 min): BG051028.D\data.ms (-1958) (-)

3-Nitroaniline

Concen: 53.91 ng

RT: 14.763 min Scan# 1918

Delta R.T. 0.007 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Tgt Ion:138 Resp: 105697



Ion Ratio Lower Upper

138 100


108 10.0 8.3 12.5

92 123.0 96.4 144.6

Abundance Scan 1919 (14.763 min): BG051029.D\data.ms



Abundance Scan 1919 (14.763 min): BG051029.D\data.ms (-1828) (-)



Abundance Scan 1954 (14.968 min): BG051028.D\data.ms (-1958) (-)

2,4-Dinitrophenol

Concen: 53.72 ng

RT: 14.969 min Scan# 1954

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Tgt Ion:184 Resp: 55237



Ion Ratio Lower Upper

184 100


63 70.3 68.3 102.5


154 64.0 62.6 93.8

Abundance Scan 1954 (14.969 min): BG051029.D\data.ms



Abundance Scan 1954 (14.969 min): BG051029.D\data.ms (-1864) (-)









Dibenzo[1,2-b:4,5-b']difuran
 Concen: 48.73 ng
 RT: 15.251 min Scan# 1955
 Delta R.T. 0.001 min
 Lab File: BG051029.D
 Acq: 13 Nov 2021 18:48




Instrument : BNA_G
 ClientSampleId : BG051029.D
 SSTDICC050


Tgt Ion:	168	Resp:	486592
Ion	Ratio	Lower	Upper
168	100		
139	40.3	33.4	50.2
169	14.1	12.3	18.5



4-Nitrophenol
 Concen: 53.85 ng
 RT: 15.057 min Scan# 1969
 Delta R.T. -0.005 min
 Lab File: BG051029.D
 Acq: 13 Nov 2021 18:48

Tgt Ion:	139	Resp:	84940
Ion	Ratio	Lower	Upper
139	100		
109	81.3	116.3	156.3#
65	109.6	105.0	145.0









2,4-Dinitrotoluene
Concen: 53.78 ng
RT: 15.215 min Scan# 1996
Delta R.T. 0.001 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48




Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Tgt Ion:	165	Resp:	125658
Ion	Ratio	Lower	Upper
165	100		
63	48.2	48.0	72.0
89	80.8	65.4	98.0
182	2.7	2.0	3.0



Fluorene
Concen: 49.70 ng
RT: 15.897 min Scan# 2112
Delta R.T. 0.001 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48

Tgt Ion:	166	Resp:	381201
Ion	Ratio	Lower	Upper
166	100		
165	100.9	77.7	116.5
167	13.3	10.8	16.2



Abundance Scan 2040 (15.473 min): BG051028.D\data.ms (-259) (-)

2,3,4,6-Tetrachlorophenol

Concen: 49.53 ng

RT: 15.392 min Scan# 2040

Delta R.T. -0.081 min

Lab File: BG051029.D




Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050



Tgt Ion:232 Resp: 102778


Ion Ratio Lower Upper

232 100

131 43.1 37.2 55.8

130 3.0 2.8 4.2

166 26.1 24.0 36.0



Abundance Scan 2070 (15.649 min): BG051028.D\data.ms (-260) (-)

Diethylphthalate




Concen: 50.52 ng

RT: 15.650 min Scan# 2070

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48




Tgt Ion:149 Resp: 439327

Ion Ratio Lower Upper

149 100

177 22.5 19.0 28.6

150 12.7 10.2 15.4



Abundance Scan 2109 (15.878 min): BG051028.D\data.ms (-2109) (-)

4-Chlorophenyl-phenylether

Concen: 49.80 ng

RT: 15.879 min Scan# 2109

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Ref

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Abundance Scan 2109 (15.879 min): BG051029.D\data.ms (-2019) (-)

Raw

Sub

Tgt Ion:204 Resp: 209001

Ion Ratio Lower Upper

204 100

206 33.0 24.6 37.0

141 58.9 47.8 71.8

Abundance

15.879

100000

50000

0

Time--> 15.80 15.90

Abundance Scan 2116 (15.919 min): BG051028.D\data.ms (-2116) (-)

4-Nitroaniline

Concen: 53.24 ng

RT: 15.921 min Scan# 2116

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Ref

Abundance Scan 2116 (15.921 min): BG051029.D\data.ms

Raw

Sub

Abundance Scan 2116 (15.921 min): BG051029.D\data.ms (-2026) (-)

Raw

Sub

Abundance Scan 2116 (15.921 min): BG051029.D\data.ms (-2026) (-)

Raw

Sub

Abundance Scan 2116 (15.921 min): BG051029.D\data.ms (-2026) (-)

Raw

Sub

Abundance Scan 2116 (15.921 min): BG051029.D\data.ms (-2026) (-)

Raw

Sub

Abundance Scan 2116 (15.921 min): BG051029.D\data.ms (-2026) (-)

Raw

Sub

Abundance Scan 2116 (15.921 min): BG051029.D\data.ms (-2026) (-)

Raw

Sub

Abundance Scan 2116 (15.921 min): BG051029.D\data.ms (-2026) (-)

Raw

Sub

Abundance Scan 2116 (15.921 min): BG051029.D\data.ms (-2026) (-)

Raw

Sub

Abundance Scan 2116 (15.921 min): BG051029.D\data.ms (-2026) (-)

Raw

Sub

Abundance Scan 2116 (15.921 min): BG051029.D\data.ms (-2026) (-)

Raw

Sub

Abundance Scan 2116 (15.921 min): BG051029.D\data.ms (-2026) (-)

Raw

Sub

Tgt Ion:138 Resp: 108616

Ion Ratio Lower Upper

138 100

92 59.5 32.3 72.3

108 93.8 132.4 172.4#

Abundance

15.921

40000

20000

0

Time--> 15.80 16.00

Abundance Scan 2159 (16.172 min): BG051028.D\data.ms (-2159) (-)

Azobenzene

Concen: 46.19 ng

RT: 16.173 min Scan# 2159

Delta R.T. 0.001 min

Lab File: BG051029.D


Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

Client SampleId :

SSTDICC050



Tgt Ion: 77 Resp: 464958




Ion Ratio Lower Upper

77 100

182 21.6 7.9 47.9

105 19.2 0.0 37.9

51 28.4 27.3 67.3



Abundance Scan 2402 (17.600 min): BG051028.D\data.ms (-2402) (-)

Phenanthrene-d10

Concen: 20.00 ng

RT: 17.601 min Scan# 2402

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48




Tgt Ion: 188 Resp: 253431

Ion Ratio Lower Upper

188 100

94 8.9 5.4 8.0#

80 10.9 3.7 5.5#



Abundance Scan 2125 (15.972 min): BG051028.D\data.ms (-2165) (-)

4,6-Dinitro-2-methylphenol

Concen: 55.09 ng

RT: 15.973 min Scan# 2125

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Tgt Ion:198 Resp: 83865

Ion Ratio Lower Upper

198 100

51 49.1 29.4 69.4

105 48.4 15.1 55.1

Abundance Scan 2125 (15.973 min): BG051029.D\data.ms

Ref

51.0 77.0 105.0 134.0 168.0 198.0

m/z-->

Raw

51.0 77.0 105.0 134.0 168.0 198.0

m/z-->

Abundance Scan 2125 (15.973 min): BG051029.D\data.ms (-2035) (-)

Sub

51.0 77.0 105.0 134.0 167.9 198.0

m/z-->

Abundance

15.973

50000

40000

30000

20000

10000

0

Time-->

Abundance Scan 2146 (16.096 min): BG051028.D\data.ms (-2166) (-)

n-Nitrosodiphenylamine

Concen: 48.33 ng

RT: 16.097 min Scan# 2146

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Tgt Ion:169 Resp: 348021

Ion Ratio Lower Upper

169 100

168 72.0 56.3 84.5

167 38.5 28.4 42.6

Abundance Scan 2146 (16.097 min): BG051029.D\data.ms

Ref

51.0 83.5 115.0 141.0 169.1 198.0

m/z-->

Raw

51.0 83.5 115.0 141.0 169.1 198.0

m/z-->

Abundance Scan 2146 (16.097 min): BG051029.D\data.ms (-2056) (-)

Sub

51.0 83.5 115.0 141.0 169.1 198.0

m/z-->

Abundance

16.097

200000

150000

100000

50000

0

Time-->

Abundance Scan 2262 (16.777 min): BG051028.D\data.ms (-2262) (-)

4-Bromophenyl-phenylether

Concen: 50.29 ng

RT: 16.778 min Scan# 2262

Delta R.T. 0.001 min

Lab File: BG051029.D


Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050





Tgt Ion:248 Resp: 128426

Ion Ratio Lower Upper


248 100

250 96.1 76.1 114.1

141 76.5 65.4 98.2



Abundance Scan 2262 (16.778 min): BG051029.D\data.ms (-2172) (-)



Abundance Scan 2283 (16.901 min): BG051028.D\data.ms (-2283) (-)

Hexachlorobenzene

Concen: 49.78 ng

RT: 16.902 min Scan# 2283

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48



Tgt Ion:284 Resp: 126344

Ion Ratio Lower Upper


284 100

142 37.8 27.0 40.6

249 32.3 28.2 42.2



Abundance Scan 2283 (16.902 min): BG051029.D\data.ms (-2193) (-)



Abundance Scan 2306 (17.036 min): BG051028.D\data.ms (-2269) (-)

Atrazine

Concen: 29.85 ng

RT: 17.037 min Scan# 2306

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050

Tgt Ion:200 Resp: 84668

Ion Ratio Lower Upper

200 100

173 27.7 3.7 43.7

215 51.8 26.9 66.9

Abundance Scan 2306 (17.037 min): BG051029.D\data.ms

Ref

58.1 92.6 132.0 164.1 200.1

m/z-->

Raw

58.1 92.5 132.0 164.1 200.1 281.1

m/z-->

Abundance Scan 2306 (17.037 min): BG051029.D\data.ms (-2216) (-)

Sub

58.1 92.5 132.0 164.1 200.1 281.1

m/z-->

Abundance

60000

40000

20000

0

17.00 17.10

Time-->

Abundance Scan 2341 (17.241 min): BG051028.D\data.ms (-2374) (-)

Pentachlorophenol

Concen: 46.30 ng

RT: 17.243 min Scan# 2341

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Tgt Ion:266 Resp: 79991

Ion Ratio Lower Upper

266 100

268 65.0 49.0 73.4

264 59.6 49.0 73.6

Abundance Scan 2341 (17.243 min): BG051029.D\data.ms

Ref

60.0 94.9 129.9 164.9 201.9 233.8 265.8

m/z-->

Raw

60.0 94.9 129.9 164.9 201.9 233.8 265.8

m/z-->

Abundance Scan 2341 (17.243 min): BG051029.D\data.ms (-2251) (-)

Sub

60.0 94.9 129.9 164.9 201.9 233.8 265.8

m/z-->

Abundance

50000

40000

30000

20000


10000

0

17.20 17.30

Time-->

Abundance Scan 2409 (17.641 min): BG051028.D\data.ms (-2470) (-)



Phenanthrene

Concen: 48.00 ng

RT: 17.642 min Scan# 2409

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050



Tgt Ion:178 Resp: 641058

Ion Ratio Lower Upper


178 100

176 20.2 16.0 24.0


179 16.6 14.2 21.4



Abundance Scan 2409 (17.642 min): BG051029.D\data.ms (-2319) (-)



Abundance Scan 2424 (17.729 min): BG051028.D\data.ms (-2472) (-)



Anthracene

Concen: 48.02 ng

RT: 17.736 min Scan# 2425

Delta R.T. 0.007 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48



Tgt Ion:178 Resp: 637327

Ion Ratio Lower Upper


178 100


176 19.1 14.6 21.8

179 16.4 13.4 20.0



Abundance Scan 2425 (17.736 min): BG051029.D\data.ms (-2334) (-)









Carbazole
Concen: 47.85 ng
RT: 18.000 min Scan# 2472
Delta R.T. 0.001 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48




Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Tgt Ion:	167	Resp:	632909
Ion	Ratio	Lower	Upper
167	100		
166	23.4	17.4	26.2
139	13.8	10.7	16.1




Di-n-butylphthalate
Concen: 48.18 ng
RT: 18.529 min Scan# 2560
Delta R.T. -0.005 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48

Tgt Ion:	149	Resp:	747825
Ion	Ratio	Lower	Upper
149	100		
150	9.9	7.5	11.3
104	6.3	6.6	10.0#



Abundance Scan 2749 (19.639 min): BG051028.D\data.ms (-275) (-)



Fluoranthene

Concen: 46.23 ng

RT: 19.640 min Scan# 1

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050




Tgt Ion:202 Resp: 758966

Ion Ratio Lower Upper


202 100

101 11.9 0.0 24.7

203 18.7 0.0 38.5



Abundance Scan 3134 (21.901 min): BG051028.D\data.ms (-3175) (-)



Chrysene-d12

Concen: 20.00 ng

RT: 21.896 min Scan# 3133

Delta R.T. -0.005 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48




Tgt Ion:240 Resp: 210208

Ion Ratio Lower Upper


240 100

120 10.3 4.2 6.2#

236 26.4 21.8 32.6






Abundance Scan 2779 (19.815 min): BG051028.D\data.ms (-2777) (-)




Benzidine
Concen: 24.23 ng
RT: 19.816 min Scan# 2779
Delta R.T. 0.001 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48

Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Tgt Ion:184 Resp: 165056
Ion Ratio Lower Upper
184 100
185 14.4 10.7 16.1
183 12.6 9.4 14.0






Abundance Scan 2811 (20.003 min): BG051028.D\data.ms (-2778) (-)




Pyrene
Concen: 50.68 ng
RT: 20.004 min Scan# 2811
Delta R.T. 0.001 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48

Tgt Ion:202 Resp: 755785
Ion Ratio Lower Upper
202 100
200 21.9 15.8 23.8
203 17.9 14.6 21.8



Abundance Scan 2842 (20.185 min): BG051028.D\data.ms (-2752) (-)



Terphenyl-d14

Concen: 100.74 ng

RT: 20.186 min Scan# 2842

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050



Tgt Ion:244 Resp: 1086648

Ion Ratio Lower Upper


244 100

212 8.6 6.9 10.3


122 11.1 5.0 7.6#



Abundance Scan 2842 (20.186 min): BG051029.D\data.ms (-2752) (-)



Abundance Scan 2958 (20.867 min): BG051028.D\data.ms (-2930) (-)



Butylbenzylphthalate

Concen: 52.97 ng

RT: 20.868 min Scan# 2958

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48



Tgt Ion:149 Resp: 334190

Ion Ratio Lower Upper


149 100

91 76.9 63.4 95.2


206 20.9 18.4 27.6



Abundance Scan 2958 (20.868 min): BG051029.D\data.ms (-2868) (-)



Abundance Scan 3130 (21.877 min): BG051028.D\data.ms (-3120) (-)



Benzo(a)anthracene

Concen: 50.20 ng

RT: 21.878 min Scan# 3130

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050



Tgt Ion:228 Resp: 698280

Ion Ratio Lower Upper


228 100

226 28.8 22.7 34.1


229 21.4 16.0 24.0



Abundance Scan 3130 (21.878 min): BG051029.D\data.ms (-3040) (-)



Abundance Scan 3114 (21.783 min): BG051028.D\data.ms (-3102) (-)



3,3'-Dichlorobenzidine

Concen: 70.47 ng

RT: 21.784 min Scan# 3114

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48



Tgt Ion:252 Resp: 324849

Ion Ratio Lower Upper


252 100

254 63.0 52.4 78.6

126 10.9 5.9 8.9#



Abundance Scan 3114 (21.784 min): BG051029.D\data.ms (-3024) (-)



Abundance Scan 3142 (21.948 min): BG051028.D\data.ms (-3136) (-)

Chrysene

Concen: 50.60 ng

RT: 21.949 min Scan# 3142

Delta R.T. 0.001 min

Lab File: BG051029.D


Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050





Tgt Ion:228 Resp: 661993

Ion Ratio Lower Upper


228 100

226 32.2 23.1 34.7

229 20.4 15.3 22.9



Abundance Scan 3142 (21.949 min): BG051029.D\data.ms (-3052) (-)



Abundance Scan 3107 (21.742 min): BG051028.D\data.ms (-3094) (-)

Bis(2-ethylhexyl)phthalate

Concen: 51.95 ng

RT: 21.743 min Scan# 3107

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48



Tgt Ion:149 Resp: 465628

Ion Ratio Lower Upper


149 100

167 26.8 22.3 33.5


279 3.7 3.8 5.8#



Abundance Scan 3107 (21.743 min): BG051029.D\data.ms (-3017) (-)



Abundance Scan 3323 (23.011 min): BG051028.D\data.ms (-3323) (-)



Di-n-octyl phthalate

Concen: 52.50 ng

RT: 23.012 min Scan# 3323

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDICC050




Tgt Ion:149 Resp: 784912

Ion Ratio Lower Upper


149 100

167 1.6 1.4 2.2

43 11.3 8.6 12.8



Abundance Scan 3713 (25.303 min): BG051028.D\data.ms (-3713) (-)



Perylene-d12

Concen: 20.00 ng

RT: 25.298 min Scan# 3712

Delta R.T. -0.005 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48




Tgt Ion:264 Resp: 217426

Ion Ratio Lower Upper


264 100

260 23.5 19.4 29.2

265 21.5 16.9 25.3





Abundance Scan 4380 (29.222 min): BG051028.D\data.ms (-4387) (-)




Indeno(1,2,3-cd)pyrene
Concen: 49.63 ng
RT: 29.223 min Scan# 4387
Delta R.T. 0.001 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48

Instrument :
BNA_G
ClientSampleId :
SSTDICC050


Tgt Ion:	276	Resp:	751624
Ion Ratio	Lower	Upper	
276	100		
138	24.3	5.1	7.7#
277	25.5	19.8	29.6



Abundance Scan 4380 (29.223 min): BG051029.D\data.ms (-4290) (-)





Abundance Scan 3528 (24.216 min): BG051028.D\data.ms (-3528) (-)





Benzo(b)fluoranthene
Concen: 50.39 ng
RT: 24.211 min Scan# 3527
Delta R.T. -0.005 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48

Tgt Ion:	252	Resp:	671013
Ion Ratio	Lower	Upper	
252	100		
253	22.9	16.2	24.4
125	9.8	3.0	4.4#



Abundance Scan 3527 (24.211 min): BG051029.D\data.ms (-3438) (-)





Benzo(k)fluoranthene

Concen: 48.95 ng

RT: 24.287 min Scan# 3540

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48

Instrument :

BNA_G

ClientSampleId :

SSTDIC050





Tgt Ion:252 Resp: 641272

Ion Ratio Lower Upper

252 100

253 23.6 18.0 27.0

125 11.1 3.1 4.7#



Benzo(a)pyrene

Concen: 50.24 ng

RT: 25.139 min Scan# 3685

Delta R.T. 0.001 min

Lab File: BG051029.D

Acq: 13 Nov 2021 18:48




Tgt Ion:252 Resp: 568803

Ion Ratio Lower Upper


252 100

253 21.3 16.6 25.0

125 10.4 3.5 5.3#





Abundance Scan 4391 (29.286 min): BG051028.D\data.ms (-4391) (-)




Dibenzo(a,h)anthracene
Concen: 49.28 ng
RT: 29.293 min Scan# 4391
Delta R.T. 0.007 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48

Instrument :
BNA_G
ClientSampleId :
SSTDICC050


Tgt Ion:	278	Resp:	617302
Ion Ratio	Lower	Upper	
278	100		
139	14.3	4.6	6.8#
279	23.1	17.9	26.9



Abundance Scan 4392 (29.293 min): BG051029.D\data.ms (-4301) (-)





Abundance Scan 4589 (30.450 min): BG051028.D\data.ms (-4592) (-)



Benzo(g,h,i)perylene
Concen: 50.02 ng
RT: 30.456 min Scan# 4590
Delta R.T. 0.007 min
Lab File: BG051029.D
Acq: 13 Nov 2021 18:48

Tgt Ion:	276	Resp:	613683
Ion Ratio	Lower	Upper	
276	100		
277	25.2	18.4	27.6
138	20.7	7.0	10.6#



Abundance Scan 4590 (30.456 min): BG051029.D\data.ms (-4499) (-)

