


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG111418\
 Data File : BG037996.D
 Acq On : 14 Nov 2018 14:25
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG111318.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 13 16:39:28 2018



AutoFind: Scans 2480, 2481, 2482; Background Corrected with Scan 2473

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.0	23083	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	30.7	19685	PASS
70	69	0.00	2	0.6	120	PASS
127	198	10	80	43.7	27960	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	64053	PASS
199	198	5	9	6.8	4351	PASS
275	198	10	60	27.1	17330	PASS
365	198	1	100	3.3	2145	PASS
441	198	0.01	100	9.2	5908	PASS
442	198	50	100	61.0	39093	PASS
443	442	15	24	19.3	7538	PASS